


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	Kannur
2	Name of Hospital	GH Thalassery
3	Name of Equipment with Make, Model and Serial Number	Refrigerator Make:- Nidicocon Model:- 235L
4	Equipment ID & Barcode	1311433
5	Date of purchase / Year of manufacture / Installation Date	10-4-2011
6	Warranty details (Yes/No)	No
7	*AMC/ CAMC Period agreed at the time of purchase	No
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	15-3-2023
9	Action taken	checked and found gas leaked due to freezer cabinet damage
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Compressor, Condensor, Relay with OLP, Thermostat, Freezer cabinet with door need to replace/damaged
12	Cost of spares (specify parts and cost)	Quotation attached


13	Asset Value	9988/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	117%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix report
16	Reasons for recommending the equipment as BER	This machine installed on 10-4-2011 and aged upto 11 years comparing the life span used, repairing and replacement of spares are beyond economic. so we recommend for condemnation
17	Name & Signature of CYRIX Authority	Rasilraj 

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


checked & identified that the service cost economically not feasible and aged upto 11 years. So recommended for BER as per BER Protocol.

  
11/04/2023

Signature of JC BM (NHM)

CONSULTANT BIOMEDICAL  
NATIONAL HEALTH MISSION  
KANNUR - 670002

Seema N




JSO

Date 29/3/23



Seal



Signature of  
Superintendent, Medical Officer (i/c)

GENERAL HOSPITAL  
THALASSERY - 870 101



**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



SERVICE PROVIDER  
Tender No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

No. : 104532

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>GH Thalassery</u>	Call Registration Date : <u>15-3-23</u>
Address <u>Thalassery</u>	Caller ID : <u>41109</u>
<u>Kannur</u>	Date of Visit : <u>29-3-23</u>
Ph : <u>9447343229</u>	Asset No. : <u>1311433</u>
	EQPT Name : <u>Refrigerator</u>
	Manufacture <u>vidicon</u> Model : <u>235L</u>
	S. No. <u>NA</u> Dept. <u>LAB</u>

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Machine not working

**Action Taken :** checked and found that gas leaked due to freezer gland damage. also compressor, condenser, Relay with O.P, Thermostate are damaged. This machine installed on 10-4-2011 and aged upto 11+ years. Comping the life span used, repairing and replacement of spairs are beyond economic. so we recoment the equipment for condemnation

Completed  Date : ..... Time : ..... Spare Required


Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Rasilraj</u>	<u>29-3-23</u>	<u>12:45PM</u>	<u>1:30PM</u>

**Customer Remark** Completed  Pending

Service Engineer Name : <u>Rasilraj</u>	Customer Name : <u>Seema</u>
Signature : <u>Rasilraj</u>	Signature : <u>Seema</u>
Date : <u>29-3-23</u>	Date : <u>29/3/23</u>
Contact Number : <u>9995521436</u>	Contact Number : <u>9446659624</u>
	Designation : <u>JSO</u>
	Hospital Seal :



**Superintendent  
GENERAL HOSPITAL  
THALASSERY - 670 10**

# SAJITH K

KUNNUMCHALIL HOUSE  
PALAYAD POST  
PHONE NO: 8943706200  
AC REFRIGERATION SERVICE& REPAIRING

Date: 23.03.2023

## ESTIMATE

NAME: WHIRLPOOL REFRIGERATION GH THALASSERY

SL.NO	ITEM	AMOUNT
1.	WHIRLPOOL REFRIGERATOR COMPRESSOR	3700/-
2.	FREEZER & FREEZER DOOR	2500/-
3.	CONDENSOR	1000/-
4.	RELAY OLP	600/-
5.	THERMOSTAT	400/-
6.	GAS CHARGING	2500/-
7.	SERVICE CHARGE	1000/-
		11700/-

  
Sajith k  
**SAJITH K**  
AC Refrigeration Service & Electrical Repairing  
Kunnumchalil House  
Palayad (P.O), Thalassery  
Mob: 8943706200



Engineering  
 and  
 Construction  
 by  
 26/11/22

FRI DHT Refrigerator 250L

21-2-2002 Old Stock Book III  
 Page No. 1107  
 Condemned as per order number CH/245/2022 dt. 29/7/2022  
 by the Superintendent A.H. Helarseni  
 23-2-10 Received from Store 1 3 ? Rs 17000/-  
 (Whirlpool - 250 litre)  
 Purchase order No P3 2867/10 dt 4-10-2010  
 S.No - INA 103300910

10-4-2011 Received from store 1 4 Rs 9988/-  
 (Videocon Refrigerator  
 VAG 243 - 235 litre)  
 (Transferred to OTI as loan basis)  
 Received on 10/11/11  
 (Stock Book Vol V - Instruments  
 of Store Page 66) # # 7

23-8-2011 Received for free with Fully- 1 5 Rs  
 Auto Analyzer - Bis his 24i  
 (Whirlpool - 280 litre (Genuine))  
 (Free Supply) Model - GENIUS - 310 litre  
 12883  
 St. No - INC - 110815413

6-8-11  
 16-8-11 Transferred to OTI 1 4 Rs  
 - Videocon - Stock Book of instruments  
 Vol II Page No 113 of OTI

Archived at per order B2/219/17  
 All A-3-17 of Superintendent A.H. Helarseni # 3 10