

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	CH CHANCHA NASSERY
3	Name of Equipment with Make, Model and Serial Number	CAUTERY ALAN ELECTRONIC SYSTEM ELSY-360M+ ASL-M1207024
4	Equipment ID & Barcode	0512142
5	Date of purchase / Year of manufacture / Installation Date	5/10/2012 17/1/2018.
6	Warranty details (Yes/No)	NO
7	*AMC/ CAMC Period agreed at the time of purchase	3 Years
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	18/9/2023
9	Action taken	Checked the equipment & tried to service the main board.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	Need to replace the main board of the equipment.
12	Cost of spares (specify parts and cost)	End of Life.

13	Asset Value	167400/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	End of life
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	By service report & recommends the equipment for RBER
16	Reasons for recommending the equipment as BER	The parts of the equipment is currently not available since the equipment is outdated. The OEM letter for end of life has received & recommend for condemnation
17	Name & Signature of CYRIX Authority	Seegun Suresh

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value - 167400
 cost of spare - EOL
 life period - 5 year -

The Authorized dealer of OEM submitted EOL report (JBS/service/gub) 001/09/2023 29/9/23
 As per RBER guidelines and above report recommended as RBER

16/10/25
 Signature of JC BM (NHM)

CONSULTANT BIOMEDICAL ENGINEER
 NATIONAL HEALTH MISSION
 KOTTAYAM

Date

Seal

Signature of Superintendent / Medical Officer (i/c)

Superintendent
 General Hospital
 Chandanathery



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 132639

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Peta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com E-mail : hemp.kl@cyrix.in

Health Facility : GH Changanassery	Call Registration Date : 18/9/2023
Address : Changanassery Kothayam	Caller ID : 69949
Ph : 98400131531	Date of Visit : 29/9/2023
	Asset No. : 0512142
	EQPT Name : Country Machine
	Manufacture : Alan Model : ELSY 360MT
	S. No. ASN - M20202 Dept. OT

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : *Charging & sagging battery not working. Equipment was taken for servicing.*

Action Taken : *Equipment returned & handedover to customer. The main board of the equipment is damaged. The spares of this machine is currently not available since the machine is outdated & the OEM has given an End of life certificate. Hence recommending for replacement. Completed Date : 29/9/2023 Time : 2:15pm Spare Required*

Spare Replaced	<input type="checkbox"/> Requested	<input type="checkbox"/>	Description	Qty.	Part Number	PR Number
			Cyrix Engineer			
			<i>Sreya Suresh</i>			
				Date	Start Time	End Time
				<i>29/9/2023</i>	<i>3:35pm</i>	<i>4:15pm</i>

Customer Remark
Equipment Rd, Poozhattam

Service Engineer Name : *Sreya Suresh*

Signature : *[Signature]*

Date : *29/9/2023*

Contact Number : *7902854749*

Completed **Pending**

Customer Name : *Poozhattam*

Signature : *[Signature]*

Date : *29/9/2023*

Contact Number : *Poozhattam*

Designation : *[Signature]*

Hospital Seal : *29/9/23*



END OF LIFE LETTER

JBS/Service/Gen/001/09-2023

29.09.2023

To,

M/S. Cyrix Healthcare Private Limited
1 St Floor, 30/64 1 B,
Kochi Dhanushkodi Road
Poonithura, Ernakulam-682038

(Kind Attn : Mr. Afsal-Zonal Manager.)

Respected Sir,

Sub: Alan make Diathermy End of Life letter – Reg.

Ref: Taluk Head Quarters Hospital, Changanacherry.

Make : ALAN ; Model : ELSY 360 M+ ; S.NO: 1204024.

As reported, the Alan make surgical diathermy was not working and completely breakdown.
We understood that this machine was already served more than 12years and it is more than the expected life of the machine.

Due to continuous technological updation, This particular model was outdated. Hence the spares are not available. So it is not possible to service it.

Hence we recommend you to treat the machine as End of Life.

Thanking you
for JBS Meditec India (P) Ltd.,
Service Department,
Kaarthikeyan.S/P,
Ph : 9442253991



“Sri Ragavendra Tower”, 3rd Floor, No-34, Co-operative E-colony, Behind Kumudham Nagar,
Villankurich Road, Coimbatore – 641035 Ph: 9443367915, 9443366752 Email ID: info@jbsmeditec.com

Diathermy Alligature • Argon Plasma Coagulator (APC) Saline-

Quotation for ALAN Brand Surgical Diathermy (EL-SY 360 L+ & M+) - Reg

6 messages

JBS marketing <marketing@jbsmeditec.com>

Fri, Sep 29, 2023 at 1:52 PM

To: afsalcyrix@gmail.com

Cc: M Mahendran <mm@jbsmeditec.com>

Dear Sir,

Greetings from JBS Meditec India Private Limited.

As per discussion with our M.D. Mr.M.Mahendran sir, Here we have enclosed our **ALIAN Brand Surgical Diathermy Quotation** and **EOL letter** for your reference.

Please find the attachment and do the needful.

Thank youWith best regards,
(S.Ambiga - Marketing Dept.)**JBS Meditec India Private Ltd,****"Sri Ragavendra Tower,"3rd Floor, No-34,****Co-operative E-colony,****Villankurichi Road,****Coimbatore – 641035****Ph: 9443367915.****3 attachments** **EOL letter 29.9.23.pdf**
392K **Quotation for Dia M+ 29.9.23.pdf**
5829K **Quotation for Dia L+ 29.9.23.pdf**
14111K**Afsal Y** <afsalcyrix@gmail.com>

To: JBS marketing <marketing@jbsmeditec.com>

Cc: M Mahendran <mm@jbsmeditec.com>

Fri, Sep 29, 2023 at 1:55 PM

Thank you for your mail.

[Quoted text hidden]

Afsal Y <afsalcyrix@gmail.com>

Fri, Sep 29, 2023 at 1:55 PM

EOL letter

[Quoted text hidden]

3 attachments

 **EOL letter 29.9.23.pdf**
392K

 **Quotation for Dia M+ 29.9.23.pdf**
5829K

 **Quotation for Dia L+ 29.9.23.pdf**
14111K

Sun, Oct 1, 2023 at 9:47 AM

Afsal Y <afsalcyrix@gmail.com>

To: JBS marketing <marketing@jbsmeditec.com>

Cc: M Mahendran <mm@jbsmeditec.com>

Dear Sir/ Madam,

Kindly revert back these three files after adding signature and seal since we want to submit it to the Government hospital.

Do the needful on an urgent basis.

Have a nice day.

Regards

AFSAL YUNUZE

Zonal Manager (South Region)- Kerala

CYRIX HEALTHCARE Pvt. Ltd

30/6/41B, Petta, Poonithura, Kochi-38

Ph:+919895061437

[Quoted text hidden]

Tue, Oct 3, 2023 at 10:17 AM

JBS marketing <marketing@jbsmeditec.com>

To: afsalcyrix@gmail.com

Cc: M Mahendran <mm@jbsmeditec.com>

Dear sir,

Please find the attachment and do the needful.

[Quoted text hidden]

3 attachments

 **M+.pdf**
993K

 **L+.pdf**
1010K

 **covering letter.pdf**
219K

Wed, Oct 4, 2023 at 10:32 AM

Afsal Y <afsalcyrix@gmail.com>

To: JBS marketing <marketing@jbsmeditec.com>

Cc: M Mahendran <mm@jbsmeditec.com>

Thanks for the mail.

AFSAL YUNUZE
Zonal Manager (South Region)- Kerala
CYRIX HEALTHCARE Pvt. Ltd
30/6/41B, Petta, Poonithura, Kochi-38
Ph: +919895061437

[Quoted text hidden]

Certificate of Distributorship

ALAN

ISO 9001:2015

ISO 13485:2016

EU CE 1639



This serves to certify that

JBS Mediitec (I) Pvt. Ltd.

Sri Ragavendra Tower, Villankurichi Road, Coimbatore, Tamil Nadu-641035

is

an **Authorized Distributor**

for **Medical Devices and Accessories**

manufactured by **ALAN**

for the territory of

Tamil Nadu, Pondicherry & Kerala.



GURUPRASAD IYER

Director

This certificate is valid upto 30th April 2024