

245

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Kottayam
2	Name of Hospital	Phc Thalassadu
3	Name of Equipment with Make, Model and Serial Number	BP Apparatus Make: Diamond Model: Deluxe Sn: 41879
4	Equipment ID & Barcode	0543495
5	Date of purchase / Year of manufacture / Installation Date	16-03-2015
6	Warranty details (Yes/No)	No
7	*AMC/ CAMC Period agreed at the time of purchase	No
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	13-11-2023, #79828
9	Action taken	Serviced.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	Required mercury, BP glass tube, BP bladder BP bulb with valve, BP connector etc.
12	Cost of spares (specify parts and cost)	Mercury - 1200 (with glass tube) BP bladder - 80 BP bulb - 60 BP glass washer - 30 BP connector - 20

Total spare cost: 1390/-

13	Asset Value	2100
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	66.2 %
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix.
16	Reasons for recommending the equipment as BER	According to clause 5.3.14.1, equipment is recommended for condemnation Tender NO: KMSCL/EP/1424(R)/698R021
17	Name & Signature of CYRIX Authority	Nidhish Roy, <u>NBS</u>

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

As set value = 2100
 Cost of Spare : 1390
 Life period. → 8 year.
 As per the RBER guidelines recommended as RBER

[Signature]
 24/11/23
 Signature of JC BM (NHM)
 CONSULTANT BIOMEDICAL ENGINEER
 NATIONAL HEALTH MISSION
 KOTTAYAM

Date 13/11/23



[Signature]
 Signature of
 Superintendent / Medical Officer (i/c)
 MEDICAL OFFICER IN CHARGE
 PRIMARY HEALTH CENTRE
 THALAYASSALAM P O
 KOTTAYAM DIST



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 160664

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency
30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Service Report

Health Facility .. Phc Thalavady	Call Registration Date : 13-11-2023
Address Thalavady	Caller ID : 79808
..... Kottayam	Date of Visit : 13-11-2023
Ph : 9496424656	Asset No. : 0543495
	EQPT Name : BP Apparatus
	Manufacture Diamond Model : deluxe
	S. No. 41879 Dept. OP

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : BP glass tube broken, Required spares - BP cuff, BP bladder and Bulb with valve.

Action Taken : Checked. BP glass tube broken and mercury lost. Required all spare parts for reconditioning. Not repairable economically. So recommended for BER.

Completed Date : 13-11-2023 Time : 12.38 pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Nidhish Roy	13-11-2023	11.59 pm	12.38 pm

Customer Remark Completed Pending

Service Engineer Name : Nidhish Roy
Signature : *Nidhish Roy*
Date : 13-11-2023
Contact Number : 9072575296

Customer Name : Dr Prima Thomas
Signature : *Dr Prima Thomas*
Date : 13/11/23
Contact Number : 9946808584
Designation : Medical Officer in Charge
Hospital Seal : *Thalavady PHC Kottayam Dist*



Stock & Issue Register

75

Name of Item B.f. A.P.P.A.R.A.T.U.S. (Diamond)

Date	Particulars	Bill No. & Date	Quantity Received	Quantity Issued	Quantity Balance	Remarks
16.3.15	IN. NO. 167246 of 13.3.15 Kamruja Community Pharmacy Kottayam	167246 13.3.15	1 NO		1 NO.	1 Rs. 2100
17.3.15	Issued to Staff Nurse. Senthil			1 NO.	NULL	Sub. Pag 245.

5
104
100



