


246

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Kottayam
2	Name of Hospital	Phc Thalassadd
3	Name of Equipment with Make, Model and Serial Number	BP Apparatus Make: Life line Model: Ultra sn: 88478
4	Equipment ID & Barcode	0543498
5	Date of purchase / Year of manufacture /Installation Date	08-08-2018
6	Warranty details (Yes/No)	No
7	*AMC/ CAMC Period agreed at the time of purchase	No
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	13-11-2008, # 79807
9	Action taken	Serviced.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	Required mercury, BP glass tube, bladder, BP bulb, BP connector etc.
12	Cost of spares (specify parts and cost)	Mercury - 1200 (with glass tube) BP bladder - 80 BP bulb - 60 BP connector - 20 Total spare cost = 1360/-


13	Asset Value	2128
14	*Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	63.9%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix.
16	Reasons for recommending the equipment as BER	According to clause 5.3.14.4, equipment is recommended for condemnation. Tender NO: KMSCL/EP/TAR(R)/698/2021 Clause - 5.3.14.1
17	Name & Signature of CYRIX Authority	Nidhish Roy 

*Not mandatory #Based on the period of life and value as per the BER guidelines


* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value - 2128
 Cost of spare : 1360
 Life period - 5 year
 As per the RBER guidelines recommended as RBER


 Signature of JC BM (NHM)
 CONSULTANT BIOMEDICAL ENGINEER
 NATIONAL HEALTH MISSION
 KOTTAYAM

Date 13/11/2021


 Seal
 THALASSERY
 PIN - 688 553


 MEDICAL OFFICER IN CHARGE
 PRIMARY HEALTH CENTRE
 Signature of THALANADU P O
 Superintendent / Medical Officer (i/c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 160662

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>Pbc Thalassadu</u></p> <p>Address <u>Thalassadu</u> <u>Kottayam</u></p> <p>Ph : <u>9496424656</u></p>	<p>Call Registration Date : <u>13-11-2023</u></p> <p>Caller ID : <u>79807</u></p> <p>Date of Visit : <u>13-11-2023</u></p> <p>Asset No. : <u>0543498</u></p> <p>EQPT Name : <u>BP Apparatus</u></p> <p>Manufacture <u>Life Line</u> Model : <u>Citra</u></p> <p>S. No. <u>88498</u> Dept. <u>OP</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : BP glass tube broken, Mercury lost,
Required BP bladder, BP bulb with valve.

Action Taken : Checked. Low mercury. Required all
spares of BP apparatus for reconditioning.
Not repairable economically. So recommended
for BER

Completed Date : 13-11-2023 Time : 12:05pm Spare Required

Spare Replaced Requested

1.	Description	Qty.	Part Number	PR Number
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Nidhish Roy</u>	<u>13-11-2023</u>	<u>11:35am</u>	<u>12:05pm</u>

Customer Remark Completed Pending

Service Engineer Name : Nidhish Roy
Signature : [Signature]
Date : 13-11-2023
Contact Number : 9070575296

Customer Name : Dr Pritha Thomas
Signature : [Signature]
Date : 13-11-2023
Contact Number : 9446805784
Designation : Medical Officer in Charge
Hospital Seal : KOTTAYAM Dist





**NATIONAL HEALTH MISSION
UNDER
MAINTENANCE PROGRAMME
BIOMEDICAL EQUIPMENT**



No. : 180665

CYRIX
HEALTHCARE PVT LTD

Service Provider
Order No. WO-3715021-2022098

ISO 13485 : 2013 & ISO 9001-2008 CERTIFIED COMPANY / AERIAL SERVICE AGENCY
3084 F B, Patta Junction, Poonithur, Kochi - 682 038, Kerala
Ph : 98422 92500 Website : www.cyrix.com E-mail : help@cyrix.in

Service Report

Call Registration Date : 13-11-2023	Health Facility : Pbc Thalavada
Caller ID : 98422	Address : Thalavada
Date of Visit : 13-11-2023	Kottayam
Asset No. : 021348	Ph : 98422
EPT Name : Bp Apparatus	
Manufacturer Life Line Model : Clifox	
S.No. 021348 Dept. OP	

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Bp glass tube broken & mercury lost
Replaced Bp bladder, Bp bulb with valve.

Action Taken : Checked. Low mercury. Replaced all
spare of Bp apparatus for reconditioning.
Not responsible economically. go accordingly
for Bp

Completed Date : 13-11-2023 Time : 12:30pm
Spare Required

Spare Relieved	Requested	Description	Qty	Part Number	PR Number
<input type="checkbox"/>	<input type="checkbox"/>				

Cylix Engineer	Date	Start Time	End Time
Michael Roy	13-11-2023	11:30am	12:30pm

Customer Remark: Completed Pending

Service Engineer Name : Michael Roy
Signature :

Date : 13-11-2023

Contact Number : 98422

Hospital Seal : KOTTAYAM DIST
Designation : NIGHT ENGINEER
Control Number : 180665
Signature :

Customer Name : Pbc Thalavada
Signature :



20/3/18 Received from Motherkennel Invoice no 1599 20 March 2018.

Item	Rate	Qty	Rs
Plate	120	4	480

Total 480/-

Issued the above items to IHT
 bal nil
 Pharmaceutical

5/5/18 Received from Suresh Durg, Inv no 61434 dated 05/05/2018

Item	MRP	Rate	Qty	GST%	Sgst	Cgst	GST	Amount
B.P. Apparatus	3000	1900	1	12	114	114	228	2125

Issued the above items to S/N. bal nil
 Pharmaceutical

5/5/18 Received from Suresh Durg, Inv no 61435 dated 5/5/18

Sl No	Item	MRP	Rate	Qty	Amnt	GST%	Sgst	Cgst	GST	Amnt
1	Digital Thermometer	235	200	1	120	18	10.50	10.50	21.00	2160
2	Stethoscope	1049	650	1	650	12	40.50	40.50	81.00	7410

Gross value 800

GST 103.20

Net Amount 903.20
 Issued the above items to S/N. bal. nil

Pharmaceutical

8/5/18 Received from Jindalpackal paper mill dated 8/5/18

Sl No	Item	Unit price	Qty	Gross value
1	Cone	7	33	247
2	Cone	8	26	208

Total 505

Discount 5

Total 500

Issued the above items to pharmaceutical & bal. nil

Pharmaceutical

Received 8/5/18

