


241

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	Pathanamthitta
2	Name of Hospital	phc Pandakara Jhekkakara
3	Name of Equipment with Make, Model and Serial Number	Laboratory microscope Labovisione. splendour 8175927
4	Equipment ID & Barcode	ID: 79094 Barcode: 0340868
5	Date of purchase / Year of manufacture / Installation Date	found excess on store Vankarate. in 31/8/2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	N/A
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	08/11/2023
9	Action taken	Recommended BER
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	N/A
12	Cost of spares (specify parts and cost)	N/A

13	Asset Value	RS: 15400
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	YES
16	Reasons for recommending the equipment as BER	fully rusted and damaged
17	Name & Signature of CYRIX Authority	Sijo Zachariah 

\*Not mandatory **#Based on the period of life and value as per the BER guidelines**


\* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Physically verified on 13/11/2023. Recommended for BER.


  
16/11/2023  
Signature of JC BM (NHM)


Date



Signature of  
Superintendent / Medical Officer (i/c)









# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 127104

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <i>PHC Pandalam Thekkekkara</i></p> <p>Address <i>Thadayi, Pandalam Thekkekkara Pathanamthitta, Pin: 691525</i></p> <p>Ph : <i>9745275905</i></p>	<p>Call Registration Date : <i>08/11/2023</i></p> <p>Caller ID : <i>79094</i></p> <p>Date of Visit : <i>08/11/2023</i></p> <p>Asset No. : <i>0340868</i></p> <p>EQPT Name : <i>Laboratory microscope</i></p> <p>Manufacture <i>Labovision</i> Model : <i>Splendour</i></p> <p>S. No. <i>3175927</i> Dept. <i>LAB</i></p>
--	--

**Service Classification : Breakdown Call**  **PMS**  **Calibration**  **Cust.Training**

**Problem Identified :** *Not working*  
*On checking found that the microscope fully damaged all the parts are rusted*

**Action Taken :** *checked the machine found that the all part of the microscope. Objective's eye piece, Stage, Adjusting knob, light source and all other parts are fully damaged with rust machine fully rusted not repairable. Re-converted to BER*

**Completed**  **Date :** *08/11/2023* **Time :** *03:50 pm* **Spare Required**

**Spare Replaced**  **Requested**

Description	Qty.	Part Number	PR Number
1.			
2. <i>x/a</i>	<i>Na</i>	<i>Na</i>	<i>Na.</i>
3.			

Cyrix Engineer	Date	Start Time	End Time
<i>Sijo Zachariah</i>	<i>08/11/2023</i>	<i>03:00 pm</i>	<i>03:50 pm</i>

**Customer Remark** **Completed**  **Pending**

<p><b>Service Engineer Name :</b> <i>Sijo Zachariah</i></p> <p><b>Signature :</b> <i>[Signature]</i></p> <p><b>Date :</b> <i>08/11/2023</i></p> <p><b>Contact Number :</b> <i>8113954221</i></p>	<p><b>Customer Name :</b> <i>[Signature]</i></p> <p><b>Signature :</b> <i>[Signature]</i></p> <p><b>Date :</b> <i>[Signature]</i></p> <p><b>Contact Number :</b> <i>[Signature]</i></p> <p><b>Designation :</b> <i>[Signature]</i></p> <p><b>Hospital Seal :</b> <i>[Signature]</i></p>
--	---



*[Signature]*  
 മെഡിക്കൽ ഉപകരണങ്ങൾ  
 പരിപാലനം, മെയിൻറനൻസ്  
 ഫോൺ: 98472 99500 | 913517  
 pandalam@cyrix.com





**NATIONAL HEALTH MISSION  
UNDER  
MAINTENANCE PROGRAMME  
BIOMEDICAL EQUIPMENT**



SERVICE PROVIDER  
Tender No WO-3715021-2021088

**CYRIX**  
HEALTHCARE PVT LTD

No. : 127104

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency  
30164 F, Patta Junction, Poonithur, Kochi - 682 038, Kerala  
Ph : 98472 88200 Website : www.cyrrix.com E-mail : temp.kl@cyrrix.in

**Service Report**

Call Registration Date : 02/11/2023  
 Caller ID : F.P.O.P  
 Date of Visit : 02/11/2023  
 Asset No. : 0340 & 68  
 EQPT Name : Laboratory microscope  
 Manufacturer Model : SPL  
 S. No. 317298 Dept. LAB  
 Health Facility : The Pambam Health Center  
 Address : Thekkayil, Pambam, Thekkayil  
 Ph : 954729002

Service Classification : Breakdown Call  PMS  Calibration  Cust. Training

Problem Identified : Not working  
 All the parts are checked and replaced. The machine is fully changed.

Action Taken : Checked the machine found that all parts of the microscope, objective, eyepiece, stage, platform, body, lens cap and all other parts are fully checked and replaced. The microscope is fully checked and replaced. No reported to BCR.

Completed  Date : 02.11.2023 Time : 03.20 PM  
 Spare Required

Spare Replaced  Requested

Description	Qty	Part Number	PR Number
1	1	1	1
2	1	1	1
3	1	1	1

Cyril Engineer	Date	Start Time	End Time
S. S. Sankar	02/11/2023	09:00 AM	03:20 PM

Customer Remark : Completed  Pending


Service Engineer Name : S. S. Sankar  
 Signature : [Signature]  
 Date : 02/11/2023  
 Contact Number : 9139 21111  
 Hospital Seal : [Seal]  
 Designation : [Blank]  
 Contact Number : [Blank]  
 Signature : [Signature]  
 Customer Name : [Blank]

Phc Pandalam Thekkelaga

Laboratory microscope

Call ID : 081112 79094

Barcode : 0340868





From,

Medical Officer/c

P.H.C Pandalam Thekkela


To,

CYRIX

Sir,

ചന്ദ്രൻ തമ്പുരാൻ പ്രാഥമികാരോഗ്യ കേന്ദ്രത്തിൽ ഉപയോഗ കൃത്യമായ താലൂക്കുതലത്തിൽ ഉൾപ്പെടുത്തി ഉപകരണങ്ങൾ Condensation ഉപകരണത്തിൽ അധികൃതരുടെ സഹായം നൽകി സ്വീകരണത്തിന് അപേക്ഷിക്കുന്നു.

S.No.	Items	Date of Receipt	Remarks
1.	Autoclave	22/3/2017	
2.	Sterilizer	08/12/1997	
3.	Needle Destroyer	9/4/2012	
4.	Microscope Binocular	31/8/2019	found excess on store verification in 2019
5.	Microscope Binocular	23/3/2006	
6.	Calorimeter	25/3/2006	



എന്ന,

ഓഫീസ്, ചന്ദ്രൻ തമ്പുരാൻ  
 പ്രാഥമികാരോഗ്യ കേന്ദ്രം  
 പാണ്ടലം തേക്കേല, തൃശ്ശൂർ ജില്ല  
 ഫോ: 04784 - 812817



