



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 108025

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>PHC Ayancheri</u>	Call Registration Date : <u>23/11/23</u>
Address : <u>Kozhikode</u>	Caller ID : <u>82016</u>
Ph : <u>7907973618</u>	Date of Visit : <u>28/11/23</u>
	Asset No. : <u>1141793</u>
	EQPT Name : <u>UPS for Semi Auto Biochemistry Analyser</u>
	Manufacture <u>APC</u> Model : <u>BX600C</u>
	S. No. <u>B21916007659</u> Dept. <u>Laboratory</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Machine not getting ON

Action Taken : Checked and found Mainboard Battery, Transformer and Relay unit defective need to replace those spares for further checking and working condition of equipment

Completed Date : 04/12/23 Time : 10:30am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <u>na</u>	<u>na</u>	<u>na</u>	<u>na</u>
2. <u>na</u>	<u>na</u>	<u>na</u>	<u>na</u>
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Abhil Suresh S</u>	<u>04/12/23</u>	<u>10am</u>	<u>10:30am</u>

Customer Remark

Completed

Pending

Service Engineer Name : Abhil Suresh S
Signature : [Signature]
Date : 04/12/23
Contact Number : 9995706213



Customer Name : DICKSON F S
Signature : [Signature]
Date : 04/12/23
Contact Number : 04/12/23
Designation :
Hospital Seal :

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Kozhikode
2	Name of Hospital	PHC Ayancheri
3	Name of Equipment with Make, Model and Serial Number	UPS for Semi Auto Biochemistry Make: APC Model: BX 600 Serial: BC 21916007659 No
4	Equipment ID & Barcode	82016 & 1141793
5	Date of purchase / Year of manufacture / Installation Date	12/09/2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC No CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	23/11/2023
9	Action taken	Checked and found main board, Battery, Transformer & Relay unit defective
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not recommending for repair
12	Cost of spares (specify parts and cost)	Battery - Rs 1100 Main Board - Rs 2500 Other components - Rs 500 [Transformer, Relay etc] Rs 4100 <u>Rs 4100</u>

13	Asset Value	₹4500
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	91%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service Report Attached & Vendor quotation
16	Reasons for recommending the equipment as BER	Checked & Found Main Board, Battery Transformer, Relay unit other components defective. Equipment installed on 12/09/2019 completed 4 years and 3 months. The repair cost is beyond economical. Met both criteria So recommended for condemnation process
17	Name & Signature of CYRIX Authority	Akhil Suresh S <i>Akhil Suresh</i>

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Signature of JC BM (NHM)

Date _____ Seal _____



Superintendent / Medical Officer (i/c)

Hudya

GSTIN: 32AAJFB3520H1ZY
SSI No: 3204020549
SSI Date: 31-7-04

Ph: 0496 2525176 (Shop)
2512680 (Service)

Battery House

BATTERY, UPS, HUPS, SOLAR & INVERTER SALES & SERVICE
Narayana Nagar Ground, Vatakara - Pin: 673 101



Whole sale & Retail Outlet

Quotation

Date: 01/12/2020

To,
The Manager
Cyrix Health Care Pvt. Ltd.

Sir,

Good via APC - UPS - Company checked and
found Complaint. spare rate is

- I. Battery - Rs. 1100
- II. Main board - Rs. 2500
- III. Other Components
(Transformer, Relay, etc.) } Rs. 500

Rs. 4100

(Rupees four thousand and one hundred only)

Vatakara



to BATTERY HOUSE
Manager