



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 143539

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>CHC PAPPINISSERY</u></p> <p>Address <u>PAPPINISSERY</u> <u>KANNUR</u></p> <p>Ph : <u>8848665721</u></p>	<p>Call Registration Date : <u>4/10/23</u></p> <p>Caller ID : <u>72792</u></p> <p>Date of Visit : <u>6/10/23</u></p> <p>Asset No. : <u>1330876</u></p> <p>EQPT Name : <u>BP apparatus</u></p> <p>Manufacture <u>Omson</u> Model : <u>HEM 8712</u></p> <p>S. No. <u>20191136391V6</u> Dept. <u>OP</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : checked the equipment found BP mainboard, display and pump defective. equipment installed on 8/12/17 aged upto 6+ years comparing lifespan equipment used, repairing cost and replacement cost beyond economical repairable and so we recommended to condemnation

Completed Date : 6/10/23 Time : 11:30 Spare Required

Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>ASWIS</u>	<u>6/10/23</u>	<u>10:30</u>	<u>11:30</u>

Customer Remark Completed Pending

Service Engineer Name : ASWIS
Signature :
Date : 6/10/23
Contact Number : 8301012759

Customer Name : Bindumok CS
Signature :
Date : 6/10/23
Contact Number : 8848665721
Designation : Senior NSG Officer
Hospital Seal :




Medical Officer -
Community Health Cent
Pappinissery

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	CHC PAPPINISSERY
3	Name of Equipment with Make, Model and Serial Number	BP Apparatus OMRON HEM 8712 20191136391V6
4	Equipment ID & Barcode	72792 & 1330876
5	Date of purchase / Year of manufacture /Installation Date	8/12/2017
6	Warranty details (Yes/No)	No
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	4/10/2023
9	Action taken	checked and found mainboard, display and pump defected
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No Recommendation.
12	Cost of spares (specify parts and cost)	quotation attached


13	Asset Value	2200/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	106%.
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service Report Attached.
16	Reasons for recommending the equipment as BER	Checked the equipment found BP mainboard, display and pump defective. equipment installed on 31/12/17. aged upto 6+ years. Comparing lifespan equipment used, repairing cost and replacement cost beyond economical repairable and so we recommended to condemnation.
17	Name & Signature of CYRIX Authority	Aswin Ramachandran. 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


checked & identified that mainboard, display and pump are defective. aged up to 6 years. service cost economically not feasible as per BER protocol.


14/10/2023

Signature of JC BM (NHM)

Bindumole CS Body
Senior NSG officer
Ph: 884866572
Date 6/10/2023


Seal


Signature of Superintendent/Medical Officer (i/c)



AR MEDICARE

AR MEDICARE

ROOM NO: 865, 2ND FLOOR
AL-RAHA ARCADE, THAZHE CHOVA
KANNUR-670018

EMAIL: armedicare07@gmail.com

MOB: 7994654108, 9446057350

OUR REF : AR/QTN/109-2023

DATE : 09/10/2023

TO
CYRIX HEALTHCARE
ERNAKULAM

Respected Sir,

SUB: QUOTATION FOR SAMSO DIGITAL WEIGHING MACHINE PARTS

Kindly note with reference to the discussions had with yourself regarding the above mentioned subject, we are pleased to submit our very special competitive offer for your reference and further perusal.

In case of any further clarifications kindly feel free to contact us.

Thanking you and assuring you of our best attention and engineering services at all times.

Waiting for your earliest reply.

Yours truly,

For AR MEDICARE,
KANNUR.



TRUST US WITH YOUR NEEDS

OUR REF : AR/QTN/109-2023
DATE : 09/10/2023

SLNO:	DESCRIPTION	TOTAL(INCL.ALL TAXES)
01 02	SAMSO DIGITAL WEIGHING MACHINE MAINBOARD LOAD CELL 4 NOS	1400/- 800/-
	TOTAL	2,200/-

TERMS & CONDITIONS OF SALE

Payment Terms : Complete Payment Within 30 Days

Delivery : Within 4 Days from the date of purchase order .



AR MEDICARE

TRUST US WITH YOUR NEEDS