



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 154949

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility Phe Kelakam</p> <p>Address Kelakam, kannur Kerala</p> <p>Ph : 9074278714</p>	<p>Call Registration Date : 13-11-23</p> <p>Caller ID : 79784</p> <p>Date of Visit : 22-11-23</p> <p>Asset No. : 1345310</p> <p>EQPT Name : Sterilizer</p> <p>Manufacture Lifeplus Model : NA</p> <p>S. No. NA Dept. Dressing room.</p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Pressure gauge, pressure indication whistle, gaiked.
Coil 2000w, Door locking knobbed are defective.

Action Taken : checked the equipment and found pressure gauge, pressure indication whistle, Door Locking knobbed, coil 2000w and gaiked are defective - requested vendor for quotation. Equipment was installed on 5-10-11 and aged over 12 years. comparing the life span used the repairing and replacement costs are beyond economical. so we recommended for condemnation as per kmscl tender clause 5.3.14.1

Completed Date : 22-11-23 Time : 1:30 PM Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.				
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
Shyamkrishnan PPK	22-11-23	12:00 PM	1:30 PM

Customer Remark Completed Pending

<p>Service Engineer Name : Shyamkrishnan PPK</p> <p>Signature : <i>[Signature]</i></p> <p>Date : 22-11-23</p> <p>Contact Number : 8281188514</p>	<p>Customer Name : D. SHABANA K</p> <p>Signature : <i>[Signature]</i></p> <p>Date : 22/11/2023</p> <p>Contact Number : 9074278715</p> <p>Designation : MEDICAL OFFICER</p> <p>Hospital Seal : PRIMARY HEALTH CENTRE KELAKAM</p>
--	---



**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)


PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	PIHC KELAKAM.
3	Name of Equipment with Make, Model and Serial Number	NAME : STERLIZER MAKE : LIFE PLUS MODEL : NA SERIALNUMBER: NA
4	Equipment ID/ Barcode	1345310
5	Date of purchase/ Year of manufacture/Installation Date	5-10-11
6	Warranty details	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC / CAMC
8	Date of breakdown/ Date of registration of complaint	13-11-23
9	Action taken	Checked the equipment & Found Coil 2000w, Pressure gauge, pressure - indication whistle, Door locking. Knob set, gauges are defective. Enquired for quotation.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Details of service required (Recommendations for repair)	Not recommended
12	Cost of spares (specify parts and cost)	Coil 2000 w - 3500 pressure gauge : - 2500 pressure indication whistle - 1500 Gasket - 1500 Door locking knob set - 1000 Total - 10000 + GST



Phang
Dr. Sha Bina K

**MEDICAL OFFICER
PRIMARY HEALTH CENTRE
KELAKAM**

13	Asset Value	15000
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	66.66 %.
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX	Service report attached.
16	Reasons for recommending the equipment as BER	checked the equipment found pressure gauge, pressure indicating whistle, coil 2000w, Door locking knob set, gaskets are defective, requested vendor for quotation. Equipment is installed on 5-10-11 & aged more than 12 years comparing the lifespan used the repairing & replacement costs are beyond economic so we recommend for condemnation as per KMSCL tender clause 5.3.14.1
17	Signature of CYRIX Authority	Slyamkrishna PPK 

* Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that pressure gauge, indicating whistle, coil, door locking knob, gasket are defective and also major parts rusted. Equipment aged up to 12 years above, so service cost economically not feasible. Recommended for BER as per BER.

Prepared:



Signature of JC BM (NHM)

27/11/2023

CONSULTANT BIOMEDICAL
NATIONAL HEALTH MISSION
KANNUR

Used for more than 12 yrs.
Recommended for Condemnation

Date 22/11/2023

Signature of
Superintendent/Medical Officer (i/c)





MEDICAL OFFICER
PRIMARY HEALTH CENTRE
KELAKAM

Phang
MEDICAL OFFICER
PRIMARY HEALTH CENTRE
KELAKAN





Phary

MEDICAL OFFICER
PRIMARY HEALTH CENTER
SELAKANG





Sreeraj Surgicals

Aiswarya Commercial Complex, Nr. Ashirvad Hospital, Thavakkara, Kannur - 670 002
Ph: 0497 2706307. Mob: 9349846307, 9387529848
Email: sreerajsurgical@gmail.com, rajeshsreerajsurgical@gmail.com
www.sreerajsurgical.com Follow us on:  www.facebook/ sreerajsurgicals

Date:

20-11-2023

QUOTATION

TO,

CYRIX HEALTHCRAE
THRIPOONITHARA
ERNAKULAM

SIR,

WE ARE VERY PLEASURE TO ENQUIRY OF RATE OF FOLLOWING ITEMS.

S.L.NO	ITEMS	AMOUNT
1	2000 W COIL	3500.00
2	PRESSURE GUAGE	2500.00
3	PRESSURE INDICATION WHISTLE	1500.00
4	GASKET	1500.00
5	DOOR LOCKING KNOBE SET	1000.00

GST INCLUDED

WE EXPECT YOUR VALUABLE ORDERS AND KIND CO.OPERATION WITH US.

THANKING YOU.



Authorised Dealer:

