



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 143119

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility ...PHC...KANNAPURAM.....	Call Registration Date : ...15-11-23.....
AddressKANNUR.....	Caller ID :80329.....
Ph :9497597113.....	Date of Visit :15-11-23.....
	Asset No. :1342623.....
	EQPT Name :NEBULIZER.....
	ManufactureAEROLIFE..... Model : CN113 ARVIDO
	S. No.AA08994..... Dept.OP.....

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :found compressor in defective condition.....

Action Taken :Checked the equipment and found compressor in defective condition. Found equipment was installed on 01-01-2013 and aged upto 10 years. While comparing the life span used, the repairing cost and replacement cost of spares are beyond economic. Hence we recommending for condemnation.....

Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Jineshkumar	15-11-23	1pm	2pm

Customer Remark Completed Pending

Service Engineer Name : Jineshkumar
Signature :
Date : 15-11-23
Contact Number : 9895706435

Customer Name : Mr. Mohd. Abdul Rasheed
Signature :
Date : 15-11-23
Contact Number : 9497597113
Designation :
Hospital Seal :




**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	PHC KANNAPURAM
3	Name of Equipment with Make, Model and Serial Number	Name : NEBULIZER MAKE : AEROCARE MODEL : CN117 ARVIDO SL NO : AA03994
4	Equipment ID & Barcode	1342623 #80329
5	Date of purchase / Year of manufacture /Installation Date	01-01-2013
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC / CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	15-11-23
9	Action taken	Checked the equipment and found the repairing cost and replacement cost of spare are beyond economical. so recommending for BER
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	Not recommending for repair
12	Cost of spares (specify parts and cost)	Quotation attached


13	Asset Value	1353
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	147.81%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	SERVICE REPORT ATTACHED
16	Reasons for recommending the equipment as BER	Checked the equipment and found compressor in defective condition found equipment was installed on 01-01-2013 and aged upto 10 years while comparing the lifespan used, the repairing cost and replacement cost of spares are beyond economical. Hence recommending for BER.
17	Name & Signature of CYRIX Authority	JINECKUMAR PP 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that compressor & piston are defective. Aged upto 10 years above. So the repairs are economically ~~not~~ not feasible. As per BER protocol recommended for BER.



11/11/2023.
Signature of JC BM (NHM)

CONSULTANT BIOMEDICAL
NATIONAL HEALTH MISSION
KANNUR - 670002



Date 15-11-2023

Seal



Signature of Superintendent Medical Officer (i/c)

മെഡിക്കൽ ഓഫീസർ (i/c)
ദർമ്മപുരി ആരോഗ്യ സേവന മിഷൻ
കണ്ണൂർ, പി.ഒ. നമ്പർ: 670 331
ഫോൺ: 04972863468
dhckannapuram@gmail.com



AR MEDICARE

AR MEDICARE

ROOM NO: 865, 2ND FLOOR
AL-RAHA ARCADE, THAZHE CHOVVA
KANNUR-670018
EMAIL: armedicare07@gmail.com
MOB: 9446057350, 7994654108

OUR REF : AR/QTN/128-2023
DATE : 15/11/2023

TO
CYRIX HEALTHCARE
ERNAKULAM

Respected Sir,

SUB: QUOTATION FOR NEBULIZER PARTS

Kindly note with reference to the discussions had with yourself regarding the above mentioned subject, we are pleased to submit our very special competitive offer for your reference and further perusal.

In case of any further clarifications kindly feel free to contact us.

Thanking you and assuring you of our best attention and engineering services at all times.

Waiting for your earliest reply.

Yours truly,



For AR MEDICARE,
KANNUR.

TRUST US WITH YOUR NEEDS

OUR REF : AR/QTN/128-2023


DATE : 15/11/2023

SLNO:	DESCRIPTION	TOTAL(INCL.ALL TAXES)
01	AEROCARE NEBULIZER COMPRESSOR WITH PISTON	2000/-
	TOTAL	2,000/-

TERMS & CONDITIONS OF SALE

Payment Terms : Complete Payment Within 30 Days

Delivery : Within 4 Days from the date of purchase order .



AR MEDICARE

TRUST US WITH YOUR NEEDS