



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 143291

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>PHC PARIYARAM</u></p> <p>Address <u>KANNUR</u></p> <p>Ph : <u>9497602192</u></p>	<p>Call Registration Date : <u>11-10-2023</u></p> <p>Caller ID : <u>74146</u></p> <p>Date of Visit : <u>12-10-2023</u></p> <p>Asset No. : <u>1341610</u></p> <p>EQPT Name : <u>BP APPARATUS</u></p> <p>Manufacture <u>MEDVISION</u> Model : <u>1018</u></p> <p>S. No. <u>N/A</u> Dept. <u>MAIN CENTER</u></p>
--	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Found main board, display and bladder in defective condition

**Action Taken :** Checked the equipment and found main board, display and bladder in defective condition. Found equipment was installed on 21-08-2019 and aged up to 4 years. While comparing the life span used, the repairing cost and replacement cost of spares are beyond economic. Hence we recommending for condemnation.

Completed  Date : 12-10-23 Time : 11 am Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Jineshkumar PP</u>	<u>12-10-23</u>	<u>9am</u>	<u>11am</u>

**Customer Remark** Completed  Pending

<p><b>Service Engineer Name :</b> <u>Jineshkumar P P</u></p> <p><b>Signature :</b> </p> <p><b>Date :</b> <u>12-10-23</u></p> <p><b>Contact Number :</b> <u>9895766435</u></p>	<p><b>Customer Name :</b> <u>Dr. Jithu</u></p> <p><b>Signature :</b> </p> <p><b>Date :</b> <u>12/10/2023</u></p> <p><b>Contact Number :</b> <u>9895766435</u></p> <p><b>Designation :</b> <u>for Medical Officer</u></p> <p><b>Hospital Seal :</b> </p>
---	---




**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	PHC PARIYARAM
3	Name of Equipment with Make, Model and Serial Number	NAME: BP APPARATUS MODEL: 1018 MAKE: MEDVISION SERIAL NUMBER: NOT AVAILABLE
4	Equipment ID & Barcode	1341610
5	Date of purchase / Year of manufacture / Installation Date	21-08-2019
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC / CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	11-10-2023
9	Action taken	Checked the equipment and found the repairing cost and spare replacement are beyond economical. Equipment aged upto 4 years. so recommending for construction.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	—


13	Asset Value	2200/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	97%.
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	SERVICE REPORT ATTACHED
16	Reasons for recommending the equipment as BER	Checked the equipment and found the main board, display and bladder in defective condition. Equipment was installed on 21-03-2019 and aged up to 4 years. While comparing the lifespan used and the repairing cost and replacement of spares are beyond economical. Hence we recommending for condemnation.
17	Name & Signature of CYRIX Authority	JINESH KUMAR P P 

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that main board issue, bladder defective. aged up to 4 years. The repairing cost economically not feasible. So recommended as per BER protocol.

  
20/10/2023  
Signature of JC BM (NHM)

CONSULTANT IN BIOMEDICAL  
NATIONAL HEALTH MISSION  
KANNUR - 870002

Date \_\_\_\_\_



  
for Dr. Ranjith  
Signature of \_\_\_\_\_  
MEDICAL OFFICER  
PRIMARY HEALTH CENTRE  
KANNUR





AR MEDICARE

## AR MEDICARE

ROOM NO: 865, 2<sup>ND</sup> FLOOR  
AL-RAHA ARCADE, THAZHE CHOVVA  
KANNUR-670018

EMAIL: [armedicare07@gmail.com](mailto:armedicare07@gmail.com)

MOB: 7994654108, 9446057350

OUR REF : AR/QTN/110-2023

DATE : 12/10/2023

TO  
CYRIX HEALTHCARE  
ERNAKULAM

Respected Sir,

**SUB: QUOTATION FOR DIGITAL BP APPARATUS PARTS**

Kindly note with reference to the discussions had with yourself regarding the above mentioned subject, we are pleased to submit our very special competitive offer for your reference and further perusal.

**In case of any further clarifications kindly feel free to contact us.**

Thanking you and assuring you of our best attention and engineering services at all times.

Waiting for your earliest reply.

Yours truly,

For AR MEDICARE,  
KANNUR.



**TRUST US WITH YOUR NEEDS**

OUR REF : AR/QTN/110-2023  
DATE : 12/10/2023

SLNO:	DESCRIPTION	TOTAL(INCL.ALL TAXES)
01	MEDVISION DIGITAL BP APPARATUS MAINBOARD DISPLAY BLADDER	1200/- 800/- 150/-
	TOTAL	2,150/-

**TERMS & CONDITIONS OF SALE**

**Payment Terms** : Complete Payment Within 30 Days

**Delivery** : Within 4 Days from the date of purchase order .



AR MEDICARE

TRUST US WITH YOUR NEEDS