



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 151505

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>GNH Thalassery</u></p> <p>Address <u>Thalassery</u> <u>Kannur</u></p> <p>Ph : <u>9947868404</u></p>	<p>Call Registration Date : <u>11-10-23</u></p> <p>Caller ID : <u>74229</u></p> <p>Date of Visit : <u>12-10-23</u></p> <p>Asset No. : <u>1311736</u></p> <p>EQPT Name : <u>Suction apparatus</u></p> <p>Manufacture <u>Supreme</u> Model : <u>supremepius</u></p> <p>S. No. <u>NA</u> Dept. <u>MICU</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Machine not working.

Action Taken : checked the machine and found that suction motor and suction bottle are damaged. This machine was installed on 25/9/2018 and aged up to 5 years. repairing cost and replacement of spairs are beyond economic so we recommend this machine for condemnation.

Completed Date : 12-10-23 Time : 3:30pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>Rasilraj</u>	<u>12-10-23</u>	<u>2:30pm</u>	<u>3:30pm</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Rasilraj</u></p> <p>Signature : <u>Rasilraj</u></p> <p>Date : <u>12-10-23</u></p> <p>Contact Number : <u>9995521436</u></p>	<p>Customer Name : <u>Pilly Faaran</u></p> <p>Signature : <u>Pilly Faaran</u></p> <p>Date : <u>12/10/23</u></p> <p>Contact Number : <u>Number office Ci-I</u></p> <p>Designation :</p> <p>Hospital Seal :</p>
---	---



10/10/23
Superintendent
GENERAL HOSPITAL
THALASSERY - 670 101

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl No	Particulars	Details
1	Name of District	Kannur
2	Name of Hospital	GM Thalassery
3	Name of Equipment with Make, Model and Serial Number	Suction apparatus Make: Supreme Model: Supreme plus sno : NA.
4	Equipment ID/ Barcode	1311736
5	Date of purchase/ Year of manufacture/Installation Date	25/9/2018
6	Warranty details	Nil
7	*AMC/ CAMC Period agreed at the time of purchase	Nil
8	Date of breakdown/ Date of registration of complaint	11-10-23
9	Action taken	Motor checked and found motor complaint.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Details of service required (Recommendations for repair)	Suction motor and suction bottle are complaint.
12	Cost of spares (specify parts and cost)	—

13	Asset Value	10,200
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	75%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	CYRIX report.
16	Reasons for recommending the equipment as BER	checked the machine and found suction bottle and suction motor are complained. This machine was installed on 25-9-18 and aged upto 5 years. replacement of spairs are beyond economic. so we recommend this machine for condemnation
17	Signature of CYRIX Authority	Rasibraj <i>Rasibraj</i>

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that suction motor, bottle ~~motor~~ are defective. aged up to 5 years. installed on 25/9/18. its service cost economically not feasible as per guidelines. so recommended for BER as per BER protocol.

Rasibraj Signature of JC BM (NHM)

CONSULTANT BIOMEDICAL
NATIONAL HEALTH MISSION
KANNUR - 670002

Pritty Jeevan
Pritty

Date 12/10/23



[Signature]
16/10/23
Superintendent
GENERAL HOSPITAL
THALASSERY - 670 101
Signature of
Superintendent Medical Officer (i/c)



Sreeraj Surgical

Aiswarya Commercial Complex, Nr. Ashirvad Hospital, Thavakkara, Kannur - 670 002
Ph: 0497 2706307. Mob: 9349846307, 9387529848
Email: sreerajsurgical@gmail.com, rajeshsreerajsurgical@gmail.com
www.sreerajsurgical.com Follow us on  www.facebook/ sreerajsurgical

Date:

12-10-2023

QUOTATION

TO,

CYRIX HEALTHCRAE
THRIPOONITHARA
ERNAKULAM

SIR,

WE ARE VERY PLEASURE TO ENQUIRY OF RATE OF FOLLOWING ITEMS.

S.L.NO	ITEMS	AMOUNT
1	SUCTION MOTOR -SUPREME PLUS	6800.00
2	SUCTION BOTTLE -SUPREME PLUS	750.00

GST INCLUDED

WE EXPECT YOUR VALUABLE ORDERS AND KIND CO.OPERATION WITH US.

THANKING YOU.



Authorised Dealer:

