



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 145340

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrrix.in

<p>Health Facility ... <u>CHC MATTANNUR.</u></p> <p>Address ... <u>Mattannur.</u> <u>KANNUR.</u></p> <p>Ph : <u>9400683145.</u></p>	<p>Call Registration Date : <u>31/10/2023.</u></p> <p>Caller ID : <u>77550.</u></p> <p>Date of Visit : <u>04/11/2023.</u></p> <p>Asset No. : <u>1330311</u></p> <p>EQPT Name : <u>Centrifuge.</u></p> <p>Manufacture <u>REMI.</u> Model : <u>R-4C.</u></p> <p>S. No. <u>NA</u> Dept. <u>Laboratory.</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working.

Action Taken : checked the equipment and found the Motor, Rheostate and Transformer are defective, equipment is installed on 24/04/2012 and aged upto more than 11 years. Comparing the life span used. The repairing and replacement cost are beyond economical. So we recommend for condemnation as per KMSCL Tender clause 5.3.14.1

Completed Date : 04/11/2023. Time : 11:00 AM. Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrlix Engineer	Date	Start Time	End Time
<u>Sajith</u>	<u>04/11/2023.</u>	<u>9:00 am</u>	<u>11:00 AM.</u>

Customer Remark Completed Pending

Service Engineer Name : Sajith

Signature : [Signature]

Date : 04/11/2023

Contact Number : 8606345195

Customer Name : COMMUNITY HEALTH CENTRE

Signature : [Signature]

Date : 4.11.2023

Contact Number : 9400683145

Designation : LT GN I

Hospital Seal :



**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No.	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	CHC MATTANNUR
3	Name of Equipment with Make, Model and Serial Number	CENTRIFUGE REMI R-4C
4	Equipment ID/ Barcode	1330311
5	Date of purchase/ Year of manufacture/Installation Date	24/04/2012
6	Warranty details	No
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC.
8	Date of breakdown/ Date of registration of complaint	31/10/2023.
9	Action taken	checked the equipment and - Found Motor, Rheostat and Transformer are defective.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Details of service required (Recommendations for repair)	Not Recommended.
12	Cost of spares (specify parts and cost)	—

13	Asset Value	9651-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	128.3%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Service report attached.
16	Reasons for recommending the equipment as BER	Checked the equipment and found the Motor, Rheostat and Transformer are defective. equipment is installed on 24/11/2012 and aged upto more than 11 years. considering the lifespan used the repairing and replacement cost are beyond economical. So we recommend for condemnation as per KMSCL Tender - clause 5.8.14.1
17	Signature of CYRIX Authority	Sayanth 04/11/2023

*Not mandatory #Based on the period of life and value as per the BER guidelines

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that motor, transformer, and rheostat are in defective. and aged upto 11 years. its service cost economically not feasible. so recommended for BER. as per the quotation. *Perth*
Signature of JC BM (NHM)
10/11/2023

CONSULTANT BIOMEDICAL
NATIONAL HEALTH MISSION
KANNUR - 670002

Shyja M.P, *Sh*
Lablechnician Gr I

Date 4.11.2023



Shyja
MEDICAL OFFICER
COMMUNITY HEALTH CENTRE
KANNUR
Superintendent Medical Officer (I/c)



Kochin Inlab Equipments India Pvt Ltd

KI/RE/254/23-24

Dated :03.11.2023

To,
Cyrix Healthcare Pvt Ltd
Pononithura,Ernakulam -682038

Sub: Repair Estimate for R4C Centrifuge

With reference to the above, we are giving our estimate as per the request.

Sl.No	Description	Rate	Qty	GST	Amount
1	Electric Motor	6,000.00	1.00	18.00%	7,080
2	Transformer	2,000.00	1.00	18.00%	2,360
3	RHEOSTATE	2,500.00	1.00	18.00%	2,950
				G.Total:	12,390

Rupees Twelve Thousand Three Hundred And Ninty Only

Kindly confirm the above estimate .

- 1 .Payment : 100% Advance**
- 2.Mention GST No: in the purchase orders.**
- 3.Delivery will be made in 4 to 5 weeks from the date of order acceptance by the company service Department**

Thanking you,

Yours faithfully,

For Kochin Inlab Equipments India Pvt Ltd



Authorized Signatory

Door No.28/2391-D,Vihaan,Chilavannur,Kadavanthara.P.O, Kochi-682020
Tel - 0484-2316938, CIN: U52311KL2008PTC022135 Email:info@inlabinstruments.com,

ENCARE

ENCARE

acer

Preventive Maintenance Schedule kept
02/11/19
04

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7889
BAR CODE - (8004-890615225) 1330311

WARNING

1. READ OPERATING INSTRUCTIONS BEFORE USE
2. PROVIDE PROPER EARTHING
3. FOR VIBRATION FREE PERFORMANCE
 - a. ENSURE LEVEL AND STABILITY
 - b. BALANCE CENTRIFUGE TUBES
 - c. ENSURE USE OF RUBBER CUSHIONS FOR GLASS TUBES
4. BRING SPEED KNOB TO 'OFF' AND INCREASE SPEED GRADUALLY
5. CHECK CARBON BRUSHES PERIODICALLY
6. CHECK GLASS FUSE FOR NON-WORKING OF UNIT.

REMI R-4C

MIN

OFF

RPM

MAINS



