



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 145342

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrilx.com | E-mail : bemp.kl@cyrilx.in

<p>Health Facility <u>CHC MATTANNUR</u></p> <p>Address <u>Mathannur</u> <u>KANNUR</u></p> <p>Ph : <u>9400 68345</u></p>	<p>Call Registration Date : <u>31/10/2023</u></p> <p>Caller ID : <u>77552</u></p> <p>Date of Visit : <u>04/11/2023</u></p> <p>Asset No. : <u>1330313</u></p> <p>EQPT Name : <u>UPS</u></p> <p>Manufacture <u>BPE</u> Model : <u>NA</u></p> <p>S. No. <u>M02L17B00152 Dept. Laboratory</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : checked the equipment and found the Main board, Transformer and batteries are damaged. Equipment was installed on 01/04/2016 and aged upto more than 7 years. Comparing the lifespan used repairing and replacement cost are beyond economical. So we recommend for condemnation as per KMSCL Tender clause 5.3.14.1

Completed Date : 04/11/2023 Time : 1.00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrilx Engineer	Date	Start Time	End Time
	<u>04/11/2023</u>	<u>11.00am</u>	<u>1.00pm</u>

Customer Remark Pending

<p>Service Engineer Name : <u>Sayan</u></p> <p>Signature : </p> <p>Date : <u>04/11/2023</u></p> <p>Contact Number : <u>8606345195</u></p>	<p>Customer Name : <u>Shyja H.P</u></p> <p>Signature : </p> <p>Date : <u>4.11.2023</u></p> <p>Contact Number : <u>9400683145</u></p> <p>Designation : <u>LT CW I</u></p> <p>Hospital Seal :</p>
---	---




Pending
MEDICAL OFFICER
COMMUNITY HEALTH CENTRE
MATTANNUR

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)


PROFORMA

Sl No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	CHC MATTANNUR
3	Name of Equipment with Make, Model and Serial Number	UPS. BPE. Sn: M02L17B00152
4	Equipment ID/ Barcode	1330313.
5	Date of purchase/ Year of manufacture/Installation Date	01/04/2016.
6	Warranty details	No.
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown/ Date of registration of complaint	31/10/2023.
9	Action taken	checked the equipment and - Found Mainboard, Transformer Batteries are damaged.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Details of service required (Recommendations for repair)	Not recommended.
12	Cost of spares (specify parts and cost)	—

13	Asset Value	4500-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	113.3 %
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Service report attached.
16	Reasons for recommending the equipment as BER	checked the equipment and found the mainboard, Transformer and batteries are damaged. Equipment was installed on 01/01/2016 aged upto more than 7 years. comparing the lifespan used repairing and replacement cost are beyond economical. So we recommend for condemnation as per KNSCL Tender clause 5.3.4.1
17	Signature of CYRIX Authority	Sayasth  04/11/2023.

*Not mandatory #Based on the period of life and value as per the BER guidelines

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
checked & identified that main board, Transformer and batteries are damaged, aged up to 7 years and its service cost economically not feasible so recommended for BER as per BER protocol.

10/11/2023.
CONSULTANT BIOMEDICAL
NATIONAL HEALTH MISSION
KANNUR - 670002
Signature of JC BM (NHM)

Shuja M.P
Lab Technicians in I
Stn
Date 4.11.2023


MEDICAL OFFICER
COMMUNITY HEALTH CENTRE
KANNUR
Sub-Intendent Medical Officer (I/c)

IT WORLD

COMPUTER SALES & SERVICE

1st Floor., Malabar Plaza, Near SBT ATM
MATTANUR.

Mob: 9744041752, Office: 8089136301

No. **104**

Date 11/11/2023.....

QUOTATION

To CYRix Healthcare Pvt Ltd
Pononithura, Ennakulam - 682038

Mob No..... Email :.....


Sl. No.	DESCRIPTION	Qty	Unit	Price
1	main board	1		1600
2	Transformer	1		1300
3	Battery	2		2200
			Total	5100



For IT WORLD

Signature

On-Line UPS



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-690615225) 1330313

