


248

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Pathanamthitta
2	Name of Hospital	PHC Poamadam
3	Name of Equipment with Make, Model and Serial Number	Equipment : Pharmacy fridge Make : LG Model : Intello cool Serial no : N/A
4	Equipment ID & Barcode	0340980
5	Date of purchase / Year of manufacture /Installation Date	09/04/2011
6	Warranty details (Yes/No)	NO
7	*AMC/ CAMC Period agreed at the time of purchase	N/A
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	07/10/2023
9	Action taken	Recommended for BER
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	N/A
12	Cost of spares (specify parts and cost)	N/A


13	Asset Value	RS : 26494
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	YES
16	Reasons for recommending the equipment as BER	Fully Rusted
17	Name & Signature of CYRIX Authority	Shiju John Chacko 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Physically verified on 13/11/2023. Recommended for BER.


 16/11/2023
 Signature of JC BM (NHM)

Date 9/11/23.


 Seal


 Signature of Superintendent / Medical Officer (i/c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 159610

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>PHC PRAMADAM</u></p> <p>Address <u>PRAMADAM</u> <u>PATHANAMHITTA</u></p> <p>Ph : <u>9072807164</u></p>	<p>Call Registration Date : <u>07/11/2023</u></p> <p>Caller ID : <u>78878</u></p> <p>Date of Visit : <u>9/11/2023</u></p> <p>Asset No. : <u>0340980</u></p> <p>EQPT Name : <u>Pharmacy Refrigerator</u></p> <p>Manufacture <u>LG</u> Model : <u>Intello Cool</u></p> <p>S. No. <u>NA</u> Dept. <u>Observation Rooms</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Fully Rusted

Action Taken : Checked the machine. found that machine is in fully damaged and rusted condition.
Recommended for BER.

Completed Date : 9/11/23 Time : 1:00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <u>-NA-</u>	<u>-NA-</u>	<u>-NA-</u>	<u>-NA-</u>
2. <u>-NA-</u>	<u>-NA-</u>	<u>-NA-</u>	<u>-NA-</u>
3. <u>-NA-</u>	<u>-NA-</u>	<u>-NA-</u>	<u>-NA-</u>


Cyrix Engineer	Date	Start Time	End Time
<u>Shiju John Chacko</u>	<u>9/11/23</u>	<u>10:00am</u>	<u>1:00pm</u>

Customer Remark Completed Pending

Fully rusted

Service Engineer Name Shiju John Chacko
Signature :
Date : 9/11/23
Contact Number : 9605381275

Customer Name : Appanna ps.
Signature :
Date : 9/11/23
Contact Number : 9605381275
Designation : Assistant
Hospital Seal : 04682306524





**NATIONAL HEALTH MISSION
UNDER
MAINTENANCE PROGRAMME
BIOMEDICAL EQUIPMENT**



No.: 125610

CYRIX
HEALTHCARE PVT LTD

SERVICE PROVIDER
Tender No. WO-21024-202389

ISO 13284 : 2012 & ISO 9001:2008 CERTIFIED COMPANY / IASB Approved Service Agency
30847 B, Park Junction, Poonjithur, Kochi - 682 028, Kerala
PH : 92472 98500 Website : www.cyrrix.com | E-mail : enquiry@cyrrix.in

Service Report

Health Facility: P.H.C. PRAMADAM	Call Registration Date: 07/11/2023
Address: PRAMADAM	Caller ID: 78878
PR: PATNAMAMHITTA	Date of Visit: 9/11/2023
PR: POTSOJITHA	Asset No.: 0340980
	EQPT Name: Pharmacy Refrigerator
	Manufacturer: LG
	Model: WJL1000
S No.: NA	Dept. Classification: NA

Service Classification: Breakdown Call PMS Calibration Cust. Training

Problem Identified:

Fully Broken

Action Taken: Checked the machine. Found that machine is in fully damaged and worst condition. Recommended for BER.

Completed Date: 9/11/23 Time: 1:00pm Spare Required

Description	Qty	Part Number	PR Number
1. NA	-	NA	NA
2. NA	-	NA	NA
3.			

Cyrix Engineer: Srinivas Chakra
Date: 9/11/23
Start Time: 10:00am
End Time: 1:00pm

Customer Remark: Fully over
 Completed Pending

Service Engineer Name: Srinivas Chakra
Signature: [Signature]
Date: 9/11/23
Contact Number: 982381232
Hospital Seal: [Seal]
Destination: [Address]
Contact Number: [Number]



PATHAN

Fridge

7

Date		Receipt	Issue	Balance	Initials
9/4/2011	Refrigerator LG model Wellocool rate 26,494	1		1	[Signature]
9/4/2011	Issued to Injection room		1	Nil	



MEDICAL OFFICER
P.H. CENTRE
PRAMADAM
V. KOTTAYAM P.O.
PATHANAMTHITTA

