


243

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

| Sl. No | Particulars | Details |
|--------|--|------------------------------------|
| 1 | Name of District | Pathanamthitta, |
| 2 | Name of Hospital | Phc Panalalam thekkakara. |
| 3 | Name of Equipment with Make, Model and Serial Number | sterilizer (Cocher type Autoclave) |
| 4 | Equipment ID & Barcode | ID: 78196, Barcode: 0340836 |
| 5 | Date of purchase / Year of manufacture / Installation Date | 22/03/2017 |
| 6 | Warranty details (Yes/No) | No Warranty |
| 7 | *AMC/ CAMC Period agreed at the time of purchase | N/A |
| 8 | Date of breakdown (Date of registration of complaint through email/ Toll free) | 03/11/2023 |
| 9 | Action taken | Recommended for BER |
| 10 | Present status of the equipment (Fully damaged / partially damaged) | Fully damaged. |
| 11 | Recommendations for repair (required service details) | N/A |
| 12 | Cost of spares (specify parts and cost) | N/A |

| | | |
|----|--|---|
| 13 | Asset Value | RS : 15000/- |
| 14 | # Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value | Na, |
| 15 | Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) | Yes |
| 16 | Reasons for recommending the equipment as BER | Fully damaged. |
| 17 | Name & Signature of CYRIX Authority | Sijo Zachariah  |

*Not mandatory #Based on the period of life and value as per the BER guidelines

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Physically verified on 13/11/2023. Recommend for BER

Slydo
16/11/2023

Signature of JC BM (NHM)

CONSULTANT/BIO MEDICAL
NATIONAL HEALTH MISSION
PATHANAMTHITTA

Date



Seal

Signature of Superintendent / Medical Officer (i/c)



PHC
Pandalam Thekkekara
Ph: 04734 - 223517
mail: phcpandalamthekke@smm...



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 127102

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

| | |
|--|--|
| <p>Health Facility <u>DHC Pandalam Thekku</u></p> <p>Address <u>Thayyil, Pandalam thekku kare</u> <u>Pothanmthitta, Pin : 691525</u></p> <p>Ph : <u>9745 2759 05</u></p> | <p>Call Registration Date : <u>03/11/2023</u></p> <p>Caller ID : <u>78196</u></p> <p>Date of Visit : <u>08/11/2023</u></p> <p>Asset No. : <u>0340836</u></p> <p>EQPT Name : <u>Sterilizer (Cooker type Autoclave)</u></p> <p>Manufacture <u>Na</u> Model : <u>Na</u></p> <p>S. No. <u>Na</u> Dept. <u>Palliative</u></p> |
|--|--|

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working
Checked the machine found that interior of the Autoclave and Pressure gauges are fully rusted and damaged

Action Taken : Checked the machine properly found that the interior of the autoclave fully rusted and burned out the internal parts coil and pressure gauge fully damaged not repairable. Recommended to B.E.R.

Completed Date : 08/11/2023 Time : 10:05 am Spare Required

Spare Replaced Requested

| Description | Qty. | Part Number | PR Number |
|--------------|-----------|-------------|-----------|
| 1. | | | |
| 2. <u>Na</u> | <u>Na</u> | <u>Na</u> | <u>Na</u> |
| 3. | | | |

| Cyrix Engineer | Date | Start Time | End Time |
|-----------------------|-------------------|-----------------|-----------------|
| <u>Sijo Zachariah</u> | <u>08/11/2023</u> | <u>09:00 am</u> | <u>10:05 am</u> |

Customer Remark Completed Pending

| | |
|--|---|
| <p>Service Engineer Name : <u>Sijo Zachariah</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>08/11/2023</u></p> <p>Contact Number : <u>81189 5422</u></p> | <p>Customer Name : <u>[Signature]</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>[Signature]</u></p> <p>Contact Number : <u>[Signature]</u></p> <p>Designation : <u>[Signature]</u></p> <p>Hospital Seal : <u>[Signature]</u></p> |
|--|---|



[Signature]
Dr. Asha Sreed
Palliative Care Unit, Pandalam Thekku
Ph: 04734 - 223617
E-mail: shouandalamthekku@gmail.com



NATIONAL HEALTH MISSION
UNDER
MAINTENANCE PROGRAMME
BIO-MEDICAL EQUIPMENT



No. : 127102

CYRIX
 HEALTHCARE PVT LTD

SERVICE PROVIDER
 Tender No. WO-371021-2021088

ISO 13485 : 2013 & ISO 9001:2008 CERTIFIED COMPANY AERB Approved Service Agency

20184 I B, Patta Junction, Poonithura, Kochi - 682 038, Kerala
 Ph : 98473 98500 Website : www.cyrix.com | E-mail : beng@cyrix.in

Service Report

Health Facility : The Government Medical College Hospital, Poonithura, Kochi
 Address : The Government Medical College Hospital, Poonithura, Kochi
 Ph : 98473 98500

Call Registration Date : 03/11/2023
 Dealer ID : F8194
 Date of Visit : 08/11/2023
 Asset No. : 0340836
 EQPT Name : 2.0 litre (Copper Tube Analar)
 Manufacturer : No. Model : No.
 S. No. : No. Dept. : No.

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Checked the machine found that pressure of the subcircuit and pressure gauges are fully tested and changed.

Action Taken : Checked the machine properly found that the pressure of the subcircuit was fully damaged and replaced. Checked and pressure gauges fully damaged and replaced.

Completed Date : 08/11/2023 Time : 10:30 AM Spare Required

Spare Replaced Replaced

| Description | Qty | Part Number | PR Number |
|-------------|-----|-------------|-----------|
| 1 | No | No | No |
| 2 | No | No | No |
| 3 | No | No | No |

| Cyrix Engineer | Date | Start Time | End Time |
|-----------------|------------|------------|----------|
| Dr. S. S. S. S. | 08/11/2023 | 10:30 AM | 12:00 PM |

Customer Remark : Completed Pending

Service Engineer Name : Dr. S. S. S. S.
 Signature :
 Date : 08/11/2023
 Contact Number : 98473 98500
 Hospital Seal :

Customer Name :
 Signature :
 Date :
 Contact Number :
 Hospital Seal :

Phc Pandalam thekkarg
Sterilizer
Call no: 78 196
Barcode : 0340836



From,

Medical Officer I/c

P.H.C Pandalam Thekkela

To,

CYRIX

Sir,

ചന്ദ്രൻ തമ്പുരാൻ പ്രൊഫഷണൽ ഹെൽത്ത് സർവീസ് ഉപയോഗിച്ചുള്ള തെളിവുകൾ ഉൾപ്പെടെയുള്ള Condensation ഉപയോഗിച്ച് രോഗങ്ങൾ നശിപ്പിക്കുന്നതിന് അപേക്ഷിക്കുന്നു.

Sl. No. Items Date of Receipt Remarks

| | | | |
|----|----------------------|------------|--|
| 1. | Autoclave | 22/3/2017 | |
| 2. | Stabilizer | 08/12/1997 | |
| 3. | Needle Destroyer | 9/4/2012 | |
| 4. | Microscope Binocular | 31/8/2019 | found excess on store verification in 2019 |
| 5. | Microscope Binocular | 23/3/2006 | |
| 6. | Calorimeter | 25/3/2006 | |

3/11/2023

ചന്ദ്രൻ തമ്പുരാൻ



എന്ന,

ഓഫീസ്, ഹെൽത്ത് സർവീസ്
 ഹെൽത്ത് സർവീസ്, ഹെൽത്ത് സർവീസ്
 ഹെൽത്ത് സർവീസ്, ഹെൽത്ത് സർവീസ്
 Ph: 04734 - 212617

