


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	Pathanamthitta
2	Name of Hospital	Phc Pandalam Thekkelkara
3	Name of Equipment with Make, Model and Serial Number	Laboratory microscope
4	Equipment ID & Barcode	ID: 78185, Barcode: 0340867
5	Date of purchase / Year of manufacture / Installation Date	08/12/1997
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	N/A
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	08/11/2023
9	Action taken	Recommended for BER
10	Present status of the equipment (Fully damaged / partially damaged)	fully damaged
11	Recommendations for repair (required service details)	N/A
12	Cost of spares (specify parts and cost)	N/A

13	Asset Value	RS : 15400/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Na
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Yes
16	Reasons for recommending the equipment as BER	Fully rusted
17	Name & Signature of CYRIX Authority	Sijo Zachariah 

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Physically verified on 13.11.2023. Recommended for BER

  
16/11/2023

Signature of JC BM (NHM)

CONSULTANT (BIOMEDICAL)  
NATIONAL HEALTH MISSION  
PANDALAM

Date



Signature of   
Superintendent / Medical Officer (i/c)

പണ്ടലം ആരോഗ്യ കേന്ദ്രം  
പണ്ടലം, തേക്കേകര  
ഫോൺ: 222617  
ഇമെയിൽ: phcpandalam@nhm.kerala.gov.in



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. : 127103**

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>PHC Pandalam Thekkekar</u></p> <p>Address <u>that day, Pandalam thekkekar</u> <u>Pathanamthitta, Pin: 691525</u></p> <p>Ph : <u>9745275905</u></p>	<p>Call Registration Date : <u>03/11/2023</u></p> <p>Caller ID : <u>78185</u></p> <p>Date of Visit : <u>08/11/2023</u></p> <p>Asset No. : <u>0340867</u></p> <p>EQPT Name : <u>Laboratory microscope</u></p> <p>Manufacture <u>Na</u> Model : <u>Na</u></p> <p>S. No. <u>Na</u> Dept. <u>LAB</u></p>
--	--

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** not working  
on checking found that the microscope fully damaged all  
the parts rusted.

**Action Taken :** checked the machine found that the all parts of  
the microscope, Objective's eye piece, stage, light source and  
all other parts are fully damaged with out machine not  
Repairable Recommended to B&R.

Completed  Date : 08/11/2023 Time : 11:30 am Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>Na</u>	<u>Na</u>	<u>Na</u>	<u>Na</u>
3.			

<b>Cyrix Engineer</b>	Date	Start Time	End Time
<u>Sijo Zachariah</u>	<u>08/11/2023</u>	<u>10:10 am</u>	<u>11:30 am</u>

**Customer Remark** Completed  Pending

**Service Engineer Name :** Sijo Zachariah  
**Signature :** [Signature]  
**Date :** 08/11/2023  
**Contact Number :** 811395422

**Customer Name :** [Signature]  
**Signature :** [Signature]  
**Date :** [Signature]  
**Contact Number :** [Signature]  
**Designation :** [Signature]  
**Hospital Seal :** [Signature]



Pandalam Thekkekar PHC  
 Poonithura, Kochi - 682 038  
 Phone : 98472 99500  
 Email : phspandalamthekkekar@gmail.com



**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



No: 127103

**CYRIX**  
HEALTH CARE PVT LTD

SERVICE PROVIDER  
Tender No. WC-37/2021-2022/888

ISO 13485 : 2013 & ISO 9001:2015 CERTIFIED COMPANY AERB Approved Service Agency

Service Report  
PH : 98472 88800 Website : www.cyrrix.com E-mail : [info@cyrix.in](mailto:info@cyrix.in)

Call Registration Date : 03/11/2023	Health Facility : P.H. Padalambal, Thalassery
Caller ID : F 2 182	Address : P.H. Padalambal, Thalassery
Date of Visit : 08/11/2023	Ph : 9742952902
Asset No. : 031086F	
EQT Name : Laboratory Autoclave	
Manufacturer : Hira	
Model : Hira	
S.No. : No	
Dept. : LAB	

Service Classification : Breakdown Call  PMS  Calibration  Cust.Training

Problem Identified : Autoclave not working properly. On checking found that the inner scope fully damaged. It was replaced.

Action Taken : Checked the machine found that the all parts of the autoclave. Checked the inner scope and replaced it with a new one. Report submitted to B&B.

Completed  Date : 08/11/2023 Time : 11:30 AM Spare Required

Description	Qty	Part Number	PR Number
NA	NA	X16	NA



Cyrix Engineer	Date	Start Time	End Time
S. Lakshmi	08/11/2023	10:10 AM	11:30 AM

Customer Remark  Completed  Pending

Service Engineer Name : S. Lakshmi	Customer Name :
Signature : [Signature]	Signature :
Date : 08/11/2023	Date :
Contact Number : 98472 88800	Contact Number :
Designation :	Designation :
Hospital Seat :	Hospital Seat :



Phc pandalam thekke kara  
Call ID : 78185  
Barcode : 0340867  
Laboratory microscope.





From,

Medical Officer/c

P.H.C Pandalam Thekkkara

To,

CYRIX

Sir,

ചന്ദ്രൻ തമ്പുരാൻ ചികിത്സാ കേന്ദ്രത്തിൽ ഉപയോഗിച്ചിരുന്ന താഴെ പറയുന്ന ഉപകരണങ്ങൾ Condensation ചെയ്യുന്നതിന് അനുയോജ്യമായ സാധന വിതരണത്തിന് അപേക്ഷിക്കുന്നു.

Sl. No.	Items	Date of Receipt	Remarks
1.	Autoclave	22/3/2017	
2.	Sterilizer	08/12/1997	
3.	Needle Destroyer	9/4/2012	
4.	Microscope Binocular	31/8/2019	found excess on store verification in 2019
5.	Microscope Binocular	23/3/2006	
6.	Calorimeter	25/3/2006	

3/11/2023

ചന്ദ്രൻ തമ്പുരാൻ



എന്ന,

*[Handwritten Signature]*

ഓഫീസ്, ചന്ദ്രൻ തമ്പുരാൻ  
 ചികിത്സാ കേന്ദ്രം, പാണ്ടലം തേക്കുപുരം  
 ഫോൺ: 04734 - 223317

