

247

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Kottayam
2	Name of Hospital	Phc Thalavada
3	Name of Equipment with Make, Model and Serial Number	Pharmacy Refrigerator Make: whirlpool Model: Genius 168L SN: NA.
4	Equipment ID & Barcode	0543507
5	Date of purchase / Year of manufacture / Installation Date	02-10-2007
6	Warranty details (Yes/No)	No
7	*AMC/ CAMC Period agreed at the time of purchase	No
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	31-10-2023, #77581
9	Action taken	Repaired. Replaced damaged delay. Compressor not working
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	Required a new compressor. Gas Filling required after new compressor installation. Needs to be service freezer unit.
12	Cost of spares (specify parts and cost)	Copper Strainer Filter - 150 Compressor delay set - 250 Compressor - 3500 Gas Filling charge - 2500 Freezer unit - 1500 Total spare cost: 8500/-

13	Asset Value	9700
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	87.6 %
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix
16	Reasons for recommending the equipment as BER	Machine was very old & Spare cost is very high According to clause 5.3.14.1, requesting for condemnation.
17	Name & Signature of CYRIX Authority	Nidhish Roy <u>Nds</u>

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value - 9700
 Cost of spare - 8500 (87.6%)
 Life period - 16 years
 As per the RBER guidelines recommended as RBER

[Signature]
 Signature of JC BM (NHM)
 CONSULTANT IN CHARGE
 NATIONAL HEALTH MISSION
 KUTTIYAM

Date 13/11/25

[Seal]
 SEAL
 DEPT. OF KERALA HEALTH SERVICES
 THALANADU
 PIN - 686 580
 HEALTH CENTER

[Signature]
 MEDICAL OFFICER IN CHARGE
 PRIMARY HEALTH CENTRE
 THALANADU P O
 KUTTIYAM DIST
 Signature of
 Superintendent / Medical Officer (i/c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 160661

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.ki@cyrix.in

<p>Health Facility ...Ph.c...Thalanadu</p> <p>Address Thalanadu</p> <p>..... Kottayam</p> <p>Ph : 9496424656</p>	<p>Call Registration Date : ...31-10-2023.....</p> <p>Caller ID : 77581</p> <p>Date of Visit : 02-11-2023</p> <p>Asset No. : 0543507</p> <p>EQPT Name : Pharmacy Refrigerator</p> <p>Manufacture Whirlpool Model : Genius 168L</p> <p>S. No. NR Dept. Pharmacy</p>
------------------------------------------------------------------------------------------------------------------------------------------------	-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not cooling due to compressor defective

Action Taken : Checked. Replaced damaged relay. In more inspection, compressor defective. Required New compressor, gas filling, service of freezer etc. Not repairable economically. So recommended for BER

Completed Date : 13-11-2023 Time : 10:30am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. Compressor relay	1		
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Nidhish Roy	13-11-2023	09:17am	10:40am

Customer Remark Completed Pending

<p>Service Engineer Name : Nidhish Roy</p> <p>Signature : <i>Nidhish Roy</i></p> <p>Date : 13-11-2023</p> <p>Contact Number : 9072575296</p>	<p>Customer Name : DR Prithvi Chackambor</p> <p>Signature : <i>Prithvi Chackambor</i></p> <p>Date : 13/11/2023</p> <p>Contact Number : 9946085094</p> <p>Designation : Medical Officer</p> <p>Hospital Seal : </p>
--------------------------------------------------------------------------------------------------------------------------------------------------------------------------	--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------



**NATIONAL HEALTH MISSION
UNDER
MAINTENANCE PROGRAMME
BIOMEDICAL EQUIPMENT**



No. : 180661

CYRIX

HEALTH CARE PVT LTD
SERVICE PROVIDER
Toll Free No: 1800 232 232

180 13088 : 2012 & 180 1001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

2014 B, Patta Junction, Tondiarpet, Kochi - 682 038, Kerala
PH : 98473 98500 Website : www.cyrinx.com | E-mail : help.k@cyrinx.in

Call Registration Date : 18-11-2023	Health Facility : Pbc Thalavada
Caller ID : 77521	Address : Thalavada
Date of Visit : 09-11-2023	City : Kottayam
Asset No. : 0245807	Pin : P A P A A A 5 4 5 6
EQPT Name : Pharmacy Refrigerator	
Manufacturer : Pharmacia Model : Pharmacia	
S. No. : 1719 Dept. Pharmacy	

Service Classification : Breakdown Call FMS Calibration Cust. Training

Problem Identified : Not cooling due to compressor defective

Action Taken : Checked. Replaced damaged relay. In more inspection compressor defective. Replaced New compressor gas filling, service of freezer etc. Not repairable economically. go recommendation for RFP

Completed Date : 13-11-2023 Time : 10:30am
Spare Required Requested

Description	Qty	Part Number	PR Number
1. Compressor relay	1		
2.			
3.			

Customer Remark : Completed Pending

Service Engineer Name : Michael Roy
Signature : *[Signature]*
Date : 13-11-2023
Contact Number : 98473 2962

Customer Name : DR P R WILSON
Signature : *[Signature]*
Date : *[Date]*
Contact Number : *[Number]*
Hospital Name : *[Name]*
Hospital Gen : *[Gen]*





