

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	Kottayam
2	Name of Hospital	Phc Erattupetta
3	Name of Equipment with Make, Model and Serial Number	Weighing Machine Make: Sanso Model: NA SN: NA
4	Equipment ID & Barcode	0541842
5	Date of purchase / Year of manufacture / Installation Date	17-10-2018
6	Warranty details (Yes/No)	No.
7	*AMC/ CAMC Period agreed at the time of purchase	No.
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	04-10-2023
9	Action taken	Checked. Tried serviced of board. Not working.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	Problems related to Display control board. Board is irreparable. Required new control board with display.
12	Cost of spares (specify parts and cost)	Spare not available

13	Asset Value	850
14	*Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Spare not available
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix
16	Reasons for recommending the equipment as BER	Machine aged upto 4 years and spares not available. So recommended for BER.
17	Name & Signature of CYRIX Authority	Nidhish Roy <u>Nidhish</u>

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

asset value - 850  
 cost of spare = not available  
 life period - 5 year  
 As per RBER guide line and spare not available. Recommended as RBER

[Signature]  
 16/10/23  
 Signature of JC BM (NHM)

CONSULTANT BIOMEDICAL ENGINEER  
 NATIONAL HEALTH MISSION  
 KOTTAYAM

Date

Seal

Signature of  
 Superintendent / Medical Officer (i/c)

Medical officer i/c  
 DIC Erattipetta





**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

**No. :** 144698

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility Phc Erattupetta	Call Registration Date : 04-10-2023
Address Erattupetta	Caller ID : 92718
Kottayam	Date of Visit : 07-10-2023
Ph : 9961077652	Asset No. : 0541842
	EQPT Name : Weighing Machine
	Manufacture Samsa Model : NA
	S. No. NA Dept. OP

**Service Classification :** Breakdown Call  PMS  Calibration  Cust. Training

**Problem Identified :** Battery cover damaged, slot rusted, Always showing Error on display, Sometimes display not working.

**Action Taken :** Checked. Problems related to display control board. Board unserviceable. Required new display with control board. Economically not repairable. So recommended for BEIR.

Completed  Date : 07.10.2023 Time : 11:00am

Spare Replaced  Requested

Spare Required

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer

Nidhish Roy

Date

Start Time

End Time

07.10.2023

09:30am

11:00am

Customer Remark

Completed

Pending

Service Engineer Name : Nidhish Roy  
Signature : *Nidhish Roy*

Date : 07-10-2023

Contact Number : 9092575296

Customer Name :

Signature :

Date :

Contact Number :

Designation :

Hospital Seal :

*Dr. Resmi P. Sas*

Medical officer I/C

PHC Erattupetto





**NATIONAL HEALTH MISSION  
UNDER  
MAINTENANCE PROGRAMME  
BIOMEDICAL EQUIPMENT**



No.: 14458

**CYRIX**  
HEALTHCARE PVT LTD

SERVICE PROVIDER  
Tender No. WO-27027-202009

ISO 13485 : 2013 & ISO 9001:2008 CERTIFIED COMPANY | AERB Approved Service Agency  
30/04 E, Patta Junction, Poonjar, Kochi - 682 038, Kerala  
Ph: 98472 0950 Website: www.cyrix.com | E-mail: hcm@cyrix.in

**Service Report**

Call Registration Date : 04-10-2023	Health Facility Phc Eattupetta
Caller ID : P 2418	Address Eattupetta
Date of Visit : 07-10-2023	Kottayam
Asset No. : 0211849	Pin : 686103
EQUIPMENT Name : Copilypod Machine	
Manufacture Model : M19	
S. No. M19 Dept. OP	

Service Classification : Breakdown Call  PMS  Calibration  Cost Training

Problem Identified : Battery cover damaged, slot empty  
Always showed Error on display, sometimes display not working.

Action Taken : Checked. Problems related to display control board. Board was replaced. Reducted not repairable. So recommended for RFR.

Completed  Date : 07-10-2023 Time : 11:00 AM  
Spare Replaced  Requested  Spare Required

Sl. No.	Description	Qty	Part Number	PR Number
1.				
2.				
3.				




Cyx Engineer		
Date	Start Time	End Time
07-10-2023	09:30 AM	11:00 AM

Customer Remarks  Completed  Pending

Service Engineer Name : Michael Ray	Signature :
Date : 07-10-2023	Contact Number : 984720950
Customer Name :	Signature :
Date :	Contact Number :



Medical Officer /  
Phc Eattupetta





DIGITAL WEIGHING MACHINE

42 ARDRAM - 1KML

17/10/18  
Main stock  
part no: 675  
1/11/18

Purchasing date 1:10:18

Exp. dt. - 12/2028

122ml to precheckup room

Balance - 1  
1/11/18

Recd

2 2 1

1

1

