



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 155236

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrilx.com | E-mail : bemp.kl@cyrilx.in

<p>Health Facility <u>PHC kankol Alapadamba</u></p> <p>Address <u>kannur</u></p> <p>Ph : <u>6282624302</u></p>	<p>Call Registration Date : <u>04/12/2023</u></p> <p>Caller ID : <u>83384</u></p> <p>Date of Visit : <u>05/12/2023</u></p> <p>Asset No. : <u>1341575</u></p> <p>EQPT Name : <u>Weighing machine</u></p> <p>Manufacture <u>SAMSO</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>OP</u></p>
--	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** found the machine is not switching on

**Action Taken :** checked the machine and found that the machine is switching 'on'  
Found main board, load cells are damaged  
Enguized jam spaces.

Completed  Date : 13/12/23 Time : 11:00AM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrilx Engineer	Date	Start Time	End Time
<u>Arun V</u>	<u>13/12/23</u>	<u>10:30AM</u>	<u>11:00AM</u>

**Customer Remark** Completed  Pending

<p>Service Engineer Name : <u>Arun V</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>13/12/2023</u></p> <p>Contact Number : <u>8547338287</u></p>	<p>Customer Name : <u>Dr. Omkaranth P</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>13/12/23</u></p> <p>Contact Number : <u>9744057824</u></p> <p>Designation : <u>MEDICAL OFFICER</u></p> <p>Hospital Seal : <u>PRIMARY HEALTH CENTRE</u></p>
--	---




MEDICAL OFFICER  
PRIMARY HEALTH CENTRE  
KANKOL-ALAPADAMBA  
KANNUR DIST

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	KANNUR
2	Name of Hospital	PHC KANKOL ALAPADAMBA
3	Name of Equipment with Make, Model and Serial Number	WEIGHING MACHINE Make :SAMSO Model:N/A SN:N/A
4	Equipment ID/ Barcode	1341575 #83784
5	Date of purchase/ Year of manufacture/Installation Date	09-06-2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 04-12-2023
9	Action taken	Checked the machine and found that Machine is not switching on, Found mainboard, Load cells damaged. Enquired for spares .
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	

13	Asset Value	1600/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that machine is not switching ON. Found that main board and load cells damaged. Equipment installed on 09-06-2019 aged up to 4 years and 5 months. Checked the spares in market, but spares are not available because it's a low asset cost equipment. So we recommending for condemnation.
17	Name & Signature of CYRIX Authority	ARUN V 

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph



Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that main board & load cells damaged. Aged up to 4 years. Spare not available. So recommended for BER as per BER protocol.


  
28/12/2023

CONSULTANT  
NATIONAL  
KANNUR

Signature of JC BM (NHM)

The above mentioned weighing machine was in use for around 4 years and the benefits has been completely ~~and~~  

Date: 13/12/23.



Signature of  
Superintendent/Medical Officer (I/c)

MEDICAL OFFICER  
PRIMARY HEALTH CENTRE  
KANKOL-ALAPADAMBA  
KANNUR DIST



SUBCENTRE KIT STOCK.

Date	IN.No	Received from / Issued to	Receipt	Issue	Balance	Remarks
6/9/19	IN No- are	1033 the following items received from Kanjya depot Thissur dated 28/08/19.				
	1	BP Apparatus (digital)	1		1	Rs- 1119.19
	2	Dust Bin - Black (swing type)	1		1	Rs- 194.91
	3	Dust Bin - Red ( " " )	1		1	Rs- 194.91
	4	Dust Bin - Yellow ( " " )	1		1	Rs- 194.91
	5	Examination Torch (led)	1		1	Rs- 78.13
	6	Footscope (Schramm make)	1		1	Rs- 33.06
	7	Gloves sterile 7 (Surgical)	25		25	Rs- 262.25
	8	Glucaneter strips	300		300	Rs- 2145.00
	9	Glucaneter	1		1	Rs- 312.60
	10	Haemometer Test Strips	1		1	Rs- 9000
	11	Haemometer (True HB)	1		1	Rs- 1850.00
	12	Hub cutter (Syringe & needle cutter)	1		1	Rs- 174.55
	13	Infantometer (krups)	1		1	Rs- 1581.25
	14	Kit for Testing Residual chlorine in Drinking water	1		1	Rs- 828.37
	15	Lancet 1's Con-call)	100		100	Rs- 90.00
	16	Measuring Tape	1		1	Rs- 19.49
	17	Nebulizer	1		1	Rs- 1232.13
	18	Oxygen cylinder 40 FCT B type	1		1	Rs- 5325.00
	19	Oxygen cylinder Spanner	1		1	Rs- 166.75
	20	Oxygen cylinder Trolley	1		1	Rs- 1413.35
	21	Oxygen flow meter with Humidifying Bottle.	1		1	Rs- 1179.23
	22	Salt Tasting kit for Iodine	5		5	Rs- 109.00
	23	Stadiometer - Life care	1		1	Rs- 467.80
	24	Stethoscope	1		1	Rs- 308.04
	25	Thermometer digital	1		1	Rs- 67.76
	26	Urine Strip (Glucose & Albumin)	9		9	Rs- 924.12
	27	weighing machine (digital- Paediatric)	1		1	Rs- 4239.44
	28	weighing machine (digital - Adult)	1		1	Rs- 1608.04
6/9/19		All the above items are issued to JPHN.				