

108

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**


Recommendations for Beyond Economic Repair (BER)

PROFORMA


Sl. No	Particulars	Details
1	Name of District	Kottayam
2	Name of Hospital	Phc Vizhikathodu
3	Name of Equipment with Make, Model and Serial Number	BP Apparatus Make: Lifecare Model: BSX516 SN: NA.
4	Equipment ID & Barcode	0543629
5	Date of purchase / Year of manufacture / Installation Date	19-12-2000
6	Warranty details (Yes/No)	No
7	*AMC/ CAMC Period agreed at the time of purchase	No
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	10-10-2023
9	Action taken	Checked. Serviced display
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	Required new display unit
12	Cost of spares (specify parts and cost)	1000 Internal PCB - 1000 Mo box 250 <u>1250/-</u>

Date 11/10/2023

Seal




Signature of
Superintendent / Medical Officer (t/c)


 Signature of Superintendent / Medical Officer (t/c)

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset Value - 1500
 Cost of spare - 1250 (83.33%)
 - 2 year 11 months life period
 - 2 year 11 months life period
 Cost of spare - 1250 (83.33%)
 Life period - 2 year 11 months
 Life period - 2 year 11 months


 Signature of JC BM (NHM)

NATIONAL HEALTH SERVICE
 KARNATAKA GOVT HEALTH SERVICES
 BANGALORE

* Attach Photograph

*Not mandatory #Based on the period of life and value as per the BER guidelines

13	Asset Value	1500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	83.33%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix
16	Reasons for recommending the equipment as BER	Display not repairable. Spare parts not available. So recommended for BER.
17	Name & Signature of CYRIX Authority	Nidhish Roy



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 144708

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility Phc Vizhikkathodu	Call Registration Date : 10-10-2023
Address Vizhikkathodu	Caller ID : 73904
Kottayam	Date of Visit : 10-10-2023
Ph : 9961238991	Asset No. : 0543609
	EQPT Name : BP Apparatus
	Manufacture Life care Model : B5X516
	S. No. NA Dept. OP

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Display not working

Action Taken : Checked. Display Faulty issue. Unreparable display unit. Economically not repairable. So recommended for BER.

Completed **Date :** 10-10-2023 **Time :** 03:40pm **Spare Required**

Spare Replaced **Requested**

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			


Cyrix Engineer	Date	Start Time	End Time
Nidhish Roy	10-10-2023	10:40pm	03:30pm

Customer Remark Completed Pending

Service Engineer Name : Nidhish Roy
Signature :
Date : 10-10-2023
Contact Number : 9072575296

Customer Name : Dr. Sreedhar Sivaradas Viswanathan
Signature :
Date : 11/10/2023
Contact Number : 9633756098
Designation : Medical Officer
Hospital Seal :



Contact number : ८०५२२२०८६
 Date : १०-१०-२०२३
 Signature : 
 Service Engineer name : वि.वि. शर्मा

Mobile no :
 Designation :
 Contact number :
 Date : 10/10/2023
 Signature :
 Customer name :

Customer Name

Completed

Pending

वि.वि. शर्मा

10-10-2023

10-10-2023

03-30PM

SLX Engineer

Date

Start Time

End Time

3

2

1

Description

SN

PR Number

PR Number

Service Request

Completed

Completed Date : 10-10-2023 Time : 03:30PM

Service Request

नोट : उपरोक्त कार्य के लिए सर्विस एंजिनियर द्वारा उपरोक्त स्थान पर जाकर निम्नलिखित कार्य किए गए हैं।
 Action Taken : डिवाइन का कार्य किया गया।

Problem Identified :

डिवाइन का कार्य किया गया।

Service Classification : Breakdown call BMS Calibration Servicing

PR : ८०५२२२०८६

SN : 1111111111

Manufacturer : BEA

EDBT Name : वि.वि. शर्मा

Address : वि.वि. शर्मा

Asset No : ०२२२०८

Date of Visit : १०-१०-२०२३

Health Facility : वि.वि. शर्मा

Called : २२२२

Call Registration Date : १०-१०-२०२३

Service Report PR : 88475 88200 Mobile : www.slx.com | E-mail : slx@slx.com

HEALTH CARE SLX LTD

SLX

SERVICE PROVIDER

NO : 14708



**NATIONAL HEALTH MISSION
 UNDER
 MAINTENANCE PROGRAMME
 BIOMEDICAL EQUIPMENT**



BP Apparatus

Date	Received	Issued	Balance	Remarks
14-12-14	Received from the store	1	1	vsh
18-7-19	Received from the store (LED display)	2	3	amp
19/12/2020	Received from the store (digital)	2	5	2
	(1) Mercury BP apparatus Ekometer (-7/21/14) cl'd to unserviceable instruments register page no. 7. Condemned as per order no. H99 PTCV dtd 20/12/21			
18/11/19	2 LED BP apparatus → hardware equipment register page no. 6	1	4	A
19/12/20	2 digital BP apparatus → ms equipment & instruments register			26

