

109

REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Idukki
2	Name of Hospital	THLPH Neelambalam
3	Name of Equipment with Make, Model and Serial Number	Refridgerator Vitecon, S-332 SN- 9812052601310038
4	Equipment ID/ Barcode	0620715
5	Date of purchase/ Year of manufacture	5/4/2006
6	Warranty details	NONE
7	*AMC/ CAMC Period agreed at the time of purchase	NO
8	Date of breakdown/ Date of registration of complaint	7/10/23
9	Action taken	Trying to Service
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	Checked the machine OBM-said that the freezer is defective and Refilled gas and welded the
12	Cost of spares (specify parts and cost)	



SUPERINTENDENT
 THE HOSPITAL
 NEDUMKANDAM
 IDUKKI-685553

Signature of Superintendent/Medical Officer (A)

Handwritten signature
 21/10/23

Date 21/10/23

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
 Consultant Biomedical Engineer
 National Health Mission
 Alappakalam
 Tel: 0471-685603
 Signature of JC BM (NHM)
Handwritten signature
 11/11/23

*** Attach Photograph**

*Not mandatory #Based on the period of life and value as per the BER guidelines

13	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	
14	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	
15	Reasons for recommending the equipment as BER	<p> <i>Handwritten text:</i> checked the machine the machine is very old. the reason spare is very defective. need to repair the gas and working parts are very costly for maintenance so recommend for beyond economic repair </p>
16	Signature of CYRIX Authority	<p> <i>Handwritten signature</i> Vinayak </p>



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 146195

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>T.H.O M</u>	Call Registration Date : <u>7/10/23</u>
Address <u>Nedumkandam</u>	Caller ID : <u>73534</u>
..... <u>Nedumkandam, Dukk</u>	Date of Visit : <u>20/10/23</u>
Ph : <u>9446103775</u>	Asset No. : <u>0620715</u>
	EQPT Name : <u>Refridgrator</u>
	Manufacture <u>Videacon</u> Model : <u>n11</u>
	S. No. <u>332</u> Dept. <u>lab</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : not working, No cooling

Action Taken : Checked the machine OEM - said that it's very old machine, the freezer is chiller type need to be replaced and gas welding and gas filling. We are recommend for Beyond Economical Repair

Completed Date : 20/10/23 Time : 4pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Vineesh Vincent</u>	<u>20/10/23</u>	<u>02 pm</u>	<u>7 pm</u>

Customer Remark Completed Pending

Service Engineer Name : Vineesh Vincent
Signature : Vineesh
Date : 20/10/23
Contact Number : 7593649124

Customer Name : Nedumkandam, Kochi
Signature : [Signature]
Date : 20/10/23
Contact Number : 98472 99500
Designation : Superintendent
Hospital Seal : T.H.O HOSPITAL NEDUMKANDAM DUKKI 685555



20/10/23
21/10/23



Handwritten notes at the top right corner of the form.

Contract Number: 223845154
 Date: 06/10/03
 Location: J. D. HOSPITAL SUPERINTENDENT
 Date: 06/10/03
 Signature: [Signature]
 Contract Name: [Name]

Description		FR number	Part Number	Qty
1				
2				
3				

Completed: Date: 06/10/03 Time: 5:00 AM
 Date required: Description: [Blank]
 Remarks: [Blank]

Problem identified: [Blank]

<p>Service Classification: Breakdown call <input checked="" type="checkbox"/> PM <input type="checkbox"/> Call repair <input type="checkbox"/> Call training <input type="checkbox"/></p> <p>Health Facility: J. D. H.</p> <p>Category: [Blank]</p> <p>Requester: [Blank]</p> <p>Approved by: [Blank]</p> <p>Date: 06/10/03</p> <p>Call ID: 223845154</p> <p>Call registered date: 06/10/03</p>	<p>Equipment Name: [Blank]</p> <p>Model: [Blank]</p> <p>Serial Number: [Blank]</p> <p>Part Number: [Blank]</p> <p>Quantity: [Blank]</p> <p>Order Number: [Blank]</p>
---	--

Service Report: [Blank]

HEALTH CARE PARTNER

CYBIX

NO: 223845154

TECHNICAL SUPPORT CENTER

THE NATIONAL HEALTH MISSION UNDER MAINTENANCE PROGRAMME BIOMEDICAL EQUIPMENT

EW S.N. REFRIGERATION
O. JUNCTION, ADIMALI

ph: 9447294560


SERVICE CENTRE

Date... 20/10/23...

To,
Cyrix Health Care
Ernakulam.

Dear Sir,

The refrigerator of THQH Medumk
andam is very old and the spare parts
(Evaporator) is not available in market.
So the refrigerator is not repairable.



Thanks

May

