


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Tripvandrum
2	Name of Hospital	WEC Thaycaud
3	Name of Equipment with Make, Model and Serial Number	Pharmacy Refrigerator Videocon
4	Equipment ID & Barcode	0114937
5	Date of purchase / Year of manufacture / Installation Date	24/3/2010
6	Warranty details (Yes/No)	No
7	*AMC/ CAMC Period agreed at the time of purchase	Nil
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	18/9/23
9	Action taken	Checked and found leg stand broken, Compressor damaged gasket damaged. Machine is rusted
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Checked and found leg Stand broken, Compressor and gasket damaged Machine is rusted
12	Cost of spares (specify parts and cost)	Not repairable economically

13	Asset Value	19,990/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Not repairable
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Attached
16	Reasons for recommending the equipment as BER	Checked and found log stand damaged, gaslet damaged, compressor not working and internal leak.
17	Name & Signature of CYRIX Authority	

*Not mandatory **#Based on the period of life and value as per the BER guidelines**


* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Compressor Damage.

So RBER.


MANEESHA MOHAN M.R.
 Junior Consultant (Biomedical).
 National Health Mission
 Thiruvananthapuram


 10-10-2023

Signature of JC BM (NHM)

Date




 Signature of Superintendent / Medical Officer (JC)

SUPERINTENDENT
 Govt. W & C Hospital, Thycad
 Thiruvananthapuram - 14



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 146794

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>WELC Thaycaud</u></p> <p>Address : <u>Trivandrum</u></p> <p>Ph : <u>9400771713</u></p>	<p>Call Registration Date : <u>18/9/23</u></p> <p>Caller ID : <u>70035</u></p> <p>Date of Visit : <u>20/9/23</u></p> <p>Asset No. : <u>0114937</u></p> <p>EQPT Name : <u>Pharmacy Refrigerator</u></p> <p>Manufacture : <u>Videcon</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>Store</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : Checked and found Leg stand broken door gasket damaged. External leak. Compressor not working. Machine is rusted. Hence recommend the machine to BER.

Completed Date : 20/9/23 Time : 5pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Salika S</u>	<u>20/9/23</u>	<u>4pm</u>	<u>5pm</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Salika S</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>20/9/23</u></p> <p>Contact Number : <u>6282656760</u></p>	<p>Customer Name : <u>Bindu M P</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>20/9/23</u></p> <p>Contact Number : <u>9400771713</u></p> <p>Designation : <u>Sub Insp</u></p> <p>Hospital Seal : </p>
---	---



SERVICE REPORT

Dear Sir / Madam,

Date-05/10/23

This is in reference with our engineer's inspection on your defunct Refrigerator, Make: Videocon, Model: Nil, Sl. No. Nil, Barcode-0114937 / Caller ID-70035 – W&C Thycaud, Tvm

The refrigerator base is very weak as it is very old and identified compressor failure and internal leakage, also found beading damaged.

Hence, this refrigerator cannot be repaired and we suggest that it can be condemned.

We shall submit the condemnation certificate if required from your end.

Thanks and regards

Yours faithfully,

Precise Equipment's & Services



Authorized Signatory

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP) 

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)
0114937 

