


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Thiruvananthapuram
2	Name of Hospital	CHH Thiruvananthapuram
3	Name of Equipment with Make, Model and Serial Number	Autoclave - Coolbox type Shiva
4	Equipment ID & Barcode	0110682.
5	Date of purchase / Year of manufacture / Installation Date	14/03/09
6	Warranty details (Yes/No)	No
7	*AMC/ CAMC Period agreed at the time of purchase	No
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	03/10/2023
9	Action taken	Checked the machine and found Bearings on the machine
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Machine cannot be repaired Small bearings found on the machine due to obsolescence.
12	Cost of spares (specify parts and cost)	No



13	Asset Value	Nil
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Nil
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Please find the attached copy of Cyrix Service report along with the RBER Protocols
16	Reasons for recommending the equipment as BER	Small leakages found on the machine due to obsolescence. This leakage cannot be repaired so leakage cannot be averted.
17	Name & Signature of CYRIX Authority	Mohamed Sajeed M <i>[Signature]</i>

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Machine too old and breakages in body
So RBER

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram

[Signature]
10.11.2023

Signature of JC BM (NHM)

Date _____



Signature of *[Signature]*
 Superintendent / Medical Officer (i/c)
**Superintendent
General Hospital
Thiruvananthapuram**



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 140827

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>CH Thiruvananthapuram</u></p> <p>Address : <u>CH Junction</u> <u>Thiruvananthapuram</u></p> <p>Ph : <u>8590512545</u></p>	<p>Call Registration Date : <u>08/10/2023</u></p> <p>Caller ID : <u>72589</u></p> <p>Date of Visit : <u>06/10/2023</u></p> <p>Asset No. : <u>0110682</u></p> <p>EQPT Name : <u>Autoclave - Pressure type</u></p> <p>Manufacture : <u>Shive</u> Model : <u>N/A</u></p> <p>S. No. : <u>N/A</u> Dept. : <u>Dental</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :
Checked the machine and found small leakages in the machine.

Action Taken :
Small leakages found on the machine. Cannot close arrest the leakages of the machine. So she recommending for condemnation.

Completed Date : 06/10/2023 Time : 12:00 Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.				
2.	<u>RBER</u>	<u>N/A</u>	<u>N/A</u>	<u>N/A</u>
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Mohamed Sajid M</u>	<u>06/10/2023</u>	<u>11:00</u>	<u>12:00</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Mohamed Sajid M</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>06/10/2023</u></p> <p>Contact Number : <u>8293499127</u></p>	<p>Customer Name : <u>[Signature]</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>06/10/2023</u></p> <p>Contact Number : <u>8590512545</u></p> <p>Designation : <u>Dental Consultant</u></p> <p>Hospital Seal : <u>[Seal]</u></p>
---	--









CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800- 425-7669

BAR CODE - 0110682



(4 14)09(264)01(2 1)0038