


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Trivandrum
2	Name of Hospital	GH Neyyethikkoni
3	Name of Equipment with Make, Model and Serial Number	nebulizer philips
4	Equipment ID & Barcode	75604 0113220
5	Date of purchase / Year of manufacture /Installation Date	13/11/2014.
6	Warranty details (Yes/No)	no
7	*AMC/ CAMC Period agreed at the time of purchase	no
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	14.10.2023
9	Action taken	motor & Piston damaged
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	checked the machine problem found in motor and piston damaged
12	Cost of spares (specify parts and cost)	NA



13	Asset Value	1353/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Not repairable
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Had the attached copy of Cyrix Service report along with RBER Report.
16	Reasons for recommending the equipment as BER	Motor and piston damaged hence recommending for condemnation.
17	Name & Signature of CYRIX Authority	Ajith George K.S

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Motor and piston damaged. Hence RBER

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram

MS
13.11.2023

Signature of JC BM (NHM)

Date

Seal

Signature of Superintendent / Medical Officer (i/c)

GENERAL HOSPITAL
NEYYATTINKARA

SUPERINTENDENT
GENERAL HOSPITAL
NEYYATTINKARA-895121



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 154389

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <i>SH Neyyattinkara</i>	Call Registration Date : <i>19.10.2023</i>
Address <i>P21 Vennalambalam</i>	Caller ID : <i>75604</i>
Ph : <i>7907169162</i>	Date of Visit : <i>28.9.2023</i>
	Asset No. : <i>013220</i>
	EQPT Name : <i>Nebuliser</i>
	Manufacture <i>Philips</i> Model : <i>NB</i>
	S. No. <i>NA</i> Dept. <i>C. ward</i>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : *not working*

Action Taken : *Checked the machine problem found that motor and piston damaged. So its recommended for beyond economical repair (BER).*

Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time

Customer Remark Completed Pending

Service Engineer Name : <i>Asif Seethi</i>	Customer Name : <i>Asha S B</i>
Signature : <i>A</i>	Signature : <i>Asha S B</i>
Date : <i>28.10.2023</i>	Date : <i>28/10/23</i>
Contact Number : <i>7593847111</i>	Contact Number : <i>9895204395</i>
	Designation : <i>Senior Nsg-officer</i>
	Hospital Seal : <i>General Hospital Neyyattinkara</i>

SUPERINTENDENT GENERAL HOSPITAL NEYYATTINKARA-695124

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7669
BAR CODE - (8004-890615225) 0113220

REF 1130529 Philips Home Nebulizer
SN 120364 230V ~ 50Hz 0.5A
LOT 1210-CNMU 30 min / 30 min O
(01)06946159503013 IP21
(10)1210-CNMU CE 0123
(21)120364
Respironics Respiratory Drug Delivery (UK) Ltd
Chichester Business Park, City Fields Way
Tangmere, Chichester PO20 2FT UK
Made in China 1130536 R00



PHILIPS

Home Nebulizer

