


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA


Sl. No	Particulars	Details
1	Name of District	Tiruvardur.
2	Name of Hospital	CIPHC Kankrammeda.
3	Name of Equipment with Make, Model and Serial Number	Nebulizer, Aub Health Care Sate town, 08 2018.
4	Equipment ID & Barcode	76035 - 0141214.
5	Date of purchase / Year of manufacture /Installation Date	20/4/2018.
6	Warranty details (Yes/No)	no
7	*AMC/ CAMC Period agreed at the time of purchase	NT.
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	21/10/23.
9	Action taken	checked the machine
10	Present status of the equipment (Fully damaged / partially damaged)	fully damaged.
11	Recommendations for repair (required service details)	checked and found that the machine was affected by flood on 15/10/23. Water went inside the machine and not found now.
12	Cost of spares (specify parts and cost)	It can't be repair economically.


13	Asset Value	1353/—
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Not repairable
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Attached.
16	Reasons for recommending the equipment as BER	checked the machine - Motor and piston fully damaged. It can't be repair economically, therefore proposing it for RBER.
17	Name & Signature of CYRIX Authority	Panav. P.S. 

*Not mandatory #Based on the period of life and value as per the BER guidelines


* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
 Machine damaged. it is fully damaged by the blood on 15-10-2023.
 So RBER.




 1-11-2023
 Signature of JC BM (NHM)

Date _____ Seal _____




 Signature of Superintendent / Medical Officer (i/c)

MEDICAL OFFICER
 UPHC, KANNANMOOLA, TVPM



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 158389

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>CIPAK</u></p> <p>Address <u>Kannammook</u> <u>Tarandrum</u></p> <p>Ph : <u>9444 62 1084</u></p>	<p>Call Registration Date : <u>21/10/23</u></p> <p>Caller ID : <u>76035</u></p> <p>Date of Visit : <u>30/10/23</u></p> <p>Asset No. : <u>0141214</u></p> <p>EQPT Name : <u>NEBUUSER</u></p> <p>Manufacture <u>Dob Health</u> Model <u>Patapub</u> <u>care</u></p> <p>S. No. <u>0272018</u> Dept. <u>OP</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : not working

Action Taken : checked and found the machine is affected by flood on 18/10/23. water went inside the machine and not functional now. Motor, piston and tubings are fully damaged and not in repairable condition. Therefore properly

Completed Date : Time : 10:20am Spare Required


Spare Replaced Requested

1.	Description	Qty.	Part Number	PR Number
2.	—	—	—	—
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>prasan</u>		<u>10 am</u>	<u>10:30 am</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Prasan AS</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>2</u></p> <p>Contact Number : <u>9094572652</u></p>	<p>Customer Name : <u>Dr. Ayya P. Peri</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>30/10/23</u></p> <p>Contact Number : <u>828 1441 822</u></p> <p>Designation : <u>Medical Officer</u></p> <p>Hospital Seal :</p>
--	---



21/10/2023



SAFE Air **TOUCH**
Slide 

Compressor Nebulizer

Model : Safe Touch
Input : 230V / 50 Hz
Power : 80 VA
Serial No. : 082018

Product by

AnB Health Care Product

124, Mettu Street, Ayanavaram, Chennai - 600 023
Tamil Nadu, India. Ph : 0091-44-43554646
www.anbhealthcare.com - anbhealthcare@yahoo.co.in

Customer Care # +91 766 766 7464

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)
0141214



HEALTHCARE PVT LTD
CYRIX
CN: UPHCKP3 S.N:
PRO Nebulizer
ID : 0141214
CAL.ON : 14104123
CAL.DUE : 18104124
ISO 9001 : 2008 Certified Co Ph: 9744555073

