


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	Thiruvananthapuram.
2	Name of Hospital	FHC Poozhura
3	Name of Equipment with Make, Model and Serial Number	Colorimeter Leba IroniU LS-111 2019-04-220-
4	Equipment ID & Barcode	79462 8 0130075
5	Date of purchase / Year of manufacture /Installation Date	10.04.2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	10.11.2023.
9	Action taken	Try to service
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	checked the equipment. found transformer and display board complete
12	Cost of spares (specify parts and cost)	Transformer - 750/- Display board - 2,800/-



13	Asset Value	4900/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	72%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Attached
16	Reasons for recommending the equipment as BER	Checked the equipment. found Display board and transformer complete. its not economically viable. so recommending the equipment as BER
17	Name & Signature of CYRIX Authority	 16.11.2023.


\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* **Attach Photograph**


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Display board and Transformer damaged.  
So RBER.


**MANEESHA MOHAN M.R.**  
Junior Consultant (Biomedical)  
National Health Mission  
Thiruvananthapuram

  
9-12-2023

Signature of JC BM (NHM)

Date		Signature of Superintendent / Medical Officer (i/c)
------	---	--

  
Signature of  
Superintendent / Medical Officer (i/c)

  
Community Health Centre





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 154481

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>Community Health Center</u></p> <p>Address <u>Poonithura, Thiruvananthapuram</u></p> <p>Ph : <u>9495824079</u></p>	<p>Call Registration Date : <u>10.11.2023</u></p> <p>Caller ID : <u>79462</u></p> <p>Date of Visit : <u>16.11.2023</u></p> <p>Asset No. : <u>0130075</u></p> <p>EQPT Name : <u>Colorimeter</u></p> <p>Manufacture <u>Lebatronics</u> Model : <u>25111</u></p> <p>S. No. <u>2019-04-220</u> Dept. <u>Laboratory</u></p>
--	--

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Not working

**Action Taken :** Checked the equipment found equipment transformer and display board complaint. its spare cost more than equipment value. So recommending equipment as BER.


Completed  Date : 16.11.2023 Time : 3.00pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1. <u>Transformer</u>	<u>1no</u>		
2. <u>Display board</u>	<u>1no</u>		
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>Mohammed Shams</u>	<u>16.11.2023</u>	<u>1.00pm</u>	<u>3.00pm</u>

**Customer Remark** Completed  Pending



<p>Service Engineer Name : <u>Mohammed Shams</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>16.11.2023</u></p> <p>Contact Number : <u>7994042445</u></p>	<p>Customer Name : <u>R.S. Mohamed Shams</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>16.11.2023</u></p> <p>Contact Number : <u>9495824079</u></p> <p>Designation : <u>Senior Lab Technician</u></p> <p>Hospital Seal : <u>[Signature]</u></p>
--	--





**AR MEDICARE**  
ROOM NO: 865, 2<sup>ND</sup> FLOOR  
AL-RAHA ARCADE, THAZHE CHOVVA  
KANNUR-670018  
EMAIL: [armedicare07@gmail.com](mailto:armedicare07@gmail.com)  
MOB: 7994654108, 9446057350

**OUR REF : AR/QTN/131-2023**  
**DATE : 04/12/2023**

**TO**  
**CYRIX HEALTHCARE**  
**ERNAKULAM**

Respected Sir,

**SUB: QUOTATION FOR CALORIMETER SPARE PARTS**

Kindly note with reference to the discussions had with yourself regarding the above mentioned subject, we are pleased to submit our very special competitive offer for your reference and further perusal.

SLNO:	DESCRIPTION	RATE (INCL. ALL TAXES)
01	DISPLAY BOARD	2,800/-
02	TRANSFORMER	750/-
<b>TOTAL</b>		<b>3,550/-</b>

**In case of any further clarifications kindly feel free to contact us.**

Thanking you and assuring you of our best attention and engineering services at all times.  
Waiting for your earliest reply.

**TERMS & CONDITIONS OF SALE**

**Payment Terms** : Complete Payment Within 30 Days  
**Delivery** : Within 4 Days from the date of purchase order .


Yours truly,

For AR MEDICARE,  
KANNUR.



**TRUST US WITH YOUR NEEDS**





- 40 Violet
- 42 Violet Blue
- 48 Greenish Blue
- 50 Blue Green
- 52 Bluish Green
- 54 Green
- 62 Yellow
- 68 Red

**CAUTION**  
 Once turned off, wait for more than 12 sec. before operation

**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMS (ENR)  
 TOLL FREE NO: 1800 - 425-7669  
 BAR CODE - (8004-890615225) 0130075

**LABTRONICS**

ISO 9001 - 2002 CERTIFIED



ABS    %T    C<sub>1</sub>    C<sub>2</sub>    C<sub>3</sub>  
 ●    ●    ●    ●    ●  
 STD. CONC.



**MICROPROCESSOR PHOTOCOLORIMETER**  
 LT - 111



52

40  
42  
48  
50  
52  
54  
62  
68

**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)  
TOLL FREE NO: 1800 - 425-7669  
BAR CODE - (8004-890615225)  
0130075



MANUFACTURED BY  
**LABTRONIC**  
PANCHKULA (INDIA)  
Model No. [redacted]  
S.No. 2019-04-22



STOCK BOOK (MEDICINES)

Name of drug			Name of drug					
Date	Receipt	Issue	Balance	Date	Receipt	Issue	Balance	Remarks
	kg./gm./ml.	kg./gm./ml.	kg./gm./ml.		kg./gm./ml.	kg./gm./ml.	kg./gm./ml.	
		Colourimeter.						
21/3/19		monochrome series.						
		8 Filter Digital.			1	1	NIL	(Cdn)
12/4/19	colorimeter							
	model & make: LTH/Labtronics			1		2	(Damp)	
	Project name - Aadram.							
	Supplied by KMSCL.							
	SQ: No: 2019-04-220							
	(1) condemned and write off as per order							
	no 03/276/2022/che (dt 9/1/22)							
	Balance → 1 No.							
10/1/23	c/o to page no. (10) District as							
	Furniture SQ Vol. II							