


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Trivandrum
2	Name of Hospital	GH Trivandrum
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : Multipara Monitor make : GE Model : CareScape 8650 SN : SEW12131784HA
4	Equipment ID & Barcode	0111053
5	Date of purchase / Year of manufacture /Installation Date	No purchase 09/2012 (Manufacturer)
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	22/11/2023
9	Action taken	Checked monitor and damaged for Monitor Cabinet, Display, Main board, Parameter board, ECG Female connectors.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged.
11	Recommendations for repair (required service details)	Replacing new Cabinet, Main board, Display, Parameter board, ECG module unit.
12	Cost of spares (specify parts and cost)	Front Unit Assembly - 1,10,000 (Monitor Cabinet + Display) ECG Board without Microdrive Module - 1,50,000 (Main board & Parameter board)

total amount = 2,91,200

13	Asset Value	70,000/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Repairing cost higher than the Asset Cost
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Attached (OEM)
16	Reasons for recommending the equipment as BER	Major part of the machine are fully damaged.
17	Name & Signature of CYRIX Authority	Josekutty George 


*Not mandatory *#Based on the period of life and value as per the BER guidelines*

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


All major parts such as (main board, Display damaged
So RBER

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


12-12-2023
Signature of JC BM (NHM)

Date



Signature of  Superintendent
General Hospital
Thiruvananthapuram
Superintendent / Medical Officer (i/c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 160849

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>Gov. Thiruvananthapuram</u></p> <p>Address <u>Petta Junction</u> <u>Thiruvananthapuram, Kerala</u></p> <p>Ph : <u>9074198898</u></p>	<p>Call Registration Date : <u>22/11/2023</u></p> <p>Caller ID : <u>81551</u></p> <p>Date of Visit : <u>22/11/2023</u></p> <p>Asset No. : <u>0111053</u></p> <p>EQPT Name : <u>Multiphas Monitor</u></p> <p>Manufacture <u>CE</u> Model : <u>Cuescape B6870</u></p> <p>S. No <u>92W112131784HA</u> Dept. <u>Covid ICU</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Checked the machine and found display board, backlight, Hostler board, and body also damaged.

Action Taken :
Repairing display board, backlight, Hostler board and replacing body is not economically possible.

Completed Date : 28/11/2023 Time : 01:00 pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>RIBER</u>	<u>NR</u>	<u>NR</u>	<u>NR</u>
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Josekutty George</u>	<u>28/11/2023</u>	<u>10:00 am</u>	<u>01:00 pm</u>

Customer Remark Machine parts are broken. Completed Pending

<p>Service Engineer Name : <u>Josekutty George</u></p> <p>Signature :</p> <p>Date : <u>28/11/2023</u></p> <p>Contact Number :</p>	<p>Customer Name : <u>SALUMOLP</u></p> <p>Signature : <u>Soban</u></p> <p>Date : <u>28/11/23</u></p> <p>Contact Number : <u>8547672645</u></p> <p>Designation : <u>Nursing officer (NIP)</u></p> <p>Hospital Seal :</p>
---	---



THE PURCHASE MANAGER,
CYRIX HEALTH CARE PVT LTD
ERNAKULAM

PMS/QTN/208/2023
08-12-2023

QUOTATION

SL #	DESCRIPTION	HSN	GST%	QTY	RATE	AMOUNT
1	Front Unit Assembly for Touchscreen Model - CARESCAPE™ B650	9018	12	1	1,10,000.00	1,10,000.00
2	CPU Board without Micro Drive on Module Careescape B650	9018	12	1	1,50,000.00	1,50,000.00
	GST					31,200.00
	TOTAL					2,91,200.00

(Price Quoted on Exchange Basis; If defective spare parts are to be retained at Hospital,

Non-Exchange price would be 20% Extra plus applicable Taxes)

Price : INCLUDING GST AS MENTIONED ABOVE

Delivery : within 60 days after getting the confirmed Order

Payment : 100% against delivery by Cheque/ Net Transfer

Validity : One month from date of quotation

WARRANTY : One year only

PACE Marketing Services,

Service Department.



Authorized Service Dealer



Authorized Dealer
Wipro GE Healthcare







KMS-GI KIPi
TOLL FREE NO.: 1-800-270-0690
(800) 8907579 - 70010037

Preventive Maintenance Schedule A/C
0 1 0 3 0 4
0 2 0 5

TYPE B550-01
[50] SEM12131784HA

GE Healthcare & Allied Co.
GE Healthcare
P.O. Box 100000
Atlanta, GA 30308
Made in England
UL
CE
RoHS
WPA
100, 240 V ~ 14.0 VA
50/60 Hz ~ 1.44-6.00 mA



16.4 lbs
18.0 lbs
18.0 lbs

CYRIX
REALTEK & SONY 1217841A
CN: GHP 573
PRO : MULTIPARA MONITOR
ID : 0111053
CAL.ON : 27-04-2023
CAL.DUE : 26-04-2024
180-890-4398 Certified Co. Ph: 87445600

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 423-7669
BAR CODE - (8004-890615225) 0111053

388 74764220



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (REMP)

TOLL FREE NO: 1800 - 425-7669
0111054


BAR CODE - (8004-890615225)



CYRIX
HEALTHCARE ASSOCIATES
CIVIL SERVICE

CS: GEN ST
ID: 011084
CAL: 27-04-1023
CAL: 26-04-1024

CS: GEN ST
ID: 011084
CAL: 27-04-1023
CAL: 26-04-1024



REPAIR SERVICE
REPAIR SERVICE
REPAIR SERVICE