

13	Asset Value	8089
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Not available.
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix
16	Reasons for recommending the equipment as BER	Machine was donated. Machine is aged upto 17 years. Spare parts not available. So recommending BER.
17	Name & Signature of CYRIX Authority	Nidhish Roy <u>NDR</u>

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Installation Date & price need to check with main stock of FAC -

[Signature]
 MEDICAL ENGINEER
 JUNIOR CONSULTANT
 NHM
 KOTTAYAM
 Signature of JCBM (NHM)

Bindu.S *[Signature]*
 nursing officer *[Signature]*

Date 22/12/2023


Seal

[Signature]
 Dr. Sals Thomas
 Medical Officer in Charge
 Superintendent / Medical Officer (I/c)
 Marangattupilly



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 170657

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>Phc. Marangattupally</u></p> <p>Address <u>Marangattupally</u> <u>Kottayam</u></p> <p>Ph : <u>9845034356</u></p>	<p>Call Registration Date : <u>14-12-2023</u></p> <p>Caller ID : <u>25696</u></p> <p>Date of Visit : <u>18-12-2023</u></p> <p>Asset No. : <u>0541748</u></p> <p>EQPT Name : <u>Sterilizer (Autoclave)</u></p> <p>Manufacture <u>NA</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>OP</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Cover rusted, coil damaged, power cord required, Door lock damaged, washer broken, a nil, pressure gauge & valve rusted.

Action Taken : Checked. Machine is totally damaged condition. Required coil, coil power cord, washer, door locks, pressure release weight, pressure releasing knob. Economically not repairable.

Completed Date : 22-12-2023 Time : 02:50pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Nidhish Roy</u>	<u>22-12-2023</u>	<u>01:50pm</u>	<u>02:50pm</u>

Customer Remark Autoclave not working

Completed Pending




<p>Service Engineer Name : <u>Nidhish Roy</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>20-12-2023</u></p> <p>Contact Number : <u>9072575296</u></p>	<p>Customer Name : <u>Do. Saly Thomas</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>[Date]</u></p> <p>Contact Number : <u>Primary Health centre</u></p> <p>Designation : <u>Marangattupally</u></p> <p>Hospital Seal : <u>[Seal]</u></p>
---	---

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Kottayam
2	Name of Hospital	Phe Marangattupally
3	Name of Equipment with Make, Model and Serial Number	Sterilizer (Autoclave Cooker) Make: NA Model: NA Sn: NA
4	Equipment ID & Barcode	0541748
5	Date of purchase / Year of manufacture /Installation Date	27-11-2006
6	Warranty details (Yes/No)	No
7	*AMC/ CAMC Period agreed at the time of purchase	No
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	14-12-2003
9	Action taken	Checked. Serviced.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	Required 2000w coil, power cord with connector, Door washer, door locks, pressure releasing weight, pressure releasing knob.
12	Cost of spares (specify parts and cost)	Spare parts not available.



CYRIX BIOMEDICAL EQUIPMENT
PROGRAMME (B) ANCE
TOLL FREE NO: 1800 - 42
BAR CODE - (8004-890615225)
0541748











AUTOCLAVE

Date	M.S P.No		Qty	To Stock	
31/7/17	103	B/F from old substock pag 001 II page no - 66	01	01	
on 10/6/2008		Received from Standard Surgical syndicate, Kottayam, donated by Lions Club bill no 1624 Rs - 889/- dated 27/11/2006			
31/7/17		In use since 16/3/2009. 5/12/07	01.		Ver SW
15/3/2022		Received from store - Vertical autoclave	01	02	b2.
15/5/2023		Transferred to unserviceable Requirs pay no: 55 no: 1		01	Bdow b2