

25/12/22

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

| Sl. No | Particulars | Details |
|--------|--|--|
| 1 | Name of District | Kottayam |
| 2 | Name of Hospital | PHC Thiruvapur |
| 3 | Name of Equipment with Make, Model and Serial Number | Weighing Machine make: Samso model: nil s/no: nil |
| 4 | Equipment ID & Barcode | 0540462 |
| 5 | Date of purchase / Year of manufacture / Installation Date | 01-10-2018 |
| 6 | Warranty details (Yes/No) | No |
| 7 | *AMC/ CAMC Period agreed at the time of purchase | No |
| 8 | Date of breakdown (Date of registration of complaint through email/ Toll free) | 13-12-2023 |
| 9 | Action taken | checked and found not working |
| 10 | Present status of the equipment (Fully damaged / partially damaged) | Fully damaged. |
| 11 | Recommendations for repair (required service details) | Need to replace the battery connectors, rusted internal parts to make it functional. |
| 12 | Cost of spares (specify parts and cost) | spare not available |



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 172413

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

| | |
|---------------------------------|---|
| Health Facility : <u>PHC</u> | Call Registration Date : <u>13-12-2023</u> |
| Address : <u>PHC Thiruvappu</u> | Caller ID : <u>85674</u> |
| <u>Kottayam</u> | Date of Visit : <u>16-12-2023</u> |
| Ph : <u>9695264316</u> | Asset No. : <u>0540462</u> |
| | EQPT Name : <u>Weighing Machine</u> |
| | Manufacture : <u>Samso</u> Model : <u>Digital</u> |
| | S. No. : <u>-</u> Dept. : <u>opd</u> |

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Weighing Machine not Working.

Action Taken : Checked and found battery loose and leaked. Due to this connectors were rusted and broken. Internal connections and body parts are loose and tired. Hence recommended for replacement.

Completed Date : Time : Spare Required

Spare Replaced Requested

| Description | Qty. | Part Number | PR Number |
|-------------|------|-------------|-----------|
| 1. | | | |
| 2. | | | |
| 3. | | | |

| Cyrux Engineer | Date | Start Time | End Time |
|--------------------|-----------------|---------------|----------------|
| <u>Sanish Swan</u> | <u>16/12/23</u> | <u>9:30am</u> | <u>10:00am</u> |

Customer Remark : Completed Pending

MEDICAL OFFICE
PHC Thiruvappu
Kottayam
Shanish Swan
16/12/23
10:00

| | |
|--|---|
| Service Engineer Name : <u>Sanish Swan</u> | Customer Name : <u>DR. RAJESH</u> |
| Signature : <u>[Signature]</u> | Signature : <u>[Signature]</u> |
| Date : <u>16/12/2023</u> | Date : <u>16/12/2023</u> |
| Contact Number : <u>9946646571</u> | Contact Number : <u>2495658937</u> |
| | Designation : <u>Assistant Surgeon in 1/c</u> |
| | Hospital Seal : |

C-10


KMSCL  KTPL
TOLL FREE NO:- 1800-270-0690
(R004) 8907579 - 70051033

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7669 
BAR CODE - (8004-890615225) 
0540462

2
PASSED

Use for your protection. Do not touch.




| | | |
|----|--|--|
| 13 | Asset Value | 2590 |
| 14 | *Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value | Spare not available |
| 15 | Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) | Lytic report recommends the equipment to be replace |
| 16 | Reasons for recommending the equipment as BER | The repairing of the equipment is economically not feasible. Hence recommended the equipment as BER. |
| 17 | Name & Signature of CYRIX Authority |  Sanku Swan |


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
 Asset value - 2590
 Cost of spare = spare not available
 Life period - 5 years 2 months
 (Installation date - 11/10/2018)
 As per the BER guidelines recommended as BER


 Jayaram
 CONSULTANT BIOMEDICAL ENGINEER
 NATIONAL HEALTH INSTITUTION
 KOTTAYAM

Date _____


 Signature of Superintendent / Medical Officer (i/c)
 11.50 AM
 22-12-2018



MEDICAL OFFICER
 PHC Thiruvarpur
 Kottayam

2
PASSED



NOTHINGWAZ MACHINE
(DISTRIBUTION)

| | | | | | |
|---|---|----------|----------|-------------------|---------------------|
| <p>1/10/18 (Ss Andacio Page:18)</p> | <p>Received from karunya medicine depot as per 1000 11457 dt 1/10/18.</p> | <p>2</p> | <p>2</p> | <p>2</p> | <p>2590/- 1</p> |
| <p>1/10/18</p> | <p>issued to s/n</p> | <p></p> | <p>2</p> | <p>NIL Rd</p> | <p>9/- 1</p> |