


REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	PHC Thalavayam
3	Name of Equipment with Make, Model and Serial Number	Nebulizer Infi Neb model : CN-01WB SN : 033002
4	Equipment ID & Barcode	85643 & 0543432
5	Date of purchase / Year of manufacture / Installation Date	30/12/2008
6	Warranty details (Yes/No)	NIL
7	AMC/ CAMC Period agreed at the time of purchase	NIL
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	14/12/2023
9	Action taken	checked motor Complaint, Piston washer Complaint.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Motor Complaint and piston washer Complaint.
12	Cost of spares (specify parts and cost)	Motor : 800 Piston washer : 300 <hr/> 1100

13	Asset Value	1300/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	84%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX
16	Reasons for recommending the equipment as BER	Motor complain and Piston washer complain
17	Name & Signature of CYRIX Authority	ANAL BROW 

\*Not mandatory \*Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NH&M


Asset value - 7500/-

Cost of spare = 1100 (84%)


Installation Date: 30/12/2008


Life period - 1.5 years


As per the BER guidelines recommended as BER.

  
Signature of JC. BM (NH&M)

Date



For  Signature of  
Medical Officer (PH&M) (1/1)  
Primary Health Centre  
Polevarodam, PIN 687



**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



Act: SERVICE PROVIDER  
Tender No. WO-272321-282/2018

**CYRIX**  
HEALTHCARE PVT LTD

No. : 166576

ISO 13485 : 2012 & ISO 9001:2008 CERTIFIED COMPANY: AERB Approved Service Agency

**Service Report**

30/64 1 B, Patta Junction, Poonithara, Kochi - 682 018, Kerala  
Ph : 08472 99500 Website : www.cyrix.com | Email : temp.k@cyrix.in

Health Facility: <u>phc</u>	Call Registration Date: <u>14/12/2023</u>
Address: <u>Thalayarathur</u>	Caller ID: <u>85643</u>
<u>1007900</u>	Date of Visit: <u>19/12/2023</u>
Ph: <u>9496930857</u>	Asset No.: <u>0543432</u>
	EQPT Name: <u>nebulizer</u>
	Manufacture: <u>Inf. Neb</u> Model: <u>CN-014B</u>
	S. No. <u>033002</u> Dept. <u>OP</u>

Service Classification : Breakdown Call  PMS  Calibration  Cust. Training

Problem Identified : no fixes

Action Taken : checked motor Complaint and piston washer Complaint. It is economically inoperable. So it is recommended for ROR.

Completed  Date : \_\_\_\_\_ Time : \_\_\_\_\_ Spare Required

Spare Replaced  Requested

Description	Qty	Part Number	PR Number
1. <u>na</u>	<u>na</u>	<u>na</u>	<u>na</u>
2.			
3.			


  

Cyrix Engineer	Date	Start Time	End Time
<u>Anil BABU</u>	<u>19/12/2023</u>	<u>10:30AM</u>	<u>12:30PM</u>

Customer Remark: Equipment not working Completed  Pending

Service Engineer Name: Anil BABU  
Signature: \_\_\_\_\_  
Date: 19/12/2023  
Contact Number: 9567575857

Customer Name: \_\_\_\_\_  
Signature: \_\_\_\_\_  
Date: \_\_\_\_\_  
Contact Number: \_\_\_\_\_  
Designation: \_\_\_\_\_  
Hospital/Dept: \_\_\_\_\_



NEBULIZER

		28 - 1300/-	
30/12/2008	Rtd from the store. Issued to op	1	2
5/10/2023	Rtd from the store	1	2.
5/10/2023	Issued to op		1



CYRIX HEALTH CARE PVT.LTD  
 BEMP SERVICE PROVIDER  
 PREVENTIVE MAINTENANCE SERVICE  
 PMS DONE ON: 15/02/2023  
 NEXT RMS DUE: 15/02/2023  
 PERFORMED BY: [Signature]

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)  
 TOLL FREE NO: 1800 - 425-7669  
 BAR CODE - (8004-8906) 5225) 0543432

**Infi-neb**

**RBER** **CYRIX™**  
 HEALTH CARE PVT LTD  
 "Genius in Healthcare"  
 Ticket No.: 85643  
 RBER approved on: 19/12/2023  
 Condemn on:  
 Remark: