



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 172394

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.ki@cyrrix.in

Health Facility : THQH IRITTY	Call Registration Date : 16/12/2023
Address : Iritty	Caller ID : 84247
KANNUR	Date of Visit : 17/12/2023
Ph : 9446 664 194	Asset No. : 1321331
	EQPT Name : NEBULISER
	Manufacture : Lifecare Model : NA
	S. No. : 1620 Dept. : WARD

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : No Fumes.

Action Taken : Checked the equipment and found no fumes. Found piston damaged and piston bearing is defective. Also found motor is defective. Enquired for spares.

Completed Date : 27/12/2023 Time : 4:00 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Sayanth	27/12/2023	3:00 PM	4:00 PM

Customer Remark Completed Pending

Service Engineer Name : Sayanth
Signature : *Sayanth*
Date : 27/12/2023
Contact Number : 8606 345 195

Customer Name : Bindu MP
Signature : *Bindu*
Date : 27-12-23
Contact Number : 9446664194
Designation : SNO
Hospital Site :




SUPERINTENDENT
TALUK HOSPITAL, IRITTY
KANNUR DIST. 671773

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	THQH IRITTY
3	Name of Equipment with Make, Model and Serial Number	EQUIPMENT NAME: NEBULIZER Make : LIFECARE Model: N/A SN: 1620
4	Equipment ID/ Barcode	1321331 #84247
5	Date of purchase/ Year of manufacture/Installation Date	25-08-2020
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 06-12-2023
9	Action taken	Checked the equipment and found no fumes, found piston damaged and piston bearing is defective also found motor is defective. Enquired for spares
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	

13	Asset Value	1353/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report attached
16	Reasons for recommending the equipment as BER	Checked the equipment and found no fumes, found piston damaged and piston bearing is defective also found motor is defective. Equipment installed on 25-08-2020 aged up to 3 years and 3 months. Checked the spares in market, but spares are not available because it's a low asset cost equipment. So we recommending for condemnation.
17	Name & Signature of CYRIX Authority	SAYANTH C 


*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that piston damaged, no fumes motor defective. Aged up to 3 years. its service cost economically not feasible. So recommended for BER as per BER protocol.



27/12/2023

Signature of JC BM (NHM)

CONSULTANT BIOMEDICAL
NATIONAL HEALTH MISSION
VANITHA

Beyond Economic Repair
as stated by Cyrix Engineer.
R. of 100% 31/12/23 res. steps may be
taken to repair/assemble where

Date: 29.12.23



Signature of
Superintendent
SUPERINTENDENT
TAMIL NADU STATE HOSPITAL
CHENNAI
ANNUR D-6/10/23

Nebulizer

		Rd	Used	Balans	Sign.
1-6-16	Tr in from General Stock Book Ad on 3-4-17 ① page no: 107	1		1	g
1-6-16	Tr in from NRHM Stock Book Ad on 29/11/2015 ② page no: 25	3		4	g
1-6-16	Tr in from HMC Stock Book Ad on 29/10/08 ③, 20/11/13 ④, 10/11/13 ⑤ page no: 10 8/2/2016 ⑥	5		9	g
1-6-16	Tr in from Stock Book of block Ad on 20/3/11 ⑦	1		10	g
16-8-16	Rd from 16 store used in wd used in wd condemned and destroyed as per order GOC(MS) NO: 63/2013/H & FWD/H & FWD dtd. 23/2/2013 Tvm.	2		12	g
			5	7	g
25/1/2019	Rd from store (Flood relief donation)	1		8	g
19-2-19	Received from the store (1001-20 & 2000f.) used in word.	5		13	
18/5/19	Tr. out to Dialysis unit (page no: 4)		1	12	
5/8/19	TR. out to unserviceable book page: 12		3	9	
6-6-2020	Received from the store ultrasonic nebulizer	1		10	
25/8/2020	Received from the store	8		18	
15/12/2020	Rd from CFLTC CFLTC page no: 112	2		20	
28/12/2020	Transferred CFLTC			18	
20/12/2020	Tr out to unserviceable register page No 20 (29/11/2015) (condemned)		3	17	g

S/2/16 ② 29/11/2015 3, 3/8/17-1

