



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 164066

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>C.M.C. VADAKKENCHERY</u>	Call Registration Date : <u>28/11/23</u>
Address <u>VADAKKENCHERY</u>	Caller ID : <u>52690</u>
<u>PALAKKAD</u>	Date of Visit : <u>30/11/23</u>
Ph : <u>701235759</u>	Asset No. : <u>0930360</u>
	EQPT Name : <u>NEBULIZER</u>
	Manufacture <u>OMRON</u> Model : <u>COMP AIR</u>
	S. No. <u>00097510</u> Dept. <u>OP</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : Checked the machine. Found that motor burned and piston rod broken. Machine installed on 10/12/16 and aged upto 7 plus years. Need to replace compress motor and piston rod for further checking.

Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>DWINIK</u>	<u>4/12/23</u>	<u>11:30</u>	<u>12:30</u>

Customer Remark Completed Pending



Service Engineer Name : DWINIK
Signature : [Signature]
Date : 4/12/23
Contact Number : 7598847155

Customer Name : MARY.K.R
Signature : [Signature]
Date : 4/12/2023
Contact Number : 9447377777
Designation : S.N.O
Hospital Seal :


[Signature]
Superintendent
Community Health Centre
Vadakkenchery, Palakkad

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	KMC YADAKKENCHERY
3	Name of Equipment with Make, Model and Serial Number	NEBULIZER MAKE : OMRON MODEL: COMP AIR: SN: 0097510
4	Equipment ID & Barcode	0930360, 82690
5	Date of purchase / Year of manufacture / Installation Date	10/12/2016
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/ NO CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	28/11/23
9	Action taken	Checked and found that motor burned and piston rod broken.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	

13	Asset Value	1,353
14	*Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	—
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX service report
16	Reasons for recommending the equipment as BER	Checked and Found that motor burned and piston rod broken. Machine installed on 10/12/2016 and outlived 7 plus years. Spare quote not submitted because spares are not available in market and recommending for condemnation
17	Name & Signature of CYRIX Authority	D N I N I S 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Recommended for BER because motor burned and piston rod broken which was installed on 10/12/2016.




3/01/24.
Signature of JC BM (NHM)

Date




Signature of
Superintendent / Medical Officer (i/c)
Community Health Centre
Vadakkencherry, Palakkad



Superintendent
Community Health Centre
Vadakkancherry, Palakkad

MARY.K.R. M.S.
Maavay¹ For Dr. Sreeja Jayasajy

B,

Sub-Stock Register
Instrument & Furniture
General + NRHM



Superintendent
Community Health Centre
Vadakkancherry, Palakkad
For the Superintendent
Jairaj

MAR 11 2008

36	Office Table c Two desks (on physical verification)	79
37	Office Table c One desk (on physical verification)	73
38	Chair wooden c 115m	75
39	Wooden stool	77
40	Bench wooden	79
41	Laboratory stool	81
42	Desk wooden	83
43	Table wooden without chair	85
44	Needle dissection burner	87, 171
45	O ₂ cylinder	89
46	Refrigerator	91, 137
47	Microscope	93
48	Iron stool	95
49	Office table (iron) c one desk	97
50	Office chair Revolving	99
51	Bed side locked	101
52	Steel Rack open	103
53	Combi chair	105
54	Standard trolley	107
55	MEDICINE TROLLEY	109
56	Back Rest	111
57	wheel chair	113
58	MINDING GAS CONNECTION	115
59	steel bin large	121
60	Steel Bin medium	125
61	Steel Bin small	127
62	Bed side screen	129
63	Needle Burner	130
63	ECU MARCHING	135
64	Stool (on physical verification)	137
65	Back rest cat Adjusted manually screws & nutting Superintendent Community Health Centre Vadakkanchery, Palakkad-688 001	22



For Dr. Sreedha
Tarak...

Nebulizer

Brought forward for
LP-Book page 11

93+337

Date	Particulars	Qty	Nil	Total	Remarks
2/2/10	Bill NO - 709 22-12-2009 Received from the store. 2 (KMC) 2			2	Nil
2014	2014 new Pajenu: 337. entered - LP Book				
10/12/16	Received from store (LP) Issued for use	1		1	Nil
10/12/16	Received from store (LP) Issued for use	1		1	Nil
10/3/2017	Received from store (LP) Block Panchayathi (AMRON) (Stock) Issued for use	2		2	Nil
3/4/17	Received from store (LP) (Panchayathi fund - (LP)) Issued for use	1		1	Nil
13/6/2020	Received from store (LP) Issued for use	1		1	Nil
27/12/2020	Received from store Issued for use	1		1	Nil



Superintendent
Community Health Centre
Vadakkencherry, Palakkad
MAYYAR
S.No

For: Dr. Seeta Jayaram