



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 153676

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>PHC</u>	Call Registration Date : <u>28/10/23</u>
Address <u>Chethalayam</u>	Caller ID : <u>77258</u>
..... <u>Wayanad, Kerala</u>	Date of Visit : <u>31/10/23</u>
Ph : <u>9961462082</u>	Asset No. : <u>1240532</u>
	EQPT Name : <u>Weighing machine</u>
	Manufacture <u>Samso</u> Model : <u>Adult</u>
	S. No. <u>NA</u> Dept. <u>Observation room</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :
..... Not getting on

Action Taken : Checked and found board and display complaint. Board and display replacement cost is beyond economic repairs, hence the weighing machine recommended for large condemnation.

Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Arjun.P</u>	<u>31/10/23</u>	<u>1:30 PM</u>	<u>2:30 PM</u>

Customer Remark Completed Pending


Service Engineer Name : <u>Arjun.P</u>	Customer Name : <u>Chethalayam</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>31/10/23</u>	Date : <u>31/10/23</u>
Contact Number : <u>7593847137</u>	Contact Number : <u>9961462082</u>
	Designation : <u>Nursing officer</u>
	Hospital Seal :

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Wayanad
2	Name of Hospital	PHC Chethalayam
3	Name of Equipment with Make, Model and Serial Number	Weighing machine make : Samsco (Adult)
4	Equipment ID/ Barcode	1240532
5	Date of purchase/ Year of manufacture/Installation Date	13/03/2019
6	Warranty details	Not Available
7	*AMC/ CAMC Period agreed at the time of purchase	No
8	Date of breakdown/ Date of registration of complaint	28/10/2023
9	Action taken	Checked the equipment found board and display faulty.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Details of service required (Recommendations for repair)	Board and display replacement cost is beyond economic. cost is more than the percentage of asset value.
12	Cost of spares (specify parts and cost)	Sensor board - 1416 Display - 448 Total - 1864/-


13	Asset Value	1600/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	116.5%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Service report attached Quotation from local vendor.
16	Reasons for recommending the equipment as BER	Checked and found - board and display complaint. Board and display cost is beyond economic/more than percentage of asset value; hence the equipment recommended for condemnation.
17	Signature of CYRIX Authority	Arjun.p. 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Board and display are complaint
Recommended for Condemnation




Signature of JC BM (NHM)
Bio-Medical Engineer
National Health Mission
(Arogyakeralam)
Wayanad-673122

BOARD & DISPLAY FAULTY.

Date 31/10/2023




Signature of Superintendent/Medical Officer

 T.C.M. 27378
 മെഡിക്കൽ ഓഫീസർ
 ചെതലായം, വയനാട് - 673 592

Ospio Healthcare Kannadikal, calicut, Kerala, GSTIN: 32AZLPT7342N1ZP Mobile: 9447338356	Quotation No. 117	Quotation Date 04/01/2024	Expiry Date 18/02/2024

BILL TO CYRIX HEALTHCARE PVT LTD Address: Poonithura, Ernakulam, GSTIN: 32AAFCC2499H2ZM State: Kerala	SHIP TO CYRIX HEALTHCARE PVT LTD Address: Poonithura, Ernakulam,
--	--

S.NO.	ITEMS	QTY.	RATE	TAX	AMOUNT
1	Load sensor for samsco digital scale	1 PCS	1,200	216 (18%)	1,416
2	Display for weigh scale	1 PCS	400	48 (12%)	448
TOTAL		2		₹ 264	₹ 1,864

HSN/SAC	Taxable Value	CGST		SGST		Total Tax Amount
		Rate	Amount	Rate	Amount	
-	400	6%	24	6%	24	₹ 48
-	1,200	9%	108	9%	108	₹ 216

Total Amount (in words) One Thousand Eight Hundred Sixty Four Rupees	
Terms and Conditions 1. Goods once sold will not be taken back or exchanged 2. All disputes are subject to [ENTER_YOUR_CITY_NAME] jurisdiction only	



STOCK Register Of EQUIPMENTS, INSTRUMENTS AND LINEN FHC. CHETHALAM

This is to certify that this Register
contains 440 Pages which is serially
numbered from 1 - 440



[Signature]
 உதவி இயக்குநர்
 காட்டு வளத்துறை
 சேதலம் - 613 092

Sno	Items	Page
40	Floor mat	81
41	Fire exti Extinguisher	83
42	Archer bag Adult	85
43	Archer bag paediatric	87
44	wheel chair	89
45	Stoacher trolley	91
46	Refrigerators	93
47	Ear care kit	95
48	Weighting machine digital	97
49	Thermometer digital	99
50	Examination torch	101
51	water purified	103
52	Height measuring stand	105
53	LED Examination light	107
54	Mattress for Examination table	109
55	Rexin pillow	111
56	Infrared thermometer	113
57	Long mirror	115
58	Spino meter	117
59	Pulse oximeter	119
60	Suction Apparatus	121
61	Tailoring Scissors	123
62	Artery forceps Straight	125
63	Thumb forceps non toothed	127
64	Needle Holder	129
65	ECCO machine	131
66	Hot water bag	133
67	Chettle container	135
68	Knee hammer	137
69	Bed sheet	139
70	Pillow cover	141
71	Oxygen Concentrator	143
72	Spinal Drain	145
73	Crash cart	147
74	Iron Cot	149
75	mattress	151
76	Digital display Refrigerator Thermometer	153
77	Air port Chair	155
78	Revolving Chair	157

