

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Ernakulam
2	Name of Hospital	Thgh Kothamangalam
3	Name of Equipment with Make, Model and Serial Number	philips, philulux 550 O+ light
4	Equipment ID & Barcode	0722365
5	Date of purchase / Year of manufacture /Installation Date	18/12/13
6	Warranty details (Yes/No)	NO
7	*AMC/ CAMC Period agreed at the time of purchase	
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	30/9/23
9	Action taken	check found transformer damaged
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	
12	Cost of spares (specify parts and cost)	EOL attached

13	Asset Value	2,36,915
14	% Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Attached.
16	Reasons for recommending the equipment as BER	Fully damaged.
17	Name & Signature of CYRIX Authority	Akhil T. Lenin Akhil T. Lenin

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Equipment transformer part is completely burned.
Equipment is 10 yrs Old. Tried to repair. But unable to do the same. Submitted BER after 30 days due to the great delay from the OEM.

Devally
11/01/24
Signature of JC BM (NHM)

DO NOT WORK

3/10/24
Dr. JAMES SIMON PEREIRA
MBBS, D'ORTHO
ORTHOPEDIC SURGEON
KERLA HEALTH SERVICES



Date

Seal



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No. : 151730

CYRIX
HEALTHCARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency
30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Service Report

Health Facility : Thgh
Address : Kothamangalam
Ernakulam
Ph : 8606002385

Call Registration Date : 30/9/23
Caller ID : 72176
Date of Visit : 3/9/23
Asset No. : 0722365
EQPT Name : OT Light
Manufacture Philips Model : Philips S50
S. No. 100023 Dept. OT

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : unit not working.

Action Taken : Check the unit, check bulb lamp connections check and found issue in transformer side totally burned and full of crust, all connections and bulb were fully damaged. unit requested for R.B.E.R.

Completed Date : 3/9/23 Time : 3:30 Spare Required

Spare Replaced <input type="checkbox"/> Requested <input type="checkbox"/>		Qty.	Part Number	PR Number	
Description					
1.					
2.					
3.					
Cyrix Engineer			Date	Start Time	End Time
<u>Akhil T. Lenin</u>			<u>3/9/23</u>	<u>12:00</u>	<u>3:30</u>

Customer Remark : Completed Pending

Service Engineer Name : Akhil T. Lenin
Signature : [Signature]
Date : 3/10/23
Contact Number : 9605974326

Customer Name : AKSA SIVAN
Signature : [Signature]
Date : 3/10/23
Contact Number :
Designation : Dr. Aes Besia Taboogist
Hospital Seal :



KERALA SURGICAL EQUIPMENT CO

Everything in Hospital Needs Under One Roof!

NON -REPAIRABLE LETTER

13-12-2023

To,

CYRIX HEALTHCARE

Sub : Servicing of OT Light (Philips)

Ref: THQH KOTHAMANGALAM - S-50

Sir,

With reference to above , one number of OT Light (Philips S-50)having severe defects ,and
The repairing cost will be of more than 80 % of the equipment Value

Hence we suggest that, the service of this equipment is not reasonable and buy a new one.

Thanking you,

For KERALA SURGICAL EQUIPMENT CO

Abby
MANAGER




41/2028, M.E.S CENTRE, NEAR TOWN HALL, BANERJI ROAD, KOCHI - 682 018. TEL: 0484 - 2369654, 2355858.

FAX: 4023232 Customer care: 9388352885/9447169654/55/ E-mail: info@keralasurgicals.com

sales@keralasurgicals.com, service@keralasurgicals.com








PHILIPS

ASS



STOCK REGISTER OF T.H.Q. HOSPITAL, KOTHAMANGALAM

Name of Article *Shadowless lamp - double d*

Date	No and date of voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of Receiver	Rem
		Brought forward					
18/12/13		RFS		Rs 1	2.36	915	-
		philips O.T light shadowless ceiling mount Double claw model 550-D40 equipment code 89-103 c Accessories					
		Sterile Handle		2	2		
		Bulb		10	10		

Rosabilly
Rosabilly