

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7869
SERIAL CODE - (8004-890615225)
0611250

ALPHANE B

NOT WORKING



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7669
SERIAL CODE - (8004-890615225)
0611250

ALPHANE B

Not working

**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 147754

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <i>OH Thodupuzha</i>	Call Registration Date : <i>11-1-2024</i>
Address <i>Thodupuzha</i>	Caller ID : <i>90143</i>
<i>Idulek</i>	Date of Visit : <i>13-1-2024</i>
Ph : <i>08848138434</i>	Asset No. : <i>0611250</i>
	EQPT Name : <i>Nebuliser</i>
	Manufacture <i>Alpha Fiberone</i> Model : <i>Na</i>
	S. No. <i>Null</i> Dept. <i>Isolation Ward</i>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : *Nebuliser piston and piston mechanism are faulty*

Action Taken : *Cheek the machine. Found that piston and piston mechanism are faulty unable to service*

Completed Date : *13-1-2024* Time : *2:00 PM* Spare Required

Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1. <i>Na</i>	<i>Na</i>	<i>Na</i>	<i>Na</i>
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<i>Vishnuprasad TJ</i>	<i>13-1-2024</i>	<i>1:30 pm</i>	<i>2:00 pm</i>

Customer Remark Completed Pending

Service Engineer Name : *Vishnuprasad TJ*
Signature : *[Signature]*
Date : *13-1-2024*
Contact Number : *9496462254*

Customer Name : *RAJINI M R*
Signature : *[Signature]*
Date : *13/1/24*
Contact Number : *828429078*
Designation : *SENIOR NURSING OFFICER*
Hospital Seal :



[Handwritten Signature]
20/1/24

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Idukki
2	Name of Hospital	DH Thodupuzha
3	Name of Equipment with Make, Model and Serial Number	Equipment :- Nebuliser Make :- Alpha Fibersonics Model :- Serial No :-
4	Equipment ID & Barcode	90143 and 0611250
5	Date of purchase / Year of manufacture / Installation Date	23-11-2020
6	Warranty details (Yes/No)	No
7	*AMC/ CAMC Period agreed at the time of purchase	1 year
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	11-1-2024
9	Action taken	Try to service
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	check the machine found that Piston and Piston mechanism are faulty unable to service
12	Cost of spares (specify parts and cost)	Nil

13	Asset Value	1875
14	% Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Nil
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	check the machine found that piston and piston mechanism are faulty unable to service so recommended to R.B.E.R
17	Name & Signature of CYRIX Authority	Vishnu prasad T T <u>B S</u> 11-1-2024

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* **Attach Photograph**


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Consultant Biomedical Engineer
National Health Mission
Arogyakeralam
Idukki-685603

~~R. S. A.~~
24/1/24

Signature of JC BM (NHM)

Date



Signature of
Superintendent
District Hospital
Thodupuzha