

MEM

START
STOP

SET



PUL
/MIN

DIA
mmHg

SYS
mmHg





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : **129219**

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>The Canchiyar</u></p> <p>Address <u>1 Canchiyar</u> <u>Cheruvu</u></p> <p>Ph : <u>9447309811</u></p>	<p>Call Registration Date : <u>11.1.24</u></p> <p>Caller ID : <u>90229</u></p> <p>Date of Visit : <u>16.1.24</u></p> <p>Asset No. : <u>0640755</u></p> <p>EQPT Name : <u>Bg Apparatus</u></p> <p>Manufacture <u>Nil</u> Model : <u>alil</u></p> <p>S. No. <u>alil</u> Dept. <u>OP</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :
Alert Working


Action Taken :
Check the Bluehore, found that Machine Menu board and Motor Complaint. No service not possible.

Completed Date : 16.1.24 Time : 4:00 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Murukh Krishna T D</u>	<u>16/1/24</u>		



Customer Remark Completed Pending

<p>Service Engineer Name : <u>Murukh Krishna T D</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>16/1/24</u></p> <p>Contact Number : <u>9847798000</u></p>	<p>Customer Name : <u>Smrigna George</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>16/1/24</u></p> <p>Contact : <u>9447309811</u></p> <p>Designation : <u>MEDICAL OFFICER IN CHARGE</u></p> <p>Hospital Seal : <u>NIDDIYAR HEALTH CENTRE</u></p>
---	---

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Belurki
2	Name of Hospital	Phe Kanahiyar
3	Name of Equipment with Make, Model and Serial Number	Rp Apparatus Make : Model : SN :
4	Equipment ID & Barcode	90229 / 0640755
5	Date of purchase / Year of manufacture / Installation Date	14/5/2019
6	Warranty details (Yes/No)	No
7	*AMC/ CAMC Period agreed at the time of purchase	One Year
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	11/1/2024
9	Action taken	try to Service
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Check the Machine, found that Machine also board and Motor not working. No Service not possible.
12	Cost of spares (specify parts and cost)	Nil

13	Asset Value	2200
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX Service Report attached
16	Reasons for recommending the equipment as BER	Check the Machine, found that Machine Motor burned and Motor damaged. So its service not possible. So its recommended for RBER.
17	Name & Signature of CYRIX Authority	Muralikrishnan, D 16/11/24

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

R#1111111111
17/11/2024

Consultant Biomedical Engineer
National Health Mission
Aravakeralam
Idukki 685603-

Signature of JC BM (NHM)

Date



Signature of
Superintendent Medical Officer (i/c)
MEDICAL OFFICER IN CHARGE
FAMILY HEALTH CENTRE
KANCHIYAR 685 51
MUKKI DIST.