







**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



SERVICE PROVIDER  
Tender No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

No. : 166585

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>THM</u>	Call Registration Date : <u>8/12/2023</u>
Address : <u>Kranvilgaal</u>	Caller ID : <u>84577</u>
<u>Kottayam</u>	Date of Visit : <u>21/12/2023</u>
Ph : <u>9947968248</u>	Asset No. : <u>0520392</u>
	EQPT Name : <u>Refrigerator</u>
	Manufacture : <u>Godrej</u> Model : <u>Cold gold</u>
	S. No. : <u>0A</u> Dept. : <u>Casualty</u>

Service Classification : Breakdown Call  PMS  Calibration  Cust. Training

Problem Identified : No cooling

Action Taken : checked compressor and relay  
Complaint, need gas charging. It is  
economically irreparable. So it  
recommended for RBER

Completed  Date : ..... Time : ..... Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>nei</u>	<u>na</u>	<u>nei</u>	<u>nei</u>
3.			


Cyrix Engineer	Date	Start Time	End Time
<u>AMAL BABU</u>	<u>21/12/2023</u>	<u>11:15 am</u>	<u>12:00 pm</u>

Customer Remark : Equipment is not working Completed  Pending   
Bundu



Service Engineer Name : <u>AMAL BABU</u>	Customer Name : <u>[Signature]</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>21/12/2023</u>	Date : <u>[Signature]</u>
Contact Number : <u>9567575983</u>	Contact Number : <u>[Signature]</u>
	Designation : <u>[Signature]</u>
	Hospital Seal : <u>[Signature]</u>




13	Asset Value	13,490
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	78.9%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX
16	Reasons for recommending the equipment as BER	Compressor and relay Complaint, need gas charging
17	Name & Signature of CYRIX Authority	ANIL BABU 

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value - 13490.  
 Cost of spare - 11100 (78.9%)  
 life period - 16 years (28/5/2007)  
 As per the BER guidelines recommended as  
 RBER

  
 9/1/2024  
 Signature of JC BM (NHM)  
 NATIONAL HEALTH MISSION  
 KOTTAYAM

Bundu  
 Sw.

Date \_\_\_\_\_ Seal \_\_\_\_\_

  
 Signature of  
 Superintendent / Medical Officer (i/c)  
 സൂപ്രണ്ടു  
 താലൂക്ക് മെഡിക്കൽ ഓഫീസ് ജൂനിയർ ഓഫീസർ  
 കുറുനിക്കുണ്ടാട് പി. ഒ. പിൻ - 686633

STOCK BOOK (MEDICINES) / INSTRUMENTS

Name of drug: Refrigerator - 300 ltr. (Gudra) Name of drug: .....

Date	Receipt	Issue	Balance	Date	Receipt	Issue	Balance	Remarks
	kg./gm./ml.	kg./gm./ml.	kg./gm./ml.		kg./gm./ml.	kg./gm./ml.	kg./gm./ml.	
27-5-07								Received from Hs - Secretary Block Panchayat, Uzhavoor at per Order No C1 - 1826/07 dt 12-3-07
					01 No		01 No	
22-5-07						01	NIC	Issued to Wz Lt/N (Substore Furniture Reg. Page No 33) M.S.M. Kal



**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	Thgh Kwanlagand
3	Name of Equipment with Make, Model and Serial Number	Refrigerator Crodrej Cold gold SO: NA
4	Equipment ID & Barcode	84577 & 0520392
5	Date of purchase / Year of manufacture / Installation Date	22/5/2007
6	Warranty details (Yes/No)	NIL
7	*AMC/ CAMC Period agreed at the time of purchase	NIL
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	8/12/2023
9	Action taken	Checked Compressor Complaint, relay Complaint.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Compressor and relay Complaint. need gas charging.
12	Cost of spares (specify parts and cost)	Compressor : 8000 Relay : 500 gas charging : 2500 <u>11000</u>

2/1/24