



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 120389

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kt@cyrix.in

Health Facility <u>FHC-KUNNAMANGALAM</u>	Call Registration Date : <u>25/01/2024</u>
Address <u>KOZHIKODE</u>	Caller ID : <u>98472</u>
<u>KERALA</u>	Date of Visit : <u>25/01/2024</u>
Ph : <u>8943401461</u>	Asset No. : <u>1142195</u>
	EQPT Name : <u>UPS</u>
	Manufacture <u>ASIA POWER</u> Model : <u>AP1050T</u>
	S. No. <u>UVO61167A00236</u> Dept. <u>LAB</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : MACHINE NOT GETTING ON

Action Taken : CHECKED AND FOUND POWER BOARD AND BATTERY DEFECTIVE, NEED TO REPLACE THOSE SPARES FOR FURTHER CHECKING AND WORKING CONDITION OF THE EQUIPMENT. ENQUIRED THE SPARE

Completed Date : 03/02/2024 Time : 01:45 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.	1		
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>PRABHATH.K</u>	<u>03/02/2024</u>	<u>01:30 PM</u>	<u>01:45 PM</u>

Customer Remark Completed Pending

machine Board damaged and battery damaged
machine 8 years old installation date 30/07/2015
Dinamik LT and SR

Service Engineer Name : PRABHATH.K
Signature
Date : 03/02/2024
Contact Number : 754384726

Customer Name Dr. Arghav
Signature
Date
Contact Number
Designation
Hospital Seal


MEDICAL OFFICER
Primary Health Centre
Kunnammangalam

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOZHIKODE
2	Name of Hospital	PHC KUNNAMANGALAM
3	Name of Equipment with Make, Model and Serial Number	Equipment :UPS for Semi Auto Biochemistry Analyser Make :AsiaPower Model :AP1050+ SN : UV060115A000236
4	Equipment ID & Barcode	92728 & 1142195
5	Date of purchase / Year of manufacture /Installation Date	30-07-2015
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	25-01-2024 (Toll free)
9	Action taken	Checked and found power board and battery defective.Enquired the spare with vendor.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	1)Power board – Rs-2400 2)7 Ah battery – 2no – Rs-1800 Total – Rs.4200

13	Asset Value	Rs . 4500/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	93%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached and vendor quotation attached
16	Reasons for recommending the equipment as BER	Checked and found power board and battery defective. Equipment out lived 8 + years. Repairing cost of the equipment is 93% both criteria met, so recommended equipment for condemnation.
17	Name & Signature of CYRIX Authority	PRABHATH.K 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Checked & found that power board and battery damaged. Machine installed on 30/07/2015. Repairing cost of the equipment is 93%. So recommended for condemnation

Consultant Bio Medical Engineer
National Health Mission
Kozhikode
Signature of JC BM (NHM)

Machine damaged
8 year old

Date

2/2/24

Seal



Signature of
Superintendent/ Medical Officer (i/c)

MEDICAL OFFICER
Primary Health Centre
Kunnammangalam

SEMI-AUTOANALYSER WITH UPS
Receipt

Received from the PNC store of PNC
Kunnamangalam on 31/7/2015

Installed in Lab 30/07/2015

I NO.


Put in use.

FAIC

Archer
3/2/15

MEDICAL OFFICER
Primary Health Centre
Kunnamangalam





Dr. S. S. S. S.

MEDICAL OFFICER
MEDICAL Health Centre
Kunnamangalam



TECHKNOW TEK Power Solutions

TO
Cyrix Healthcare Pvt Ltd

Date : 02-02-2024

Qt no : 2943

Ref:As Per Your Enquiry

We are placed to quote for ups Service

SLN	Description	Price	Unit	Amount
1	Ap 1050/24v 1, power board	2400/-	1	2400.00
2	7Ah battery	900/-	2	1800.00
Total Amount				4200.00

Tax and duty's : included
Payment : 100% Advance
Quotation validity : 15 days

Yours truly
Techknow tek power solutions
Ramanattukara
Calicut
GST: 32DAJPP0728L1ZJ
E-mail: techknowtek@gmail.com
Mob: 9746699021,9744050501



Near LIC,Airport Road,Ramanattukara calicut
Mob:9746699021,9744050501,
mail: techknowtek@gmail.com



[Signature]
MEDICAL OFFICER
Primary Health Centre
Kunnamangalam