



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 183054

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

| | |
|---|---|
| Health Facility Phc konichav..... | Call Registration Date : 17-1-2024..... |
| Address Kannur, Kerala..... | Caller ID : 91376..... |
| Ph : 9747204172..... | Date of Visit : 18-1-2024..... |
| | Asset No. : 1344379..... |
| | EQPT Name : B.P Apparatus..... |
| | Manufacture OMBON Model : hem 8712..... |
| | S. No. NA Dept. Main centre..... |

Service Classification : Breakdown Call **PMS** **Calibration** **Cust.Training**

Problem Identified : Showing error.....

Action Taken : checked the equipment and found BP reading error found problems with main board and pressure sensor. Required for spares.....

Completed **Date :** 18-1-24 **Time :** 12:30pm **Spare Required**

Spare Replaced **Requested**

| Description | Qty. | Part Number | PR Number |
|-------------|------|-------------|-----------|
| 1. | | | |
| 2. | | | |
| 3. | | | |

| Cyrix Engineer | Date | Start Time | End Time |
|------------------|---------|------------|----------|
| Shankrishnan PPK | 18-1-24 | 11:30am | 12:30pm |

Customer Remark **Completed** **Pending**

Service Engineer Name : Shankrishnan PPK
Signature :
Date : 18-1-24
Contact Number : 8281188574

Customer Name : Nareera P
Signature :
Date : 18/1/24
Contact Number : 9747204172
Designation : PHN
Hospital Seal :




**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA


| Sl. No | Particulars | Details |
|--------|--|---|
| 1 | Name of District | KANNUR |
| 2 | Name of Hospital | PHC KANICHAR |
| 3 | Name of Equipment with Make, Model and Serial Number | EQUIPMENT NAME: DIGITAL BP APPARATUS Make : OMRON Model: hem8712 SN: N/A |
| 4 | Equipment ID/ Barcode | 1344379 #91376 |
| 5 | Date of purchase/ Year of manufacture/Installation Date | 29-08-2019 |
| 6 | Warranty details (Yes/No) | No warranty |
| 7 | *AMC/ CAMC Period agreed at the time of purchase | No AMC/CAMC |
| 8 | Date of breakdown (Date of registration of complaint through email/ Toll free) | Toll free -17-01-2024 |
| 9 | Action taken | Checked the equipment and found BP reading error found problem with main board and pressure sensor. Enquired for spares |
| 10 | Present status of the equipment (Fully damaged / partially damaged) | Fully damaged |
| 11 | Recommendations for repair (required service details) | No recommendations |
| 12 | Cost of spares (specify parts and cost) | |



| | | |
|----|--|---|
| 13 | Asset Value | 2200/- |
| 14 | *Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value | |
| 15 | Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) | Cyrix Service report attached |
| 16 | Reasons for recommending the equipment as BER | Checked the equipment and found BP reading error found problem with main board and pressure sensor. Equipment installed on 29-08-2019 aged up to 4 years and 5 months. Checked the spares in market, but spares are not available because it's a low asset cost equipment. So we recommending for condemnation. |
| 17 | Name & Signature of CYRIX Authority | SAYANTH C  |

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
checked & identified that main board & pressure sensor are defective. Aged up to 4 1/2 years. its spares not available below asset value. So recommended for BER as per BER protocol.

 01/02/2024 *Signature of JC BM (NHM)

CONSULTANT BIOMEDICAL
 NATIONAL HEALTH MISSION
 KANNUR - 670002

Date: 


 Signature of
 Superintendent/Medical Officer (i/c)
 കേരളാർത്ഥശാസ്ത്ര വിഭാഗം
 കേരളാർത്ഥശാസ്ത്ര വിഭാഗം

കേരളാർത്ഥശാസ്ത്ര വിഭാഗം, P.O, കൊല്ലം - 670679



27

NATIONAL RURAL
District Health and Family
Cash/Bank Book

| Date | Voucher No. | Particulars | | |
|---|-------------|--|-------|----------|
| | | | WTISC | UTF |
| 5/12/17 | 2/17-18 | Food - SPC students } Town cleaning purpose } | | |
| | | Total | | |
| | | Balance in Bank | 1708 | 10,000/- |
| <p>At the instance of AG Audit on 8/2/18 the interest amount credited in Bank pass Book was calculated. As per the Pass Book the upto date interest as on 1/11/17 was Rs-1104/- whereas as per the Cash Book the interest amount was Rs-987/. The difference was Rs-117/ the difference amount was Rs-117/- was occurred due to debit of Bank charges of Rs-117/- By the Bank during the period from - 11/9/17 to 25/7/18</p> | | | | |
| 1/2/18 | | No transaction from 06/12/17 to 12/3/18 | | |
| 13/3/18 | | | 1600 | 8508 |
| | 7/11-18 | Nutrition day - ASHA | 400 | } |
| | 8/11-18 | Nutrition day - JPHN | 400 | |
| | 9/11-18 | Communication allowance | 800 | |
| | | Total | 1600 | |
| 10/2/18 | 11/13-18 | Digital weighing machine Digital BP monitor | | |



Signature
District Health and Family
670679



Handwritten signature

சென்னை மருத்துவ கல்லூரி
மருத்துவ, P.O. சாலை எண் - 670679