


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

421

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Invandrum,
2	Name of Hospital	
3	Name of Equipment with Make, Model and Serial Number	TH malayinkeezhu Nebulizer. PHGPS. Respronic. 1023510365.
4	Equipment ID & Barcode	89551 - 0121922
5	Date of purchase / Year of manufacture / Installation Date	21/1/21.
6	Warranty details (Yes/No)	No.
7	*AMC/ CAMC Period agreed at the time of purchase	Nil.
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	8/1/24.
9	Action taken	Checked the machine and found that piston brokened from motor.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	Checked the machine and found that piston brokened from the motor and motor failure. The piston can't be refixed. It can't be repair economically.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	1258/-
14	" Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Attached.
16	Reasons for recommending the equipment as BER	checked the machine -> Piston broken and from the motor and it can't be repaired. The piston is not available as a spare in the stock and it can't be repaired economically. Therefore proposing it for RBER.
17	Name & Signature of CYRIX Authority	Poornu P.S. 


\*Not mandatory **#Based on the period of life and value as per the BER guidelines**

\* **Attach Photograph**


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Piston broken and can't be repaired by gluing. So can be RBER.

**MANEESHA MOHAN M.R.**  
 Junior Consultant (Biomedical)  
 National Health Mission  
 Thiruvananthapuram

  
 16-01-2024  
 Signature of JC BM (NHM)

Date



  
 Superintendent / Medical Officer (i/c)  
 Superintendent  
 Headquarters Hospital  
 Malavinkil



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

No. : 158321

421

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

### Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>TH</u></p> <p>Address : <u>malayinkeeyba</u> <u>Toward sum.</u></p> <p>Ph : <u>98397 36991.</u></p>	<p>Call Registration Date : <u>8/1/24.</u></p> <p>Caller ID : <u>89551.</u></p> <p>Date of Visit : <u>9/1/24.</u></p> <p>Asset No. : <u>0121922.</u></p> <p>EQPT Name : <u>Nebulizer</u></p> <p>Manufacture <u>Philips.</u> Model <u>Respironics.</u></p> <p>S. No. <u>102351036</u> Dept. <u>Ward.</u></p>
---	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** not working.

**Action Taken :** Checked the machine and found that piston brokened from the motor. It can't be repair and it is not available as a spare in the market. It can't be repair temporarily.

Completed  Date : 9/1/24 Time : 12pm. Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Poornu.</u>	<u>9/1/24.</u>	<u>11:15am</u>	<u>12pm.</u>

Customer Remark Completed  Pending

<p>Service Engineer Name : <u>Poornu PS</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>9/1/24.</u></p> <p>Contact Number : <u>90245 70652.</u></p>		<p>Customer Name : <u>THOM, Mr</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>9/1/24</u></p> <p>Contact Number : <u>N/O Gr I</u></p> <p>Designation : <u>N/O Gr I</u></p> <p>Hospital Seal :</p>
--	--	--



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669



BAR CODE - (8004-890615225) 0121922

മേശുകുടെ നസിഡൻസ്  
നാഷണലിറ്റേഷൻ (MRA)

**CYRIX**  
HEALTHCARE PVT LTD  
S.N: 1023510365

CN: THM006  
PRO : Nebulizer  
ID : 0121922  
CAL.ON : 29-12/2023  
CAL.DUE : 28-12/2024



ISO 9001 2008 Certified Co. Ph: 9744555073

വത് T477/12  
റാക്ടർ സ്മാരക മന്ദിരം  
ഇംഗിഷൻ, മേശുകുടെ,  
കിരൂവനത്തപുരം - 695571

102



1095063

InnoSpire Essence (EU Plug)

230V ~ 50Hz 1A 30 min I / 30 min O

**SN** 1023510365

**LOT** 170513

CE 0123



(01) 0 0383730 00188 3  
(10) 170513  
(21) 1023510365



IP21



Respironics Respiratory Drug Delivery (UK) Ltd  
Chichester Business Park, City Fields Way  
Tangmere, Chichester PO20 2FT UK  
Made in China

1108975 R01

WARR



