



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 143782

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>THQH Koyilandy</u>	Call Registration Date : ... <u>8/1/24</u>
Address <u>Koyilandy</u>	Caller ID : ... <u>89467</u>
..... <u>Kozhikode</u>	Date of Visit : ... <u>10/1/24</u>
Ph : <u>9847805166</u>	Asset No. : ... <u>1150841</u>
	EQPT Name : ... <u>Digital weighing machine</u>
	Manufacture ... <u>Samsa</u> Model : ... <u>Digital</u>
	S. No. ... <u>NA</u> Dept. ... <u>Casualty</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :
..... Machine not getting on

Action Taken : Checked and found that main board and display
defective. Need to replace these spares for further checking working
condition of the equipment.

Completed Date : 16/1/24 Time : 11:30AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Akhil Raj</u>	<u>16/1/24</u>	<u>10:30AM</u>	<u>11:30AM</u>

Customer Remark Completed Pending
Recommended for condemnation Sindhu D.K Senior Nursing Officer
Ph: 9847805166

Service Engineer Name : Akhil **Customer Name :**
Signature : [Signature] **Signature :** [Signature]
Date : 16/1/24 **Date :**
Contact Number : 8943677676 **Contact Number :**
Hospital Seal : KOYILANDY, 67305 **Hospital Seal :** [Seal]


SUPERINTENDENT
Head Quarters Hospital
Koyilandy

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOZHIKODE
2	Name of Hospital	THQH KOYILANDY
3	Name of Equipment with Make, Model and Serial Number	Equipment : Digital Weighing Machine Make : Samsco Model :NA SN : NA
4	Equipment ID & Barcode	89467 & 1150841
5	Date of purchase / Year of manufacture /Installation Date	18-11-2020
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	08-01-2024 (Toll free)
9	Action taken	Checked and found main board and display defective.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	Rs . 1897/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found main board and display defective. Equipment installed on 18-11-2020 and covered upto 3+ years. Quotation not submitted since spares are not available in the market .So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AKHIL RAJ 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Checked & found that main board and display damaged. Spares needs to replaced. Machine installed on 18/11/2020. spares are not available in the market. So recommended for condemnation.

Signature of JC BM (NHM)
Consultant Bio Medical Engineer
National Health Mission

Kozhikode

Recommended for Condemnation



SUPERINTENDENT
Taluk Head Quarters Hospital
Koyilandy

Signature of
Superintendent / Medical Officer (i/c)

Date




STOCK BOOK OF

Name of article WEIGHING MACHINE

Date	Invoice No.	From whom received or to whom issued	Receipt	Issued	Balance	Batch No Exp. Date Mfg. Date	Price per Unit/MFG	Remarks
11/1/18		Brought Forward BF from old stocks books Page No. 136 Rd from the store 12/9/13 - Rs 818/- 1 municipal ward 24/10/17 - Rs 2006/- 2 1/10/13 - B. 818/- 1 Issued to use.			4			Put in General Expense municipal ward
16/3/19		Received from the store - 1 No Delta Table TOP Digital weighing scale Capacity 20kg - minim 100gm -			5 Nos		Warranty, 14/3/19 - 13/3/20 Rs - 5888/-	RY HW.
16/3/19		Issued to use (Paediatric OP)						
18/11/2020		Received from Store DIGITAL - ADULT. Scheme - MLA fund 2020-2021. Supplier: Dr. Bellaram's Sons. Batch no: 111. Rs: 1897.49/- (1).	1		6			→ Paid orn.
22/3/2022		To out to unscrutable register VOL - I Page no: - 143 Rd 12/9/13, 1/10/13		2	4 Nos -			Prabhu



SUPERINTENDENT
Taluk Head Quarters Hospital
Koyilandy



[Signature]
SUPERINTENDENT
Head Quarters Hospital
Koyilandy