

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

4/14

Recommendations for Beyond Economic Repair (BER)

PROFORMA

| Sl. No | Particulars | Details |
|--------|--|---|
| 1 | Name of District | Tarandrum. |
| 2 | Name of Hospital | |
| 3 | Name of Equipment with Make, Model and Serial Number | THOM Malay mkeezhu ECC machine, CURA. C6-3C F712180108BC |
| 4 | Equipment ID & Barcode | 86588-0121277. |
| 5 | Date of purchase / Year of manufacture / Installation Date | 13/8/2019. |
| 6 | Warranty details (Yes/No) | No. |
| 7 | *AMC/ CAMC Period agreed at the time of purchase | nil. |
| 8 | Date of breakdown (Date of registration of complaint through email/ Toll free) | 20/12/23. |
| 9 | Action taken | checked the machine printer and main board damaged. |
| 10 | Present status of the equipment (Fully damaged / partially damaged) | fully damaged. |
| 11 | Recommendations for repair (required service details) | checked the machine and found that printer head and main board damaged. The spares are not available to purchase because of OEM support is not available |
| 12 | Cost of spares (specify parts and cost) | NA |

| | | |
|----|--|--|
| 13 | Asset Value | 28,000 - |
| 14 | [#] Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value | NA |
| 15 | Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) | Attached. [Copy service report] |
| 16 | Reasons for recommending the equipment as BER | Checked the machine and found that printer and main board damaged. Spares are not available OEM not supporting for service as per tender condition as per tender clause 5.3.14.2. It can't be repaired economically. Hence propose for RBER. |
| 17 | Name & Signature of CYRIX Authority | Poornu. D.S. [Signature] |

*Not mandatory #Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

As per clause 5.3.14.2, it can't be repaired as company not present.

MANEESHA MOHAN M.R.
 Junior Consultant (Biomedical)
 National Health Mission
 Thiruvananthapuram

[Signature]
 3-1-2024
 Signature of JC BM (NHM)



Date

[Signature]
 Signature of
 Superintendent / Medical Officer (i/c)
Superintendent
 Taluk Headquarters Hospital
 Malayinkil



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 158284

AM

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

| | |
|-----------------------------|--|
| Health Facility <i>THOH</i> | Call Registration Date : <i>20/12/23</i> |
| Address <i>malayinkeeyu</i> | Caller ID : <i>86888</i> |
| <i>Tandram</i> | Date of Visit : <i>21/12/23</i> |
| Ph : <i>9246 14142</i> | Asset No. : <i>0121877</i> |
| | EQPT Name : <i>ECG machine</i> |
| | Manufacture <i>CARD</i> Model : <i>CG3C</i> |
| | S. No. <i>1912180/08</i> Dept. <i>Casualty</i> |
| | <i>EC</i> |

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : *not working*

Action Taken : *checked the machine and found that printer failure, main board damaged. Spares are not available. OEM not supporting for spare as per tender condition tender clause 5.3.14.2. It can't be repair economically.*

Completed Date : *21/12/23*, Time : *1:45pm* Spare Required

Spare Replaced Requested

| Description | Qty. | Part Number | PR Number |
|-------------|------|-------------|-----------|
| 1. | | | |
| 2. | | | |
| 3. | | | |

| Cyrix Engineer | Date | Start Time | End Time |
|----------------|-----------------|------------|---------------|
| <i>Pranav</i> | <i>21/12/23</i> | <i>1pm</i> | <i>1:45pm</i> |

Customer Remark : Completed Pending




| | |
|--|--------------------------------------|
| Service Engineer Name : <i>Pranav D.</i> | Customer Name : <i>Srinivasan. K</i> |
| Signature : <i>[Signature]</i> | Signature : <i>[Signature]</i> |
| Date : <i>21/12/23</i> | Date : <i>21/12/23</i> |
| Contact Number : <i>90975 20652</i> | Contact Number : <i>7510 235972</i> |
| | Designation : <i>Sdo.</i> |
| | Hospital Seal : |


070

CURA




CYRIX MEDICAL EQUIPMENT MAINTENANCE
 PROGRAMS (PAPER)
 TOLL FREE NO: 1800 - 425-7669
 BAR CODE - (8004-890815225)
 0121877



ERROR
 READY




 0/0




 CYRIX
 MEDICAL EQUIPMENT MAINTENANCE
 PROGRAMS (PAPER)

DDO





SN F712180108N022E  

 2017-06-13  

Model: CG-3c  

Power: AC 100V 240V 50/60Hz/50/60W 5VA

Name: Digital Electrocardiograph 

 This machine must only be operated by a qualified person(s) who are familiar with the product safety instructions.