



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 172976

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrx.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>Wec Ponnani</u>	Call Registration Date : <u>6-1-24</u>
Address <u>Ponnani</u>	Caller ID : <u>89212</u>
<u>MAPAPPURAM (DL) - KERALA</u>	Date of Visit : <u>10-1-24</u>
Ph : <u>8156994634</u>	Asset No. : <u>1011192</u>
	EQPT Name : <u>Weighing Machine</u>
	Manufacture <u>ROSSMAX</u> Model : <u>Electronic</u>
	S. No. <u>NA</u> Dept. <u>Post Natal</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : weighing machine Display blinking.

Action Taken : checked the weighing machine and found main board, load cell and display defective condition. Need to replace these spare for checking working condition of equipment.

Completed Date : 15/1/24 Time : 2-15 pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>Chm Krishnan VP</u>	<u>15/1/24</u>	<u>1-40 pm</u>	<u>2-15 pm</u>

Customer Remark Completed Pending

Dr. ASHA.R.K
SUPERINTENDENT
W & C HOSPITAL PONNANI
PO PONNANI SOUTH - 679 544



Service Engineer Name : Chm Krishnan VP
Signature : [Signature]
Date : 15/1/24
Contact Number : 9809643504

Customer Name : [Signature]
Signature : [Signature]
Date : 15/1/24
Contact Number : 8156994634
Designation : Nursing Officer
Hospital Seal :


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	MALAPPURAM
2	Name of Hospital	WOMEN & CHILD HOSPITAL PONNANI
3	Name of Equipment with Make, Model and Serial Number	WEIGHING MACHINE ROSS MAX WE-300
4	Equipment ID & Barcode	1011192
5	Date of purchase / Year of manufacture / Installation Date	21/5/2019
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/ CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	6-1-2024 - TOLL FREE
9	Action taken	CHECKED AND FOUND MAIN BOARD LOAD CELL AND DISPLAY DEFECTIVE
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NEED TO REPLACE MAINBOARD, LOAD CELL AND DISPLAY
12	Cost of spares (specify parts and cost)	



13	Asset Value	4990/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	-
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT ATTACHED.
16	Reasons for recommending the equipment as BER	CHECKED THE MACHINE AND FOUND MAINBOARD, LOADCELL AND DISPLAY DEFECTIVE. THE EQUIPMENT INSTALLED ON 21/5/2019 SINCE SPARE ARE NOT AVAILABLE IN THE MARKET, SO WE RECOMMEND THE EQUIPMENT FOR CONDEMNATION, SINCE IT IS A LOW COST EQUIPMENT
17	Name & Signature of CYRIX Authority	 Chinnirishnan VP

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Verified with the damage of loadcell display & main board. Beyond economic repair.


Signature of JC BM (NHM)

Date




Signature of
SUPERINTENDENT
W & C HOSPITAL, PONNANI
P.O. PONNANI SOUTH - 679 586

Electronic wt machine (Neonatal) 43
 (Crescent)


GOVT WOMEN & CHILDREN'S HOSPITAL, PONNANI
INSTRUMENT STOCK REGISTER

S.No	Date	Name	Received	Issued	Conde- mination	Price/ Specification	Sign of S/W/MO	Bal- ance
						RS 42297.44	Signature BME	1
	21/5/19	Rd from store SISK P-no: 185	1					
	21/5/19	Issued to ward		1				

Signature
BME NMM



Handwritten text at the top of the page, possibly a name or date, which is mostly illegible.



Handwritten signature and the text 'BME W/2019'.

