


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Tiruvandrum.
2	Name of Hospital	UPHC VATHYOOKKRAVU.
3	Name of Equipment with Make, Model and Serial Number	Centrifuge. Remi C-254/S. ZBCN-08451.
4	Equipment ID & Barcode	75482 - 0140841.
5	Date of purchase / Year of manufacture /Installation Date	21/05/2015
6	Warranty details (Yes/No)	No.
7	*AMC/ CAMC Period agreed at the time of purchase	Nil.
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	18/10/2023.
9	Action taken	Checked the machine it can't be repair economically.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	Checked the machine and found that motor failure intensity controller damaged. It can't be repair economically.
12	Cost of spares (specify parts and cost)	It can't be repair economically.


13	Asset Value	96511-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Attached.
16	Reasons for recommending the equipment as BER	checked the machine and found that problems with the motor and intensity controller. the motor was recently rewinded and got faulty again. Lid is mildly rusted also. As the major parts are faulty equipment is proposed for RBER.
17	Name & Signature of CYRIX Authority	

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
 Motor and intensity controller Complaint
 So RBER


MANEESHA MOHAN M.R.
 Junior Consultant (Biomedical)
 National Health Mission
 Thiruvananthapuram


 25-10-2023

Signature of JC BM (NHM)

Date 18/10/23



Signature of 
 Superintendent Medical Officer (i/c)
 VATTIYURKAVU, TPM



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 146550

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>UPHC</u></p> <p>Address <u>Vattiyurkavu</u> <u>Tiruvandur.</u></p> <p>Ph : <u>97479 64083</u></p>	<p>Call Registration Date : <u>18/10/23</u></p> <p>Caller ID : <u>75482</u></p> <p>Date of Visit : <u>18/10/23</u></p> <p>Asset No. : <u>0140241</u></p> <p>EQPT Name : <u>Centrifuge</u></p> <p>Manufacture <u>Remi</u> Model : <u>C-087/2</u></p> <p>S. No. <u>ZBCN-0845</u> Dept. <u>Lab.</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : motor failure.

Action Taken : checked the machine and found that problems with the motor and intensity controller. The motor was recently rewinded and got faulty again. Lid lock is mildly rusted also. As the major parts are faulty equipment is prepared for RBER.

Completed Date : 18/10/23 Time : 3pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. _____	_____	_____	_____
2. _____	_____	_____	_____
3. _____	_____	_____	_____

Cyrix Engineer	Date	Start Time	End Time
<u>Parav.</u>	<u>18/10/23</u>	<u>9-30</u>	<u>3pm.</u>

Customer Remark Completed Pending



<p>Service Engineer Name : <u>Parav.P.S.</u></p> <p>Signature : _____</p> <p>Date : <u>18/10/23</u></p> <p>Contact Number : <u>9074570652</u></p>	<p>Customer Name : <u>Fehima S</u></p> <p>Signature : _____</p> <p>Date : <u>18/10/23</u></p> <p>Contact Number : <u>730 6503707</u></p> <p>Designation : <u>Lab technician</u></p> <p>Hospital Seal : _____</p>
---	--

