

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

130

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Ttirvandrum
2	Name of Hospital	GH Ttirvandrum
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : BP Apparatus Model : Delux Manufacture : Diamond.
4	Equipment ID & Barcode	0110616
5	Date of purchase / Year of manufacture /Installation Date	12/6/2010
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AME/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	2/1/2024
9	Action taken	checked the BP Apparatus Mercury stand broken and not stable to stand in Table.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	Mercury stand broken and not stable to stand in table This condition no measuring BP in patient.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	1323
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report and photos attached.
16	Reasons for recommending the equipment as BER	Mercury stand is broken no measuring a BP in patient and not stable to stand in table.
17	Name & Signature of CYRIX Authority	Josekutty George 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Mercury stand broken. Machine is not stable to stand. So can be RBER.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


22-01-2024

Signature of JC BM (NHM)

Date




Superintendent
General Hospital
Thiruvananthapuram
Superintendent / Medical Officer (i/c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 160962

n30

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>G.H. Trivandrum</u></p> <p>Address <u>Trivandrum</u></p> <p>Ph : <u>9961391392</u></p>	<p>Call Registration Date : <u>2/1/2024</u></p> <p>Caller ID : <u>88439</u></p> <p>Date of Visit : <u>3/1/2024</u></p> <p>Asset No. : <u>0110616</u></p> <p>EQPT Name : <u>BP Apparatus</u></p> <p>Manufacture <u>Diamond</u> Model : <u>delux</u></p> <p>S. No. <u>Nil</u> Dept. <u>Ward - 1</u></p>
--	---

Service Classification : Breakdown Call **PMS** **Calibration** **Cust.Training**

Problem Identified : BP Apparatus Broken

Action Taken : checked the BP Apparatus not for repair this complaint BP Apparatus stand Broken and not stable to stand in BP Apparatus in table.

Completed **Date :** 3/1/2024 **Time :** **Spare Required**

Spare Replaced **Requested**

No.	Description	Qty.	Part Number	PR Number
1.	<u>RBER</u>			
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time

Customer Remark Not stable in position **Completed** **Pending**

<p>Service Engineer Name : <u>Josekutty George</u></p> <p>Signature : <i>[Signature]</i></p> <p>Date : <u>3/1/2024</u></p> <p>Contact Number : <u>8848230407</u></p>	<p>Customer Name : <u>Celine Leena</u></p> <p>Signature : <i>[Signature]</i></p> <p>Date : <u>3/1/24</u></p> <p>Contact Number : <u>9961391392</u></p> <p>Designation : <u>Senior Nursing Officer</u></p> <p>Hospital Seal :</p>
--	--



