


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

431

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Tiruvandrum
2	Name of Hospital	GH Tiruvandrum
3	Name of Equipment with Make, Model and Serial Number	Equipment Name: Nebulizer Make: Ready Mist SN: 196630
4	Equipment ID & Barcode	0111780
5	Date of purchase / Year of manufacture / Installation Date	7/11/2013
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	5/1/2024
9	Action taken	Checked the machine Motor, piston are damaged and motor fully rusted.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Motor, piston are damaged and motor fully rusted.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	1353
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report and photos attached.
16	Reasons for recommending the equipment as BER	Nebulizer main parts are damaged and fully rusted (motor and piston). This is the reasons for recommending the equipment as RBER.
17	Name & Signature of CYRIX Authority	Josekutty George 


*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* **Attach Photograph**


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Motor and Piston damaged.
Can be RBER.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


22-01-2024
Signature of JC BM (NHM)

Date



Signature of
Superintendent Medical Officer (I/c)
23/1/24
Superintendent
General Hospital
Thiruvananthapuram



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 160979

31

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>GH Trivandrum</u></p> <p>Address <u>Trivandrum</u></p> <p>Ph : <u>9496684098</u></p>	<p>Call Registration Date : <u>5/1/2024</u></p> <p>Caller ID : <u>88957</u></p> <p>Date of Visit : <u>6/1/2024</u></p> <p>Asset No. : <u>0111780</u></p> <p>EQPT Name : <u>Nebulizer</u></p> <p>Manufacture <u>Ready Mist</u> Model : <u>Nil</u></p> <p>S. No. <u>196630</u> Dept. <u>Ward-9</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Nebulizer not working.

Action Taken : Checked the machine Nebulizer motor fully rusted and Motor not working, piston broken. This machine can not repairing.

Completed Date : 8/1/2024 Time : 11pm Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.	RB ER			
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Josekutty George</u>	<u>8/1/2024</u>	<u>10:30pm</u>	<u>11pm</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Josekutty George</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>8/1/2024</u></p> <p>Contact Number : <u>8848230407</u></p>	<p>Customer Name : <u>Mercy Jant. wd. 9</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>8/1/24</u></p> <p>Contact Number : <u>9074548508</u></p> <p>Designation : <u>Senior Nurse officer</u></p> <p>Hospital Seal : </p>
--	--

