


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

11/24

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Thiruvananthapuram
2	Name of Hospital	CHC Kanyakumbham.
3	Name of Equipment with Make, Model and Serial Number	Nebuliser Ultraneb. BI/NB/077
4	Equipment ID & Barcode	284578 0130526
5	Date of purchase / Year of manufacture / Installation Date	04.01.2020.
6	Warranty details (Yes/No)	No
7	*AMC/ CAMC Period agreed at the time of purchase	No
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	02.01.2024
9	Action taken	checked the nebuliser found motor & piston changed
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	checked the nebuliser found piston & motor changed its economically not unserviceable
12	Cost of spares (specify parts and cost)	No

13	Asset Value	1353/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Attached
16	Reasons for recommending the equipment as BER	checked the nebuliser found Piston changed & motor burned its service economically not viable. so recommending the equipment as BER
17	Name & Signature of CYRIX Authority	


*Not mandatory #Based on the period of life and value as per the BER guidelines

* **Attach Photograph**


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Piston and motor damaged.
So can be RBER as per norms.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Institute of Biomedical Engineering & Technology
Thiruvananthapuram


12-1-2024
Signature of JC BM (NHM)

Date _____ Seal _____


Signature of Superintendent Medical Officer (i/c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 173772

24

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>Community Health Center</u></p> <p>Address <u>Kangasikumbam Thiruvananthapuram</u></p> <p>Ph : <u>9446968196</u></p>	<p>Call Registration Date : <u>02.01.2024</u></p> <p>Caller ID : <u>88457</u></p> <p>Date of Visit : <u>03.01.2024</u></p> <p>Asset No. : <u>0130524</u></p> <p>EQPT Name : <u>Nebuliser</u></p> <p>Manufacture <u>Ultraneb</u> Model : <u>BI/NA/077</u></p> <p>S. No. <u>NA</u> Dept. <u>OP</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Found motor & Piston damaged

Action Taken : checked the nebuliser found motor & Piston damaged
its service economically do not viable

Completed Date : 03.01.2024 Time : 2.00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Mohammed Sherif</u>	<u>03.01.2024</u>	<u>1.30pm</u>	<u>2.00pm</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Mohammed Sherif</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>03.01.2024</u></p> <p>Contact Number : <u>7994042445</u></p>	<p>Customer Name : _____</p> <p>Signature : <u>[Signature]</u></p> <p>Date : _____</p> <p>Contact Number : _____</p> <p>Designation : _____</p> <p>Hospital Seal : <u>0472-2831209</u></p>
---	--

Nebulizer.

116


Date	To whom issued	Received	Issued	Balance
2/5/18	Received from KROCEL (supplied as per order No. 4/NPHCE/2017, dtd 26/3/2017) Mfr Trucheck. (X-ceed Health Care) Model N8001	2 Nos		2 Nos
3/5/18	Issued to S/N		2	nil
4/11/20	Received from DDWA GVM, 19-20. Mfr SANNOFI	1 Nos		1 Nos.
19/01/20	Issued to staff nurse		1	Nil
18/7/21	Received from Fut store NIM project without invoice Mfr Qty Angel 2			2
18/7/21	Issued to S/N		2	Nil

Sub:
 SRI:
 page:

Stronics



CYRIX HEALTH CARE PVT.LTD
BEMP SERVICE PROVIDER
PREVENTIVE MAINTENANCE SERVICE
PMS DONE ON
NEED SERVICE
PERFORMED BY



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE
PROGRAMME (perip)



TOLL FREE NO: 1800 - 425-7669

**BAR CODE - (8004-890615225)
0130524**





Ultraneb Nebulizer

Model

BI/NB/077

Serial No

UN/02155

LOT

2010077

Power

160VA

Input

230V/50HZ

