



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 164143

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

| | |
|-------------------------------------|--|
| Health Facility <u>TEOH ALATHUR</u> | Call Registration Date : <u>02/01/24</u> |
| Address <u>ALATHUR</u> | Caller ID : <u>86500</u> |
| <u>PALAKKAD</u> | Date of Visit : <u>04/01/24</u> |
| Ph : <u>946139336</u> | Asset No. : <u>0920919</u> |
| | EQPT Name : <u>BP APPARATUS</u> |
| | Manufacture <u>NIL</u> Model : <u>NIL</u> |
| | S. No. <u>NIL</u> Dept. <u>LABOUR ROOM</u> |

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Machine not working

Action Taken : Checked the machine. Found that Display, main board and Pump motor is faulty. Also found that NIBP wiff leakage. Need to replace these spare for further checking working condition of equipment.

Completed Date : Time : Spare Required

Spare Replaced Requested

| Description | Qty. | Part Number | PR Number |
|-------------|------|-------------|-----------|
| 1. | | | |
| 2. | | | |
| 3. | | | |

| Cyrix Engineer | Date | Start Time | End Time |
|----------------|------|------------|----------|
| <u>DVIN'B</u> | | | |

Customer Remark Completed SUPERINTENDENT T.H. PENDING PITAI ALATHUR

Service Engineer Name : DVIN'B
Signature : [Signature]
Date :
Contact Number : 75938 47155

Customer Name : Shreeja - T-1
Signature : [Signature]
Date : 23/1/24
Contact Number : 9947488493
Designation : N-0 401
Hospital Seal

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

| Sl. No | Particulars | Details |
|--------|---|---|
| 1 | Name of District | PALAKKAD |
| 2 | Name of Hospital | THOH ALATHUR |
| 3 | Name of Equipment with Make, Model and Serial Number | BP APPARATUS MAKE: NIL MODEL: NIL SN : NIL |
| 4 | Equipment ID & Barcode | 88500 , 0920919 |
| 5 | Date of purchase / Year of manufacture /Installation Date | 3/01/2020 |
| 6 | Warranty details (Yes/No) | No Warranty |
| 7 | *AMC/ CAMC Period agreed at the time of purchase | NO AMC / NO CAMC |
| 8 | Date of breakdown(Date of registration of complaint through email/ Toll free) | 02/01/2024, TOLL FREE |
| 9 | Action taken | Checked and found that display, main board, NIBP cuff and 4 number of battery faulty. |
| 10 | Present status of the equipment (Fully damaged / partially damaged) | Fully damaged |
| 11 | Recommendations for repair (required service details) | Not recommending |
| 12 | Cost of spares (specify parts and cost) | NA |

| | | |
|----|--|--|
| 13 | Asset Value | 2200 |
| 14 | # Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value | NA |
| 15 | Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) | CYRIX service report |
| 16 | Reasons for recommending the equipment as BER | Checked and found that LED display main board, MIBP out and four numbers of batteries faulty. Machine installed on 31/10/20 and outlive 4 years. Spare quote not submitted spare are not available in market and since recommending for condemnation |
| 17 | Name & Signature of CYRIX Authority | D. N. K. D. |

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Cuff, bladder, display damaged hence recommended for BER.



Asst
27/01/24
Signature of JC BM (NHM)



K. Anand
SUPERINTENDENT
T.H.O. HOSPITAL
ALATHUR
Signature of
Superintendent / Medical Officer (i/c)

Date

Seal

| | | |
|----|---------------------------------------|-----|
| | Curette Sharp & Blunt | 173 |
| | Forceps uterine | 5 |
| 3 | Forceps Chantal | 7 |
| 4 | Forceps craniotomy & Handle | 9 |
| 5 | Forceps dissecting Plane | 11 |
| 6 | Forceps dissecting Hooked | 13 |
| 7 | Forceps introductory Miller's Murray | 15 |
| 8 | Forceps Spring Holding | 17 |
| 9 | Forceps Alice tissue | 19 |
| 10 | Hook Blunt & Excise | 21 |
| 11 | Perforator Cranium | 23 |
| 12 | Scissors Curved on flat 6" | 25 |
| 13 | Scissors dressing st | 27 |
| 14 | Scissors Mayo st 5 1/2" | 29 |
| 15 | Scissors Sharp | 31 |
| 16 | Uterine Sound | 33 |
| 17 | Speculum vaginal | 35 |
| 18 | Basin & I 8" | 37 |
| 19 | Basin & I 10" | 39 |
| 20 | Basin Wash Shallow S.S. | 41 |
| 21 | Bed pan S.S. | 43 |
| 22 | Bed Stead Iron | 45 |
| 23 | Bed Stead Maternity | 47 |
| 24 | Bed Stead Delivery | 49 |
| 25 | Bowl & I 6" | 51 |
| 26 | Bowl S.S. (Sponge Bowl) | 53 |
| 27 | Breast Pump | 55 |
| 28 | Bucket Cr. I Large + Small | 57 |
| 29 | Can Irrigator 4 pint R & L | 59 |
| 30 | Can Irrigator 2 pint L & I | 61 |
| 31 | Catheter metal male | 63 |
| 32 | Catheter Tracheal & glass mucous trap | 65 |
| 33 | Cup feeding | 67 |
| 34 | Cup Solution | 69 |
| 35 | Dressing drum Shallow 9x9 | 71 |
| 36 | Dressing drum 9x9" | 73 |
| 37 | Jar Dressing S.S. | 75 |
| 38 | Lamp dusk Inspection | 77 |
| 39 | Locker Bed Side | 79 |
| | | 81 |

Number of articles
 127
 127

Stores
 127

SUPERINTENDENT
 T.H.O. HOSPITAL
 ALATHUR



| | | |
|----|---|-----|
| 41 | mag E1 Half Pint | 83 |
| 42 | measure Graduated 100cc S.S. | 85 |
| 43 | nozzle vaginal vulconale | 87 |
| 44 | Pad Lock | 89 |
| 45 | Picture Water S.S. | 91 |
| 46 | Razor Safety | 93 |
| 47 | Scales Infant Weight No. 6 | 95 |
| 48 | Weighing machine Adult | 97 |
| 49 | Stand Irrigating | 99 |
| 50 | Stove Single Burner | 101 |
| 51 | All glass Syringe 5cc | 105 |
| 52 | All glass Syringe 2cc | 107 |
| 53 | Thermometer | 109 |
| 54 | Valsalva forceps | 111 |
| 55 | Petrol med | 113 |
| 56 | Tray Center for Dressing | 115 |
| 57 | Tray Instrument Shallow R-I | 117 |
| 58 | Tray Dressing R-I | 119 |
| 59 | Tray Instrument & Cover E-D | 121 |
| 60 | Tray Instrument & Cover S-S-14 1/2 x 10 1/2 | 123 |
| 61 | Tray kidney S-S | 125 |
| 62 | Trolley Dressing | 127 |
| 63 | Engine nozzle Adult | 129 |
| 64 | Stove & Double burner | 131 |
| 65 | Sphygmomanometer | 133 |
| 66 | Forceps midwifery | 135 |
| 67 | Vacuum Extractor | 137 |
| 68 | evum Forceps | 139 |
| 69 | Forceps Forceps | 141 |
| 70 | Electric sterilizer | 143 |
| 71 | Aluminum measure | 145 |
| 72 | Janetka stove | 147 |
| 73 | Bucket 10L | 149 |
| 74 | Basin E1 | 151 |
| 75 | E1 Tray & Lid | 153 |
| 76 | Baby Cradle | 155 |
| 77 | P.V.C knife tray 40cc | 157 |
| 78 | Director set | 159 |
| 79 | O. Cylinder stand | |



SUPERINTENDENT
T.H.V. HOSPITAL
ALATHUR


BP apparatus

132

| Date | Description | Cost | Rel | In | Out | Bal | ms |
|----------|--|------|-----|----|-----|-----|----|
| 23/1/18 | Received from store P. Padiyathil Digital (Padiyathil journal) Main Stock Book No 62 | | 1 | | 1 | 1 | ✓ |
| 23/1/18 | Rs. 2000/- Issued to ward | 2000 | | 1 | | 11 | |
| 9/2/18 | Tor. out 111 Page no | | | 1 | | 10 | ✓ |
| 2/7/18 | Tor. out 8w (1) (Digitized) Page no 133 | | | 1 | | 9 | ✓ |
| 26/1/18 | Rel from the store (Digitized) | | 1 | | | 10 | ✓ |
| 3/1/2020 | Rel from the store manual | 3000 | ✓ | 1 | | 11 | ✓ |
| 3/1/2020 | Rel from the store Led. | 2383 | ✓ | 1 | | 12 | ✓ |
| | Tor. out 6 unservicable register Page no 2 | | | 7 | | 5 | ✓ |
| 17/4/21 | 3 digital BP Apparatus Tr. Out to New Stock book B Page No. 65 & 65 | | | 3 | | 2 | ✓ |
| 17/4/21 | 2 manual BP apparatus Tr. out to Page: No. - 66 of volume code B II | | | 2 | | NIL | |



Signature
SUPERINTENDENT
T.H.Q. HOSPITAL
ALATHUR



[Handwritten signature]

SUPERINTENDENT
T.H.Q. HOSPITAL
ALATHUR