



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 175344

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Patta Junction, Poonithura, Kochi - 682 036, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bomp.kl@cyrix.in

<p>Health Facility <u>PHC ophthalmopathy</u></p> <p>Address <u>Palakkad</u> <u>Kerala</u></p> <p>Ph : <u>09544877578</u></p>	<p>Call Registration Date : <u>20/01/2024</u></p> <p>Caller ID : <u>91871</u></p> <p>Date of Visit : <u>21/01/2024</u></p> <p>Asset No. : <u>0941078</u></p> <p>EQPT Name : <u>weighing machine</u></p> <p>Manufacture <u>OMRON</u> Model : <u>HN-283</u></p> <p>S. No. <u>2014090353F</u> / Dept. <u>OP</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : machine not working

Action Taken : checked the unit Fault Found in unit not working. Found problem is sensor bushes are missing and display is damaged, need to replace these spares for further checking and working condition of the unit.


Completed Date : 29-01-2024 Time : 10:50 AM Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.				
2.				
3.				

Cyrux Engineer	Date	Start Time	End Time
<u>ABIJITH - B</u>	<u>29-01-2024</u>	<u>9:30 AM</u>	<u>10:50 AM</u>

Customer Remark Completed Pending



Service Engineer Name : ABIJITH - B

Signature : [Signature]

Date : 29-01-2024

Contact Number : 9567164651

Customer Name : [Signature]

Signature : [Signature]

Date : [Signature]

Contact Number : "മെഡിക്കൽ സൂപ്പർ ക്ലിനിക്കൽ"

Designation : കുടുംബ ആരോഗ്യ കേന്ദ്രം

Hospital Seal : മെഡിക്കൽ, പാലക്കാട്-678557

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Palakkad
2	Name of Hospital	PHC Ozhapathy
3	Name of Equipment with Make, Model and Serial Number	weighing machine make:- OMRON model:- HN-283 SN:- 2014090353F1
4	Equipment ID & Barcode	91871, 09H1078
5	Date of purchase / Year of manufacture / Installation Date	1-10-2018
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC / NO CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	20-01-2024, TOLL FREE
9	Action taken	checked the unit found problem sensor bushes are missing and display also damaged
10	Present status of the equipment (Fully damaged / partially damaged)	fully damaged
11	Recommendations for repair (required service details)	not recommendation
12	Cost of spares (specify parts and cost)	-

13	Asset Value	3800
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	—
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX service report.
16	Reasons for recommending the equipment as BER	checked the unit fault found in sensor bushes missing and display is damaged unit installed on 01/10/2018 and out live 5 plus years spare quite not submitting because spare are not available in market. since recommended for condemnation.
17	Name & Signature of CYRIX Authority	ABIJITH -B

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Display damaged and the sensor bushes missing hence recommended for BER.

Consultant (Biomedical Engg)
NHM (Arogyakeralam), Palakkad

[Signature]
31-01-24

Signature of JC BM (NHM)

Date

[Seal]

[Signature]
"മെഡിക്കൽ ഓഫീസർ ഇൻ ചാർജ്ജ്"
Superintendent / Medical Officer (i/c)
കുടുംബ ആരോഗ്യ കേന്ദ്രം
ഔലപ്പതി, പാലക്കാട്-678 557

S/No: Name

It is ^{to} Certified that this Book contains
1 to 196 Continuous pages.
From 29.09.1989
to

INVENTORY REGISTER.

VERIFIED
UP TO

21/07/18

P. Anil Kumar
Store Verification Officer (Incharge)
District Medical Office of Health
Civil Station-Palakkad



E1




FHC Ozharapathy
Inventory Register
Staff Nurse

INDEX

Sl. No.	Name of equipment	Page No.	
1.	REFRIGERATOR	6	
2.	GENERATOR	6	
3.	DEEP FREEZER	8	
4.	CHEST FREEZER (HR)	8	
5.	ELECTRO GAUARD STERILIZER	10	
6.	CHEST THERMOMETER	10	(Double entry) in substock register
7.	MINOR OPERATION TABLE	12	
8.	SHADOWLESS LAMP (PORTABLE)	12	(Double entry)
9.	Electric Sterilizer (Syringe)	13	(Double entry)
10.	Electric Sterilizer (Tred)	14	
11.	INHALER	16	
12.	Autoclave	16	
13.	Needle destroyer	17	(Double entry)
	Syringe cutter	18	
14.	Ambubag	18	(Double entry)
15.	V. Guard UPS	20	
	BATTERY		
16.	V. GUARD TROLLEY	20	
17.	STOVE (K.P.G)	22	
18.	Stethoscope	22	(Double entry)
19.	Torch light	24	
19.	Torch light	24	
20.	Height & weight measuring machine	24	
21.	DOBEL-BAVN BABYSCALE	26	
22.	Sterilizer (Small)	28	
23.	ceiling fan	28	
24.	Wall mounting fan	28	
25.	nebuliser piston type	30	
26.	BP Apparatus	30	
27.	Needle destroyer	30	
28.	Fridge, water	32	
29.	Steel Alomara	32	

S.No	Date	Name of Instrument	Qty	Life span	Date of expiry	Particulars
19	31.1.2007	Torch light Bill No: 152 (Bismi. Kjs)	1	3 yrs		250/100

S.No	Date	Name of instrument	Qty	Life span	Date of expiry	Particulars
20	01/04/2011	Height & weight measuring machine (Invoice no: 14 263) purchased from neo-surgical co. By using NRHM fund	1			28000 1
		11/10/12. Shuk entered in D2-161.				



മെഡിക്കൽ ഐ.ടി. സർവീസസ് പ്രൈവറ്റ് ലിമിറ്റഡ്
ആരോഗ്യ കേന്ദ്രം
പാലക്കാട് - 678 557

ആരോഗ്യ കേന്ദ്രം
പാലക്കാട് - 678 557

തിരു.
പാലക്കാട് - 678 557