

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Kottayam
2	Name of Hospital	General Hospital
3	Name of Equipment with Make, Model and Serial Number	Pharmacy Refrigerator Make: Videocon Model: nil
4	Equipment ID & Barcode	0510712 / 93080
5	Date of purchase / Year of manufacture / Installation Date	28/4/2006
6	Warranty details (Yes/No)	nil
7	*AMC/ CAMC Period agreed at the time of purchase	nil
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	29/1/2024
9	Action taken	Checked and found door was broken due to rust
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	The equipment was full of rust. Repairing is not possible
12	Cost of spares (specify parts and cost)	Spare cost not getting.

13	Asset Value	12000/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Spare cost not getting.
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix report attached.
16	Reasons for recommending the equipment as BER	The equipment was very old and repairing is not possible. Hence recommended the equipment as BER
17	Name & Signature of CYRIX Authority	J. Sureshwar


*Not mandatory



#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value - 12000
 Cost of spare - not maintained
 Life period - 17 years 9 months
 Cyrix reported that spare not available
 The refrigerator is 17 years old and fully damaged


 Signature of JC BM (NHM)

	<p>J.M 15/2/2019</p>	
<p>Superintendent / Medical Officer (i/c)</p>	<p>Superintendent i/c General Hospital Kottayam</p>	<p>Signature of Seal Date</p>



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 172559

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>GH</u>	Call Registration Date : <u>29/1/2024</u>
Address : <u>General hospital</u>	Caller ID : <u>93080</u>
<u>Kottayam</u>	Date of Visit : <u>08/2/24</u>
Ph : <u>9447356846</u>	Asset No. : <u>0510712</u>
	EQPT Name : <u>Pharmacy Refrigerator</u>
	Manufacture : <u>Videacon</u> Model : <u>-</u>
	S. No. : <u>-</u> Dept. : <u>Iyo 01</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Refrigerator doors detached from the equipment due to rust.

Action Taken : Found Refrigerator was very old wear and tear happened. Not repairable.


Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Sarith Sivas</u>	<u>8/2/24</u>	<u>9:00 am</u>	<u>10:00 am</u>

Customer Remark Completed Pending



Service Engineer Name : Sarith Sivas
Signature : [Signature]
Date : 8/2/24
Contact Number : 994646571

Customer Name : BABY ANITHA
Signature : [Signature]
Date : 15/02/2024
Contact Number : 9447356846
Designation : SNO
Hospital Seal :

STOCK BOOK OF

Fridge

118

NAME OF ARTICLE

Date	No. and date of invoice or bill	From whom received or to whom issued	Receipt	Issue	Balance after each transaction	Initials of receiver	Remarks
		B/F					
28/9/06	Ad from Edna Page 5	Handled by Jasco Fashion Jewellery. Videocass 165 Ltr. without stabilizer. Cost not known.	1				
2/10	Ad from store - (L.P)	Stabilizer (VOLTEK) Rs. 1900/-	1				
11	Fridge stand plastic (L.P)	Rs. 490/-	1				
11	5 cap 2 pin (0.5) (L.P)	Rs. 15/-	1				
27/11/06	6x16 Amp Power Plug. (SWITCH SOCKET & FUSE indicators) Received and issued to OJ for the use of sterilizer.		ALL				
		C/F					

